

TOTAL SOLIDS - SM2540B

SUPERVISOR: apatel

ANALYST: Aparna

Date: 04/03/2018

OVEN TEMP1: 103 °C TEMP: 104 °C

DATE & TIME IN: 04/02/2018 16:00 OUT: 04/03/2018 08:30

OVEN TEMP2: 104 °C TEMP: 105 °C

DATE & TIME IN: 04/03/2018 09:25 OUT: 04/03/2018 10:55

Run Number: LB94309

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB94309BL	LB94309BL	68.8506	68.8506	100	68.8506	68.8506	68.8506	0	0
2	J2187-01	37ST-2018	84.7612	84.7612	100	84.8856	84.8856	84.8856	0.1244	1244
3	J2187-01DUP	37ST-2018DUP	79.9949	79.9949	100	80.1205	80.1205	80.1205	0.1256	1256

A = Sample Volume (ml)
 B = Empty Dish Weight (g)
 C = Final Empty Dish + Sample weight after drying @105°C(g)
 D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TS040218

WorkList ID : 110319

Date : 4/2/2018 2:14:28 PM

2194309

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/09/2018	Water	J2187-01 H	TS	Cool 4 deg C	DIST01	L43	37ST-2018	04/02/2018	SM2540 B

Date/Time 04/02/18, 15:00
 Received by: AS
 Relinquished by: CP

Date/Time 04/02/18, 15:30
 Received by: AS
 Relinquished by: AS