



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB94313

Temp.: °C

BalanceID: WC SC-4

SLOPE.: 97.6

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.02	04/02/2018	16:45
2	CAL2	1	Water			7.01	04/02/2018	16:46
3	CAL3	1	Water			10.03	04/02/2018	16:47
4	ICV	1	Water			7.00	04/02/2018	16:48
5	CCV1	1	Water			2.01	04/02/2018	16:49
6	J2185-08	1	Solid	20.03	20	10.72	04/02/2018	16:50
7	J2185-08DUP	1	Solid	20.05	20	10.73	04/02/2018	16:51
8	J2185-17	1	Solid	20.04	20	11.45	04/02/2018	17:00
9	CCV2	1	Water			12.02	04/02/2018	17:02

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity j2185

WorkList ID : 110327

Date: 4/2/2018 4:27:00 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/02/2018	Solid	J2185-08	Corrosivity	Cool 4 deg C	PSEG01	L31	LOWER-CONCRETE-SLAB-1	04/02/2018	9045C
04/02/2018	Solid	J2185-17	Corrosivity	Cool 4 deg C	PSEG01	L31	LOWER-CONCRETE-SLAB-2	04/02/2018	9045C

Date/Time 04/02/18 4:50 PM
Received by: J
Relinquished by: W

Date/Time 8/16/14 5:45 PM
 Received by: CM
 Relinquished by: JS