



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: KETANKUMAR

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB94329

Ambient Barometric Pressure (mmHg): 755

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	82.3	04/02/2018	09:02
2	J2172-02		1	>100.0	>212.0	>212.0	04/02/2018	09:32
3	J2172-02DUP		1	>100.0	>212.0	>212.0	04/02/2018	09:50
4	J2172-03		1	>100.0	>212.0	>212.0	04/02/2018	10:15
5	J2172-04		1	>100.0	>212.0	>212.0	04/02/2018	10:40
6	J2172-05		1	>100.0	>212.0	>212.0	04/02/2018	11:02
7	J2172-06		1	>100.0	>212.0	>212.0	04/02/2018	11:30
8	J2172-07		1	>100.0	>212.0	>212.0	04/02/2018	11:52
9	J2172-08		1	>100.0	>212.0	>212.0	04/02/2018	12:15
10	J2172-09		1	>100.0	>212.0	>212.0	04/02/2018	12:40
11	J2172-11		1	>100.0	>212.0	>212.0	04/02/2018	13:02
12	J2172-12		1	>100.0	>212.0	>212.0	04/02/2018	13:25
13	J2172-13		1	>100.0	>212.0	>212.0	04/02/2018	13:45
14	J2172-14		1	>100.0	>212.0	>212.0	04/02/2018	14:15
15	J2172-15		1	>100.0	>212.0	>212.0	04/02/2018	14:40
16	J2172-16		1	>100.0	>212.0	>212.0	04/02/2018	14:58
17	J2172-17		1	>100.0	>212.0	>212.0	04/02/2018	15:16
18	J2172-18		1	>100.0	>212.0	>212.0	04/02/2018	15:42

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.33

WORKLIST(Hardcopy Internal Chain)

LB94329

WorkList Name : FP4218

WorkList ID : 110289

Date : 4/2/2018 8:31:09 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/06/2018	Water	J2172-02	Flash Point	Cool 4 deg C	PSEG01	J63	22043	03/30/2018	1010A
04/06/2018	Water	J2172-03	Flash Point	Cool 4 deg C	PSEG01	J63	22044	03/30/2018	1010A
04/06/2018	Water	J2172-04	Flash Point	Cool 4 deg C	PSEG01	J63	22045	03/30/2018	1010A
04/06/2018	Water	J2172-05	Flash Point	Cool 4 deg C	PSEG01	J63	22132	03/30/2018	1010A
04/06/2018	Water	J2172-06	Flash Point	Cool 4 deg C	PSEG01	J63	22133	03/30/2018	1010A
04/06/2018	Water	J2172-07	Flash Point	Cool 4 deg C	PSEG01	J63	22134	03/30/2018	1010A
04/06/2018	Water	J2172-08	Flash Point	Cool 4 deg C	PSEG01	J63	22178	03/30/2018	1010A
04/06/2018	Water	J2172-09	Flash Point	Cool 4 deg C	PSEG01	J63	22180	03/30/2018	1010A
04/06/2018	Water	J2172-11	Flash Point	Cool 4 deg C	PSEG01	J63	22181	03/30/2018	1010A
04/06/2018	Water	J2172-12	Flash Point	Cool 4 deg C	PSEG01	J63	22182	03/30/2018	1010A
04/06/2018	Water	J2172-13	Flash Point	Cool 4 deg C	PSEG01	J63	22183	03/30/2018	1010A
04/06/2018	Water	J2172-14	Flash Point	Cool 4 deg C	PSEG01	J63	22184	03/30/2018	1010A
04/06/2018	Water	J2172-15	Flash Point	Cool 4 deg C	PSEG01	J63	22185	03/30/2018	1010A
04/06/2018	Water	J2172-16	Flash Point	Cool 4 deg C	PSEG01	J63	22186	03/30/2018	1010A
04/06/2018	Water	J2172-17	Flash Point	Cool 4 deg C	PSEG01	J63	22187	03/30/2018	1010A
04/06/2018	Water	J2172-18	Flash Point	Cool 4 deg C	PSEG01	J63	22188	03/30/2018	1010A

Date/Time 4/2/18 8:35 AM

Received by: RS

Relinquished by: RS

Date/Time 4/2/18 15:50

Received by: RS

Relinquished by: RS