



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: KETANKUMAR

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB94330

Ambient Barometric Pressure (mmHg): 765

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	28.0	82.4	80.8	04/03/2018	09:20
2	J2187-01		1	>100.0	>212.0	>212.0	04/03/2018	09:58
3	J2187-01DUP		1	>100.0	>212.0	>212.0	04/03/2018	10:25

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.33

WORKLIST(Hardcopy Internal Chain)

LB94330

WorkList Name : FP4318

WorkList ID : 110343

Date : 4/3/2018 8:11:09 AM

Due Date	Matrix	Sample	Test	Flash Point	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/14/2018	Water	J2187-01	✓	Flash Point	Cool 4 deg C	DIST01	L43	37ST-2018	04/02/2018	1010A

Date/Time 4/3/18 8:15 AM
 Received by: RS
 Relinquished by: RS

Date/Time 4/3/18 10:30 AM
 Received by: RS
 Relinquished by: RS