

245.1


LB94333

INSTRUMENT ID: CV2

Linear

 μ Abs.:

35059



A= 0.0000e+000

B= 2.8400e-004

C= -2.4558e-002

Rho= 0.9999196

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.010	0.010	120	0.000	120				
0.05	0.050	0.058	0.008	290	0.0 %	290				
0.2	0.200	0.184	-0.016	735	0.0 %	735				
2.5	2.500	2.451	-0.049	8716	0.0 %	8716				
5.0	5.000	5.031	0.031	17802	0.0 %	17802				
7.5	7.500	7.584	0.084	26792	0.0 %	26792				
10.0	10.000	9.932	-0.068	35059	0.0 %	35059				

LB94333 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 03 Apr 2018 10:52:35

Sample ID	Extended ID	µ Abs	Conc.	Std Cond	Method	Units	Date	Type	Type
0.0 - 1	SO	120	-	0.0000	245.1	PPB	03 Apr 2018 10:55:56	S	Std
0.05 - 1	SO 05	290	-	0.0500	245.1	PPB	03 Apr 2018 10:58:05	S	Std
0.2 - 1	SO 2	735	-	0.2000	245.1	PPB	03 Apr 2018 11:00:14	S	Std
2.5 - 1	SO 2.5	8716	-	2.5000	245.1	PPB	03 Apr 2018 11:02:21	S	Std
5.0 - 1	SO 5	17802	-	5.0000	245.1	PPB	03 Apr 2018 11:04:30	S	Std
7.5 - 1	SO 7.5	26792	-	7.5000	245.1	PPB	03 Apr 2018 11:06:37	S	Std
10.0 - 1	SO 10	35059	-	10.0000	245.1	PPB	03 Apr 2018 11:08:45	S	Std
ICV72 - 1	ICV72	13803	3.8955	-245.1		PPB	03 Apr 2018 11:22:07	U	SMPL
ICB72 - 1	ICB72	8	-0.0223	-245.1		PPB	03 Apr 2018 11:24:13	U	SMPL
CCV29 - 1	CCV29	17839	5.0417	-245.1		PPB	03 Apr 2018 11:26:21	U	SMPL
CCB29 - 1	CCB29	-41	-0.0362	-245.1		PPB	03 Apr 2018 11:28:27	U	SMPL
CRA - 1	CRA	249	0.0462	-245.1		PPB	03 Apr 2018 11:33:03	U	SMPL
HighStd - 1	HighStd	35404	10.0302	-245.1		PPB	03 Apr 2018 11:35:09	U	SMPL
ChkStd - 1	ChkStd	24822	7.0249	-245.1		PPB	03 Apr 2018 11:37:16	U	SMPL
PB107917BL - 1	PBW	7	-0.0226	-245.1		PPB	03 Apr 2018 11:41:50	U	SMPL
PB107917BS - 1	LCSW	13015	3.6717	-245.1		PPB	03 Apr 2018 11:43:58	U	SMPL
J2177-01 - 1	032818-DBRI-F-D-B	-86	-0.0490	-245.1		PPB	03 Apr 2018 11:46:05	U	SMPL
J2177-02 - 1	032818-DBRI-F-D-M	14	-0.0206	-245.1		PPB	03 Apr 2018 11:48:14	U	SMPL
J2177-03 - 1	032818-DBRI-F-D-S	37	-0.0141	-245.1		PPB	03 Apr 2018 11:50:22	U	SMPL
J2177-04 - 1	032818-DBRI-F-U-B	67	-0.0055	-245.1		PPB	03 Apr 2018 11:52:31	U	SMPL
J2177-05 - 1	032818-DBRI-F-U-M	64	-0.0064	-245.1		PPB	03 Apr 2018 11:54:38	U	SMPL
CCV30 - 1	CCV30	17423	4.9236	-245.1		PPB	03 Apr 2018 11:56:44	U	SMPL
CCB30 - 1	CCB30	-50	-0.0388	-245.1		PPB	03 Apr 2018 11:58:51	U	SMPL
J2177-06 - 1	032818-DBRI-F-U-S	16	-0.0200	-245.1		PPB	03 Apr 2018 12:00:59	U	SMPL
J2177-07 - 1	032818-JETB-F-D-B	81	-0.0016	-245.1		PPB	03 Apr 2018 12:03:05	U	SMPL
J2177-08 - 1	032818-JETB-F-D-M	57	-0.0084	-245.1		PPB	03 Apr 2018 12:05:12	U	SMPL
J2177-09 - 1	032818-JETB-F-D-S	74	-0.0035	-245.1		PPB	03 Apr 2018 12:07:19	U	SMPL
J2177-10 - 1	032818-JETB-F-U-B	41	-0.0129	-245.1		PPB	03 Apr 2018 12:09:27	U	SMPL
J2177-11 - 1	032818-JETB-F-U-M	60	-0.0075	-245.1		PPB	03 Apr 2018 12:11:35	U	SMPL
J2177-12 - 1	032818-JETB-F-U-S	47	-0.0112	-245.1		PPB	03 Apr 2018 12:13:42	U	SMPL
J2177-13 - 1	032818-SIDI-E-D-B	65	-0.0061	-245.1		PPB	03 Apr 2018 12:15:51	U	SMPL
J2177-14 - 1	032818-SIDI-E-D-S	85	-0.0004	-245.1		PPB	03 Apr 2018 12:17:59	U	SMPL
J2177-15 - 1	032818-SIDI-E-U-B	60	-0.0075	-245.1		PPB	03 Apr 2018 12:20:06	U	SMPL
CCV31 - 1	CCV31	17532	4.9545	-245.1		PPB	03 Apr 2018 12:22:13	U	SMPL
CCB31 - 1	CCB31	-64	-0.0427	-245.1		PPB	03 Apr 2018 12:24:20	U	SMPL
J2177-16 - 1	032818-SIDI-E-U-S	28	-0.0166	-245.1		PPB	03 Apr 2018 12:26:28	U	SMPL
J2177-16DUP - 1	032818-SIDI-E-U-SDUP	54	-0.0092	-245.1		PPB	03 Apr 2018 12:28:35	U	SMPL
J2177-16MS - 1	032818-SIDI-E-U-SMS	14401	4.0653	-245.1		PPB	03 Apr 2018 12:30:42	U	SMPL
J2177-16MSD - 1	032818-SIDI-E-U-SMSD	14575	4.1147	-245.1		PPB	03 Apr 2018 12:32:49	U	SMPL
PB107918BL - 1	PBW	-122	-0.0592	-245.1		PPB	03 Apr 2018 12:34:57	U	SMPL
PB107918BS - 1	LCSW	12941	3.6507	-245.1		PPB	03 Apr 2018 12:37:04	U	SMPL
J2149-02 - 1	NEWBUILDING-WATERFOUNTAIN	-95	-0.0515	-245.1		PPB	03 Apr 2018 12:39:13	U	SMPL
J2149-02DUP - 1	NEWBUILDING-WATERFOUNTAINDUP	16	-0.0200	-245.1		PPB	03 Apr 2018 12:41:22	U	SMPL
J2149-02MS - 1	NEWBUILDING-WATERFOUNTAINMS	13039	3.6785	-245.1		PPB	03 Apr 2018 12:43:30	U	SMPL
J2149-02MSD - 1	NEWBUILDING-WATERFOUNTAINMSD	13363	3.7705	-245.1		PPB	03 Apr 2018 12:45:37	U	SMPL
CCV32 - 1	CCV32	17202	4.8608	-245.1		PPB	03 Apr 2018 12:47:45	U	SMPL
CCB32 - 1	CCB32	-50	-0.0388	-245.1		PPB	03 Apr 2018 12:49:53	U	SMPL
J2149-04 - 1	OLDBUILDING-WATERFOUNTAIN	-22	-0.0308	-245.1		PPB	03 Apr 2018 12:52:01	U	SMPL
J2179-01 - 1	032918-DBRI-F-D-B	35	-0.0146	-245.1		PPB	03 Apr 2018 12:54:08	U	SMPL
J2179-02 - 1	032918-DBRI-F-D-M	38	-0.0138	-245.1		PPB	03 Apr 2018 12:56:15	U	SMPL
J2179-03 - 1	032918-DBRI-F-D-S	67	-0.0055	-245.1		PPB	03 Apr 2018 12:58:22	U	SMPL
J2179-04 - 1	032918-DBRI-F-U-B	72	-0.0041	-245.1		PPB	03 Apr 2018 13:00:29	U	SMPL
J2179-05 - 1	032918-DBRI-F-U-M	62	-0.0070	-245.1		PPB	03 Apr 2018 13:02:37	U	SMPL
J2179-06 - 1	032918-DBRI-F-U-S	36	-0.0143	-245.1		PPB	03 Apr 2018 13:04:45	U	SMPL
J2179-07 - 1	032918-SIDI-E-D-B	8	-0.0223	-245.1		PPB	03 Apr 2018 13:06:53	U	SMPL
J2179-08 - 1	032918-SIDI-E-D-S	68	-0.0052	-245.1		PPB	03 Apr 2018 13:09:01	U	SMPL
J2179-09 - 1	032918-SIDI-E-U-B	32	-0.0155	-245.1		PPB	03 Apr 2018 13:11:08	U	SMPL
CCV33 - 1	CCV33	17560	4.9625	-245.1		PPB	03 Apr 2018 13:13:15	U	SMPL
CCB33 - 1	CCB33	-39	-0.0356	-245.1		PPB	03 Apr 2018 13:15:22	U	SMPL
J2179-10 - 1	032918-SIDI-E-U-S	46	-0.0115	-245.1		PPB	03 Apr 2018 13:17:30	U	SMPL
J2187-01 - 1	37ST-2018	79	-0.0021	-245.1		PPB	03 Apr 2018 13:19:37	U	SMPL
J2177-16LX5 - 1		0	-0.0246	-245.1		PPB	03 Apr 2018 13:21:44	U	SMPL
J2177-16A - 1		15616	4.4104	-245.1		PPB	03 Apr 2018 13:23:51	U	SMPL
J2149-02LX5 - 1		-122	-0.0592	-245.1		PPB	03 Apr 2018 13:25:59	U	SMPL
J2149-02A - 1		13000	3.6674	-245.1		PPB	03 Apr 2018 13:28:07	U	SMPL
CCV34 - 1	CCV34	17803	5.0315	-245.1		PPB	03 Apr 2018 13:30:15	U	SMPL
CCB34 - 1	CCB34	-46	-0.0376	-245.1		PPB	03 Apr 2018 13:32:24	U	SMPL