


7470A

LB94338
INSTRUMENT ID: CV1

Linear

 μ Abs.:

35308



A= 0.0000e+000

B= 2.8265e-004

C= -2.5314e-002

Rho= 0.9998775

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.001	-0.001	87	0.000	87				
0.2	0.200	0.202	0.002	803	0.0 %	803				
2.5	2.500	2.425	-0.075	8668	0.0 %	8668				
5.0	5.000	5.111	0.111	18170	0.0 %	18170				
7.5	7.500	7.509	0.009	26656	0.0 %	26656				
10.0	10.000	9.955	-0.045	35308	0.0 %	35308				

LB94338 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 03 Apr 2018 14:10:00

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	87	-	0.00007470A	PPB		03 Apr 2018 14:16:43	S	Std
0.2 - 1	50.2	803	-	0.20007470A	PPB		03 Apr 2018 14:18:51	S	Std
2.5 - 1	52.5	8668	-	2.50007470A	PPB		03 Apr 2018 14:21:00	S	Std
5.0 - 1	55	18170	-	5.00007470A	PPB		03 Apr 2018 14:23:08	S	Std
7.5 - 1	57.5	26656	-	7.50007470A	PPB		03 Apr 2018 14:25:16	S	Std
10.0 - 1	57.0	35308	-	10.00007470A	PPB		03 Apr 2018 14:27:25	S	Std
ICV73 - 1	ICV73	14003	3.9327	-7470A	PPB		03 Apr 2018 14:29:59	U	SMPL
ICB73 - 1	ICB73	-54	-0.0406	-7470A	PPB		03 Apr 2018 14:32:06	U	SMPL
CCV35 - 1	CCV35	18242	5.1309	-7470A	PPB		03 Apr 2018 14:34:14	U	SMPL
CCB35 - 1	CCB35	-87	-0.0499	-7470A	PPB		03 Apr 2018 14:36:20	U	SMPL
CRA - 1	CRA	752	0.1872	-7470A	PPB		03 Apr 2018 14:38:28	U	SMPL
HighStd - 1	HighStd	35663	10.0550	-7470A	PPB		03 Apr 2018 14:40:34	U	SMPL
ChkStd - 1	ChkStd	25273	7.1182	-7470A	PPB		03 Apr 2018 14:42:42	U	SMPL
PB107919BL - 1	PBW	-244	-0.0943	-7470A	PPB		03 Apr 2018 14:44:51	U	SMPL
PB107919BS - 1	LCSW	13032	3.6582	-7470A	PPB		03 Apr 2018 14:47:01	U	SMPL
PB107919TB - 1	PB107870TB	-106	-0.0553	-7470A	PPB		03 Apr 2018 14:49:08	U	SMPL
J1757-08 - 1	EO-03-033018-A	10	-0.0225	-7470A	PPB		03 Apr 2018 14:51:17	U	SMPL
J1757-10 - 1	EO-03-033018-B	15	-0.0211	-7470A	PPB		03 Apr 2018 14:53:25	U	SMPL
J1757-12 - 1	EO-03-033018-C	7	-0.0233	-7470A	PPB		03 Apr 2018 14:55:33	U	SMPL
J2175-03 - 1	TP-3	26	-0.0180	-7470A	PPB		03 Apr 2018 14:57:39	U	SMPL
CCV36 - 1	CCV36	18171	5.1108	-7470A	PPB		03 Apr 2018 14:59:45	U	SMPL
CCB36 - 1	CCB36	-93	-0.0516	-7470A	PPB		03 Apr 2018 15:01:52	U	SMPL
J2178-04 - 1	S6-TP	-1	-0.0256	-7470A	PPB		03 Apr 2018 15:04:00	U	SMPL
J2178-04DUP - 1	S6-TPDUP	16	-0.0208	-7470A	PPB		03 Apr 2018 15:06:06	U	SMPL
J2178-04MS - 1	S6-TPMS	13155	3.6930	-7470A	PPB		03 Apr 2018 15:08:13	U	SMPL
J2178-04MSD - 1	S6-TPMSD	13608	3.8211	-7470A	PPB		03 Apr 2018 15:10:20	U	SMPL
PB107920BL - 1	PBW	-114	-0.0575	-7470A	PPB		03 Apr 2018 15:12:27	U	SMPL
PB107920BS - 1	LCSW	13370	3.7538	-7470A	PPB		03 Apr 2018 15:14:35	U	SMPL
PB107920TB - 1	PB107872TB	-113	-0.0573	-7470A	PPB		03 Apr 2018 15:16:43	U	SMPL
J2172-01 - 1	22043-5-132-4-78-80	1890	0.5089	-7470A	PPB		03 Apr 2018 15:18:52	U	SMPL
J2172-10 - 1	22181-8	-19	-0.0307	-7470A	PPB		03 Apr 2018 15:21:00	U	SMPL
J2172-10DUP - 1	22181-8DUP	43	-0.0132	-7470A	PPB		03 Apr 2018 15:23:07	U	SMPL
CCV37 - 1	CCV37	17838	5.0167	-7470A	PPB		03 Apr 2018 15:25:14	U	SMPL
CCB37 - 1	CCB37	-39	-0.0363	-7470A	PPB		03 Apr 2018 15:27:20	U	SMPL
J2172-10MS - 1	22181-8MS	13398	3.7617	-7470A	PPB		03 Apr 2018 15:29:28	U	SMPL
J2172-10MSD - 1	22181-8MSD	13652	3.8335	-7470A	PPB		03 Apr 2018 15:31:35	U	SMPL
J2178-04LX5 - 1		-69	-0.0448	-7470A	PPB		03 Apr 2018 15:33:43	U	SMPL
J2178-04A - 1		13617	3.8236	-7470A	PPB		03 Apr 2018 15:35:51	U	SMPL
J2172-10LX5 - 1		-91	-0.0510	-7470A	PPB		03 Apr 2018 15:37:58	U	SMPL
J2172-10A - 1		13689	3.8439	-7470A	PPB		03 Apr 2018 15:40:06	U	SMPL
CCV38 - 1	CCV38	18106	5.0924	-7470A	PPB		03 Apr 2018 15:42:14	U	SMPL
CCB38 - 1	CCB38	-113	-0.0573	-7470A	PPB		03 Apr 2018 15:44:23	U	SMPL