

=====

Test results

Aquakem 7.2AQ1

1B94340
Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

4/3/2018 13:18

Test: ^{Reactive} ~~Total~~ CNMg/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	96.513	0.0	0.097	
ICB1	-0.434	0.0	0.004	
CCV1	240.796	0.0	0.235	
CCB1	-0.165	0.0	0.004	
J2185-08	-0.492	0.0	0.004	
J2185-08DUP	-0.527	0.0	0.004	
J2185-17	-0.375	0.0	0.004	
PB107932BL	-0.613	0.0	0.004	
CCV2	258.301	0.0	0.251	
CCB2	-0.468	0.0	0.004	

N	10
Mean	59.254
SD	104.8325
CV%	176.92

AS

Aquakem v. 7.2AQ1

Results from time period:

Tue Apr 03 11:47:18 2018

Tue Apr 03 13:14:16 2018

Sample Id	Sam/Ctr/c	Test short nam	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	-0.6088	µg/l	4/3/2018 12:27:12	
5.OPPBCN	A	Reactive CN	P	4.1817	µg/l	4/3/2018 12:27:13	
10PPBCN	A	Reactive CN	P	9.2661	µg/l	4/3/2018 12:27:14	
50PPBCN	A	Reactive CN	P	49.2007	µg/l	4/3/2018 12:27:15	
100PPBCN	A	Reactive CN	P	101.8001	µg/l	4/3/2018 12:27:16	
250PPBCN	A	Reactive CN	P	252.8346	µg/l	4/3/2018 12:27:17	
500PPBCN	A	Reactive CN	P	498.3255	µg/l	4/3/2018 12:27:18	
ICV1	S	Reactive CN	P	96.5126	µg/l	4/3/2018 12:50:47	
ICB1	S	Reactive CN	P	-0.4335	µg/l	4/3/2018 12:50:48	
CCV1	S	Reactive CN	P	240.7963	µg/l	4/3/2018 12:50:49	
CCB1	S	Reactive CN	P	-0.1649	µg/l	4/3/2018 12:50:50	
J2185-08	S	Reactive CN	P	-0.4918	µg/l	4/3/2018 12:50:52	
J2185-08DUP	S	Reactive CN	P	-0.5267	µg/l	4/3/2018 12:50:53	
J2185-17	S	Reactive CN	P	-0.3753	µg/l	4/3/2018 12:50:54	
PB107932BL	S	Reactive CN	P	-0.6135	µg/l	4/3/2018 13:14:14	
CCV2	S	Reactive CN	P	258.3012	µg/l	4/3/2018 13:14:15	
CCB2	S	Reactive CN	P	-0.4683	µg/l	4/3/2018 13:14:16	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

4/3/2018 12:28


Test Total CN

Accepted 4/3/2018 12:28

Factor 1044
Bias 0.004

Coeff. of det. 0.999921

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	-0.6088	0.0000	
2	5.0PPBCN	0.008	4.1817	5.0000	
3	10PPBCN	0.013	9.2661	10.0000	
4	50PPBCN	0.051	49.2007	50.0000	
5	100PPBCN	0.102	101.8001	100.0000	
6	250PPBCN	0.246	252.8346	250.0000	
7	500PPBCN	0.481	498.3255	500.0000	