


7471B

LB 94342
INSTRUMENT ID: CV4

Linear

 μ Abs.:

41046



A= 0.0000e+000

B= 2.4364e-004

C= 2.0620e-002

Rho= 0.9999657

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.031	0.031	41	0.000	41				
0.2	0.200	0.216	0.016	801	0.0 %	801				
2.5	2.500	2.469	-0.031	10049	0.0 %	10049				
5.0	5.000	4.947	-0.053	20222	0.0 %	20222				
7.5	7.500	7.516	0.016	30765	0.0 %	30765				
10	10.000	10.021	0.021	41046	0.0 %	41046				

LB94342 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 03 Apr 2018 16:06:23					
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	41	-	0.000	7471B	PPB	03 Apr 2018 16:09:56	S	Std
0.2 - 1	50.2	801	-	0.200	7471B	PPB	03 Apr 2018 16:12:14	S	Std
2.5 - 1	52.5	10049	-	2.500	7471B	PPB	03 Apr 2018 16:14:31	S	Std
5.0 - 1	55	20222	-	5.000	7471B	PPB	03 Apr 2018 16:16:48	S	Std
7.5 - 1	57.5	30765	-	7.500	7471B	PPB	03 Apr 2018 16:19:05	S	Std
10 - 1	510	41046	-	10.000	7471B	PPB	03 Apr 2018 16:24:49	S	Std
ICV74 - 1	ICV74	16589	4.0623	-	7471B	PPB	03 Apr 2018 16:27:32	U	SMPL
ICB74 - 1	ICB74	-130	-0.0111	-	7471B	PPB	03 Apr 2018 16:29:48	U	SMPL
CCV39 - 1	CCV39	21063	5.1524	-	7471B	PPB	03 Apr 2018 16:32:04	U	SMPL
CCB39 - 1	CCB39	-136	-0.0125	-	7471B	PPB	03 Apr 2018 16:34:20	U	SMPL
CRA - 1	CRA	829	0.2226	-	7471B	PPB	03 Apr 2018 16:36:37	U	SMPL
HighStd - 1	HighStd	41235	10.0671	-	7471B	PPB	03 Apr 2018 16:38:52	U	SMPL
ChkStd - 1	ChkStd	28228	6.8981	-	7471B	PPB	03 Apr 2018 16:41:09	U	SMPL
PB107934BL - 1	PBS	-266	-0.0442	-	7471B	PPB	03 Apr 2018 16:43:27	U	SMPL
PB107934BS - 1	LCSS	16020	3.9237	-	7471B	PPB	03 Apr 2018 16:45:46	U	SMPL
J1757-07 - 1	EO-03-033018-A	1754	0.4480	-	7471B	PPB	03 Apr 2018 16:48:03	U	SMPL
J1757-09 - 1	EO-03-033018-B	1499	0.3858	-	7471B	PPB	03 Apr 2018 16:50:20	U	SMPL
J1757-11 - 1	EO-03-033018-C	4321	1.0734	-	7471B	PPB	03 Apr 2018 16:52:38	U	SMPL
J2148-01 - 1	SB-4C	8594	2.1145	-	7471B	PPB	03 Apr 2018 16:54:55	U	SMPL
J2148-02 - 1	SB-2A	14102	3.4564	-	7471B	PPB	03 Apr 2018 16:57:12	U	SMPL
CCV40 - 1	CCV40	20602	5.0401	-	7471B	PPB	03 Apr 2018 16:59:28	U	SMPL
CCB40 - 1	CCB40	-104	-0.0047	-	7471B	PPB	03 Apr 2018 17:01:43	U	SMPL
J2148-03 - 1	SB-2B	4713	1.1689	-	7471B	PPB	03 Apr 2018 17:03:59	U	SMPL
J2148-04 - 1	SB-2C	735	0.1997	-	7471B	PPB	03 Apr 2018 17:06:16	U	SMPL
J2148-05 - 1	SB-4A	544	0.1532	-	7471B	PPB	03 Apr 2018 17:08:32	U	SMPL
J2148-06 - 1	SB-4B	2245	0.5676	-	7471B	PPB	03 Apr 2018 17:10:48	U	SMPL
J2169-01 - 1	BROOK-BROOKLYN	824	0.2214	-	7471B	PPB	03 Apr 2018 17:19:24	U	SMPL
J2169-01DUP - 1	BROOK-BROOKLYNDUP	755	0.2046	-	7471B	PPB	03 Apr 2018 17:21:40	U	SMPL
J2169-01MS - 1	BROOK-BROOKLYNMS	15478	3.7917	-	7471B	PPB	03 Apr 2018 17:23:57	U	SMPL
J2169-01MSD - 1	BROOK-BROOKLYNMSD	15486	3.7936	-	7471B	PPB	03 Apr 2018 17:26:15	U	SMPL
J2170-01 - 1	EPE-124-2(30-32)	86	0.0416	-	7471B	PPB	03 Apr 2018 17:28:33	U	SMPL
J2170-02 - 1	EPE-124-2(55-57)	1173	0.3064	-	7471B	PPB	03 Apr 2018 17:30:49	U	SMPL
CCV41 - 1	CCV41	21224	5.1916	-	7471B	PPB	03 Apr 2018 17:33:05	U	SMPL
CCB41 - 1	CCB41	-70	0.0036	-	7471B	PPB	03 Apr 2018 17:35:21	U	SMPL
J2175-01 - 1	TP-3	69449	16.9411	-	7471B	PPB	03 Apr 2018 17:37:38	U	SMPL
J2178-01 - 1	S6-TP	4995	1.2376	-	7471B	PPB	03 Apr 2018 17:39:54	U	SMPL
J2185-01 - 1	LOWER-CONCRETE-SLAB-1	351	0.1061	-	7471B	PPB	03 Apr 2018 17:42:12	U	SMPL
J2185-10 - 1	LOWER-CONCRETE-SLAB-2	2111	0.5349	-	7471B	PPB	03 Apr 2018 17:44:29	U	SMPL
PB107935BL - 1	PBS	-3	0.0199	-	7471B	PPB	03 Apr 2018 17:46:45	U	SMPL
PB107935BS - 1	LCSS	15274	3.7420	-	7471B	PPB	03 Apr 2018 17:49:02	U	SMPL
J2182-01 - 1	6-WM-DIP-1-EW@5.5	1218	0.3174	-	7471B	PPB	03 Apr 2018 17:51:19	U	SMPL
J2182-02 - 1	6-WM-DIP-1-EW@10.5	997	0.2635	-	7471B	PPB	03 Apr 2018 17:53:37	U	SMPL
J2182-03 - 1	6-WM-DIP-1-WW@5.5	1770	0.4519	-	7471B	PPB	03 Apr 2018 17:55:55	U	SMPL
J2182-04 - 1	6-WM-DIP-2-EW@4.5	1265	0.3288	-	7471B	PPB	03 Apr 2018 17:58:11	U	SMPL
CCV42 - 1	CCV42	21258	5.1999	-	7471B	PPB	03 Apr 2018 18:00:28	U	SMPL
CCB42 - 1	CCB42	-80	0.0011	-	7471B	PPB	03 Apr 2018 18:02:44	U	SMPL
J2182-05 - 1	6-WM-DIP-2-WW@4.5	1629	0.4175	-	7471B	PPB	03 Apr 2018 18:05:01	U	SMPL
J2182-06 - 1	6-WM-DIP-3-EW@3.5	1083	0.2845	-	7471B	PPB	03 Apr 2018 18:07:17	U	SMPL
J2182-07 - 1	6-WM-DIP-3-WW@3.5	1170	0.3057	-	7471B	PPB	03 Apr 2018 18:09:33	U	SMPL
J2182-08 - 1	6-WM-DIP-4-EW@3.5	2273	0.5744	-	7471B	PPB	03 Apr 2018 18:11:49	U	SMPL
J2182-09 - 1	6-WM-DIP-4-WW@3.5	1539	0.3956	-	7471B	PPB	03 Apr 2018 18:14:06	U	SMPL
J2182-10 - 1	6-WM-DIP-5-EW@3	5525	1.3667	-	7471B	PPB	03 Apr 2018 18:16:23	U	SMPL
J2182-11 - 1	6-WM-DIP-5-WW@3	6640	1.6384	-	7471B	PPB	03 Apr 2018 18:18:41	U	SMPL
J2182-12 - 1	6-WM-DIP-6-EW(2.5-5)	1091	0.2864	-	7471B	PPB	03 Apr 2018 18:30:33	U	SMPL
J2182-12DUP - 1	6-WM-DIP-6-EW(2.5-5)DUP	1148	0.3003	-	7471B	PPB	03 Apr 2018 18:32:51	U	SMPL
J2182-13 - 1	J2182-12MS	16798	4.1133	-	7471B	PPB	03 Apr 2018 18:35:07	U	SMPL
CCV43 - 1	CCV43	20490	5.0128	-	7471B	PPB	03 Apr 2018 18:37:24	U	SMPL
CCB43 - 1	CCB43	-84	0.0002	-	7471B	PPB	03 Apr 2018 18:39:40	U	SMPL
J2182-14 - 1	J2182-12MSD	18169	4.4473	-	7471B	PPB	03 Apr 2018 18:44:33	U	SMPL
J2182-15 - 1	6-WM-DIP-6-WW(2.5-5)	3618	0.9021	-	7471B	PPB	03 Apr 2018 18:46:49	U	SMPL
J2182-16 - 1	6-WM-DIP-7-EW(1-5)	4618	1.1457	-	7471B	PPB	03 Apr 2018 18:49:05	U	SMPL
J2182-17 - 1	6-WM-DIP-7-WW(1-5)	7208	1.7768	-	7471B	PPB	03 Apr 2018 18:51:21	U	SMPL
J2182-18 - 1	6-WM-DIP-8-EW(1-5)	16391	4.0141	-	7471B	PPB	03 Apr 2018 18:53:37	U	SMPL
J2182-19 - 1	6-WM-DIP-8-WW(1-5)	12711	3.1175	-	7471B	PPB	03 Apr 2018 18:55:55	U	SMPL
J2182-20 - 1	ES-8-WATER-LATERAL-1-NW@3.5	922	0.2453	-	7471B	PPB	03 Apr 2018 18:58:12	U	SMPL
J2182-21 - 1	J2182-20MS	18217	4.4590	-	7471B	PPB	03 Apr 2018 19:00:30	U	SMPL
J2182-22 - 1	J2182-20MSD	18413	4.5067	-	7471B	PPB	03 Apr 2018 19:02:48	U	SMPL
PB107936BL - 1	PBS	-76	0.0021	-	7471B	PPB	03 Apr 2018 19:05:05	U	SMPL
CCV44 - 1	CCV44	20463	5.0062	-	7471B	PPB	03 Apr 2018 19:07:22	U	SMPL
CCB44 - 1	CCB44	-67	0.0043	-	7471B	PPB	03 Apr 2018 19:09:38	U	SMPL
PB107936BS - 1	LCSS	15841	3.8801	-	7471B	PPB	03 Apr 2018 19:11:55	U	SMPL
J2183-01 - 1	ES-8-WATER-LATERAL-1-SW@3.5	1321	0.3425	-	7471B	PPB	03 Apr 2018 19:14:11	U	SMPL
J2183-02 - 1	ES-8-WATER-LATERAL-2-NW@3	1704	0.4358	-	7471B	PPB	03 Apr 2018 19:16:28	U	SMPL
J2183-03 - 1	ES-8-WATER-LATERAL-2-SW@3	1286	0.3339	-	7471B	PPB	03 Apr 2018 19:18:45	U	SMPL
J2183-04 - 1	ES-4-DWS-1-SW@4.5	13779	3.3777	-	7471B	PPB	03 Apr 2018 19:21:02	U	SMPL
J2183-05 - 1	ES-4-DWS-1-NW@4.5	12938	3.1728	-	7471B	PPB	03 Apr 2018 19:28:39	U	SMPL
J2183-05DUP - 1	ES-4-DWS-1-NW@4.5DUP	12825	3.1453	-	7471B	PPB	03 Apr 2018 19:30:56	U	SMPL
J2183-06 - 1	J2183-05MS	30178	7.3732	-	7471B	PPB	03 Apr 2018 19:36:05	U	SMPL
J2183-07 - 1	J2183-05MSD	27946	6.8294	-	7471B	PPB	03 Apr 2018 19:38:23	U	SMPL
J2183-08 - 1	ES-4-DWS-BOTTOM@5.5	1921	0.4886	-	7471B	PPB	03 Apr 2018 19:40:41	U	SMPL
CCV45 - 1	CCV45	20581	5.0350	-	7471B	PPB	03 Apr 2018 19:42:59	U	SMPL
CCB45 - 1	CCB45	-59	0.0062	-	7471B	PPB	03 Apr 2018 19:45:16	U	SMPL
J2183-09 - 1	ES-4-DWS-2-SW@5	3542	0.8836	-	7471B	PPB	03 Apr 2018 19:47:33	U	SMPL
J2183-10 - 1	ES-4-DWS-2-NW@5	3198	0.7998	-	7471B	PPB	03 Apr 2018 19:49:50	U	SMPL
J2184-01 - 1	SB-1-1(6-8)	789	0.2129	-	7471B	PPB	03 Apr 2018 19:52:07	U	SMPL
J2184-02 - 1	SB-1-2(10-12)	115	0.0486	-	7471B	PPB	03 Apr 2018 19:54:24	U	SMPL
J2184-03 - 1	SB-2-1(1-3)	1295	0.3361	-	7471B	PPB	03 Apr 2018 19:56:40	U	SMPL
J2184-04 - 1	SB-2-2(9-11)	306	0.0952	-	7471B	PPB	03 Apr 2018 19:58:57	U	SMPL
J2184-05 - 1	SB-3-1(1-3)	8198	2.0180	-	7471B	PPB	03 Apr 2018 20:01:14	U	SMPL
J2184-06 - 1	SB-3-2(9-11)	485	0.1388	-	7471B	PPB	03 Apr 2018 20:03:31	U	SMPL
CCV46 - 1	CCV46	20936	5.1214	-	7471B	PPB	03 Apr 2018 20:05:48	U	SMPL
CCB46 - 1	CCB46	-82	0.0006	-	7471B	PPB	03 Apr 2018 20:08:06	U	SMPL
J2169-01LX5 - 1		144	0.0557	-	7471B	PPB	03 Apr 2018 20:10:22	U	SMPL

LB94342 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 03 Apr 2018 16:06:23

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
J2169-01A - 1		17591	4.3065		-7471B	PPB	03 Apr 2018 20:12:39	U	SMPL
J2182-12LX5 - 1		178	0.0640		-7471B	PPB	03 Apr 2018 20:14:56	U	SMPL
J2182-12A - 1		17933	4.3898		-7471B	PPB	03 Apr 2018 20:17:13	U	SMPL
J2183-05LX5 - 1		2269	0.5734		-7471B	PPB	03 Apr 2018 20:19:30	U	SMPL
J2183-05A - 1		28849	7.0494		-7471B	PPB	03 Apr 2018 20:21:47	U	SMPL
J2175-01DLX5 - 1		12564	3.0817		-7471B	PPB	03 Apr 2018 20:24:04	U	SMPL
CCV47 - 1	CCV47	20484	5.0113		-7471B	PPB	03 Apr 2018 20:26:22	U	SMPL
CCB47 - 1	CCB47	-60	0.0060		-7471B	PPB	03 Apr 2018 20:28:39	U	SMPL