



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB94346

Slope.: 97.5

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			18.2	4.01	04/03/2018	15:15
2	CAL2	1	Water			18.2	7.00	04/03/2018	15:16
3	CAL3	1	Water			18.3	10.02	04/03/2018	15:18
4	ICV	1	Water			18.5	7.00	04/03/2018	15:19
5	CCV1	1	Water			18.4	2.01	04/03/2018	15:20
6	J2197-06	1	Water			19.1	7.22	04/03/2018	15:21
7	J2197-06DUP	1	Water			19.3	7.23	04/03/2018	15:23
8	CCV2	1	Water			19.6	12.01	04/03/2018	15:25

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph j2197

WorkList ID : 110367

Date : 4/3/2018 3:10:31 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/03/2018	Water	J2197-06	-A pH	Cool 4 deg C	PSEG01	J63	DRUM-WATER	04/03/2018	9040C

Date/Time 04/03/18 3:10 PM
 Received by: SO
 Relinquished by: CP

Date/Time 04/03/18 4:00 PM
 Received by: CP
 Relinquished by: SO