

=====

Test results

Aquakem 7.2AQ1

2094357
Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

4/4/2018 10:34

Test: Total CN mg/L

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	94.343	0.0	0.094	
ICB1	0.388	0.0	0.004	
CCV1	241.820	0.0	0.235	
CCB1	0.264	0.0	0.004	
PB107952BL	0.271	0.0	0.004	
PB107952BS	96.783	0.0	0.096	
MIDPB107952	241.593	0.0	0.235	
J2133-10	1.842	0.0	0.005	
J2133-11	8.073	0.0	0.011	
J2133-12	1.612	0.0	0.005	
CCV2	249.404	0.0	0.242	
CCB2	-0.157	0.0	0.004	

N	12
Mean	78.020
SD	106.3189
CV%	136.27

AS

Aquakem v. 7.2AQ1

Results from time period:

Wed Apr 04 10:31:29 2018

Wed Apr 04 10:34:17 2018

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	-0.2232	µg/l	4/4/2018 8:21:16	
5.0PPBCN	A	Total CN	P	4.6613	µg/l	4/4/2018 8:21:17	
10PPBCN	A	Total CN	P	9.8335	µg/l	4/4/2018 8:21:18	
50PPBCN	A	Total CN	P	50.1169	µg/l	4/4/2018 8:21:19	
100PPBCN	A	Total CN	P	98.6672	µg/l	4/4/2018 8:21:20	
250PPBCN	A	Total CN	P	253.3654	µg/l	4/4/2018 8:21:21	
500PPBCN	A	Total CN	P	498.5789	µg/l	4/4/2018 8:21:22	
ICV1	S	Total CN	P	94.3428	µg/l	4/4/2018 10:31:29	
ICB1	S	Total CN	P	0.3881	µg/l	4/4/2018 10:31:30	
CCV1	S	Total CN	P	241.8203	µg/l	4/4/2018 10:31:31	
CCB1	S	Total CN	P	0.2644	µg/l	4/4/2018 10:31:32	
PB107952BL	S	Total CN	P	0.2707	µg/l	4/4/2018 10:31:33	
PB107952BS	S	Total CN	P	96.7828	µg/l	4/4/2018 10:31:34	
MIDPB107952	S	Total CN	P	241.5932	µg/l	4/4/2018 10:31:35	
J2133-10	S	Total CN	P	1.8423	µg/l	4/4/2018 10:31:36	
J2133-11	S	Total CN	P	8.0731	µg/l	4/4/2018 10:31:37	
J2133-12	S	Total CN	P	1.6121	µg/l	4/4/2018 10:31:38	
CCV2	S	Total CN	P	249.4041	µg/l	4/4/2018 10:31:39	
CCB2	S	Total CN	P	-0.1567	µg/l	4/4/2018 10:34:17	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

4/4/2018 8:21


Test Total CN

Accepted 4/4/2018 8:21

Factor 1045
Bias 0.004

Coeff. of det. 0.999925

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	-0.2232	0.0000	
2	5.0PPBCN	0.008	4.6613	5.0000	
3	10PPBCN	0.013	9.8335	10.0000	
4	50PPBCN	0.052	50.1169	50.0000	
5	100PPBCN	0.098	98.6672	100.0000	
6	250PPBCN	0.246	253.3654	250.0000	
7	500PPBCN	0.481	498.5789	500.0000	