



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB94361

Slope.: 98.7

BalanceID: WC SC-4

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.2	4.01	04/03/2018	17:35
2	CAL2	1	Water			7.0	21.2	04/03/2018	17:36
3	CAL3	1	Water			10.0	21.2	04/03/2018	17:38
4	ICV	1	Water			21.4	7.01	04/03/2018	17:39
5	CCV1	1	Water			21.5	2.01	04/03/2018	17:40
6	J2151-02	1	Solid	20.02	20	22.1	8.87	04/03/2018	17:41
7	J2151-02DUP	1	Solid	20.03	20	22.3	8.88	04/03/2018	17:42
8	J2151-04	1	Solid	20.01	20	22.3	9.49	04/03/2018	17:44
9	J2151-06	1	Solid	20.05	20	21.6	9.25	04/03/2018	17:45
10	J2166-02	1	Solid	20.04	20	22.2	8.56	04/03/2018	17:46
11	J2197-03	1	Solid	20.02	20	22.5	8.10	04/03/2018	17:47
12	J2197-05	1	Solid	20.03	20	22.2	8.27	04/03/2018	17:49
13	J2199-02	1	Solid	20.05	20	22.3	7.97	04/03/2018	17:50
14	J2207-06	1	Solid	20.02	20	22.1	9.46	04/03/2018	17:52
15	J2209-02	1	Solid	20.06	20	22.3	6.74	04/03/2018	18:00
16	CCV2	1	Water			22.1	12.01	04/03/2018	18:02
17	J2211-01	1	Solid	20.03	20	22.5	5.75	04/03/2018	18:03
18	J2211-03	1	Solid	20.02	20	22.3	7.32	04/03/2018	18:05
19	J2212-06	1	Solid	20.01	20	20.9	6.40	04/03/2018	18:06
20	J2212-12	1	Solid	20.02	20	21.3	6.68	04/03/2018	18:09

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	J2213-01	1	Solid	20.01	20	21.5	8.55	04/03/2018	18:10
22	CCV3	1	Water			21.0	2.01	04/03/2018	18:11

2B 94361

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity j2192 WorkList ID : 110364 Date : 4/3/2018 3:06:16 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/03/2018	Solid	J2151-02	Corrosivity	Cool 4 deg C	PSEG05	J63	AU-06-040318-A	04/03/2018	9045C
04/03/2018	Solid	J2151-04	Corrosivity	Cool 4 deg C	PSEG05	J63	AU-06-040318-B	04/03/2018	9045C
04/03/2018	Solid	J2151-06	Corrosivity	Cool 4 deg C	PSEG05	J63	AU-06-040318-C	04/03/2018	9045C
04/03/2018	Solid	J2166-02	Corrosivity	Cool 4 deg C	PSEG05	J63	TR-05-040318	04/03/2018	9045C
04/03/2018	Solid	J2197-03	Corrosivity	Cool 4 deg C	PSEG01	J63	DRUM-STONES	04/03/2018	9045C
04/03/2018	Solid	J2197-05	Corrosivity	Cool 4 deg C	PSEG01	J63	SP	04/03/2018	9045C
04/03/2018	Solid	J2199-02	Corrosivity	Cool 4 deg C	PSEG01	J63	S6-A	04/03/2018	9045C
04/03/2018	Solid	J2207-06	Corrosivity	Cool 4 deg C	PSEG01	J63	SP-1	04/03/2018	9045C
04/03/2018	Solid	J2209-02	Corrosivity	Cool 4 deg C	PSEG02	M11	OILY-DEBRIS	04/03/2018	9045C
04/03/2018	Solid	J2211-01	Corrosivity	Cool 4 deg C	PSEG02	M12	DRUM-1-3-4-5-6	04/03/2018	9045C
04/03/2018	Solid	J2211-03	Corrosivity	Cool 4 deg C	PSEG02	M12	DRUM-9	04/03/2018	9045C
04/03/2018	Solid	J2212-06	Corrosivity	Cool 4 deg C	PSEG01	M11	SP-1-A	04/03/2018	9045C
04/03/2018	Solid	J2212-12	Corrosivity	Cool 4 deg C	PSEG01	M11	SP-1-B	04/03/2018	9045C
04/03/2018	Solid	J2213-01	Corrosivity	Cool 4 deg C	PSEG01	M13	DRUM-1	04/03/2018	9045C

Date/Time 04.03.18 3:10 PM
 Received by: SO
 Relinquished by: SO

Date/Time 04.03.18 3:06 PM
 Received by: SO
 Relinquished by: SO