


ISM02.4

LB 94378
INSTRUMENT ID: CV1

Linear

 μ Abs.:

42877



A= 0.0000e+000

B= 2.3535e-004

C= 9.5683e-003

Rho= 0.9997472

Accept=Accepted

glope
y = 1012x - 42877

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	g/d
0.00	0.000	0.019	0.019	40	0.000	40					1
0.2	0.200	0.223	0.023	905	0.0 %	905					12
2.5	2.500	2.472	-0.028	10464	0.0 %	10464					-1
5.0	5.000	5.050	0.050	21416	0.0 %	21416					1
7.5	7.500	7.336	-0.164	31128	0.0 %	31128					-2
10.0	10.000	10.101	0.101	42877	0.0 %	42877					1

203
4/4/18

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	40	-		0 ISM02.4	PPB	4/4/2018 16:39	Std
	0.2 S0.2	905	-		0.2 ISM02.4	PPB	4/4/2018 16:42	Std
	2.5 S2.5	10464	-		2.5 ISM02.4	PPB	4/4/2018 16:44	Std
	5 S5	21416	-		5 ISM02.4	PPB	4/4/2018 16:46	Std
	7.5 S7.5	31128	-		7.5 ISM02.4	PPB	4/4/2018 16:48	Std
	10 S10	42877	-		10 ISM02.4	PPB	4/4/2018 16:51	Std
ICV	ICV	16277	3.8404	-	ISM02.4	PPB	4/4/2018 16:57	SMPL
ICB	ICB	-69	-0.0067	-	ISM02.4	PPB	4/4/2018 16:59	SMPL
CCV051	CCV051	21018	4.9562	-	ISM02.4	PPB	4/4/2018 17:01	SMPL
CCB051	CCB051	-107	-0.0156	-	ISM02.4	PPB	4/4/2018 17:03	SMPL
PB107972BL	PBW003	19	0.014	-	ISM02.4	PPB	4/4/2018 17:06	SMPL
J2083-01	SY-6-20180326	24	0.0152	-	ISM02.4	PPB	4/4/2018 17:08	SMPL
J2083-02	SY-3DD-20180326	25	0.0155	-	ISM02.4	PPB	4/4/2018 17:10	SMPL
J2083-04	SY-6-20180326	131	0.0404	-	ISM02.4	PPB	4/4/2018 17:13	SMPL
J2083-05	SY-3DD-20180326	143	0.0432	-	ISM02.4	PPB	4/4/2018 17:15	SMPL
J2116-01	SY-3D-20180327	652	0.163	-	ISM02.4	PPB	4/4/2018 17:17	SMPL
J2116-02	J2116-01S	4779	1.1343	-	ISM02.4	PPB	4/4/2018 17:19	SMPL
J2116-03	J2116-01D	597	0.1501	-	ISM02.4	PPB	4/4/2018 17:26	SMPL
J2116-04	SY-3-20180327	223	0.0621	-	ISM02.4	PPB	4/4/2018 17:29	SMPL
J2116-05	SY-2R-20180327	44	0.0199	-	ISM02.4	PPB	4/4/2018 17:31	SMPL
J2116-06	SY-5-20180327	211	0.0592	-	ISM02.4	PPB	4/4/2018 17:33	SMPL
J2116-07	SY-2D-20180327	65	0.0249	-	ISM02.4	PPB	4/4/2018 17:35	SMPL
J2116-08	SY-3D-20180327	124	0.0388	-	ISM02.4	PPB	4/4/2018 17:38	SMPL
J2116-09	J2116-08S	4539	1.0778	-	ISM02.4	PPB	4/4/2018 17:40	SMPL
J2116-10	J2116-08D	58	0.0232	-	ISM02.4	PPB	4/4/2018 17:42	SMPL
J2116-11	SY-3-20180327	71	0.0263	-	ISM02.4	PPB	4/4/2018 17:44	SMPL
J2116-12	SY-2R-20180327	34	0.0176	-	ISM02.4	PPB	4/4/2018 17:47	SMPL
J2116-13	SY-5-20180327	124	0.0388	-	ISM02.4	PPB	4/4/2018 17:49	SMPL
J2116-14	SY-2D-20180327	24	0.0152	-	ISM02.4	PPB	4/4/2018 17:51	SMPL
PB107973BL	PBW003	150	0.0449	-	ISM02.4	PPB	4/4/2018 17:54	SMPL
CCV052	CCV052	20564	4.8493	-	ISM02.4	PPB	4/4/2018 17:56	SMPL
CCB052	CCB052	-49	-0.002	-	ISM02.4	PPB	4/4/2018 17:58	SMPL
J2136-01	PK-10D-20180328	158	0.0468	-	ISM02.4	PPB	4/4/2018 18:00	SMPL
J2136-02	PK-10S-20180328	128	0.0397	-	ISM02.4	PPB	4/4/2018 18:03	SMPL
J2136-03	PK-10I-20180328	1810	0.4356	-	ISM02.4	PPB	4/4/2018 18:05	SMPL
J2136-03D	PK-10I-20180328D	1783	0.4292	-	ISM02.4	PPB	4/4/2018 18:07	SMPL
J2136-03S	PK-10I-20180328S	5839	1.3838	-	ISM02.4	PPB	4/4/2018 18:09	SMPL
J2136-08	PK-10D-20180328	28	0.0162	-	ISM02.4	PPB	4/4/2018 18:12	SMPL
J2136-09	PK-10S-20180328	103	0.0338	-	ISM02.4	PPB	4/4/2018 18:14	SMPL
J2136-10	PK-10I-20180328	1517	0.3666	-	ISM02.4	PPB	4/4/2018 18:16	SMPL
CCV053	CCV053	20716	4.8851	-	ISM02.4	PPB	4/4/2018 18:19	SMPL
CCB053	CCB053	-141	-0.0236	-	ISM02.4	PPB	4/4/2018 18:21	SMPL