

7471B


LB 94384

INSTRUMENT ID: CV1

Linear

 μ Abs.:

42461



A= 0.0000e+000

B= 2.3412e-004

C= -1.2974e-002

Rho= 0.9998286

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.028	0.028	176	0.000	176				
0.2	0.200	0.186	-0.014	848	0.0 %	848				
2.5	2.500	2.475	-0.025	10625	0.0 %	10625				
5.0	5.000	4.948	-0.052	21188	0.0 %	21188				
7.5	7.500	7.636	0.136	32670	0.0 %	32670				
10	10.000	9.928	-0.072	42461	0.0 %	42461				

LB94384 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 04 Apr 2018 18:47:13

Sample ID	Extended ID	μ Abs.	Conc.	Std Cond	Method	Units	Date	Type	Type
0.0 - 1	S0	176	-	0.0007471B	PPB		04 Apr 2018 18:48:40	S	Std
0.2 - 1	S0.2	848	-	0.20007471B	PPB		04 Apr 2018 18:50:58	S	Std
2.5 - 1	S2.5	10625	-	2.50007471B	PPB		04 Apr 2018 18:53:16	S	Std
5.0 - 1	S5	21188	-	5.00007471B	PPB		04 Apr 2018 18:55:33	S	Std
7.5 - 1	S7.5	32670	-	7.50007471B	PPB		04 Apr 2018 18:57:50	S	Std
10 - 1		42461	-	10.00007471B	PPB		04 Apr 2018 19:02:54	S	Std
ICV76 - 1	ICV76	15933	3.7173	-7471B	PPB		04 Apr 2018 19:05:30	U	SMPL
ICB76 - 1	ICB76	-105	-0.0376	-7471B	PPB		04 Apr 2018 19:07:45	U	SMPL
CCV54 - 1	CCV54	21238	4.9593	-7471B	PPB		04 Apr 2018 19:10:02	U	SMPL
CCB54 - 1	CCB54	-85	-0.0329	-7471B	PPB		04 Apr 2018 19:12:17	U	SMPL
CRA - 1	CRA	795	0.1732	-7471B	PPB		04 Apr 2018 19:14:34	U	SMPL
HighStd - 1	HighStd	42771	10.0007	-7471B	PPB		04 Apr 2018 19:16:49	U	SMPL
ChkStd - 1	ChkStd	30338	7.0899	-7471B	PPB		04 Apr 2018 19:19:06	U	SMPL
PB107981BL - 1	PBS	1	-0.0127	-7471B	PPB		04 Apr 2018 19:23:57	U	SMPL
PB107981BS - 1	LCSS	15513	3.6190	-7471B	PPB		04 Apr 2018 19:26:13	U	SMPL
J2197-01 - 1	DRUM-STONES	658	0.1411	-7471B	PPB		04 Apr 2018 19:28:30	U	SMPL
J2197-04 - 1	SP	1884	0.4281	-7471B	PPB		04 Apr 2018 19:30:47	U	SMPL
J2199-01 - 1	S6-A	3944	0.9104	-7471B	PPB		04 Apr 2018 19:33:05	U	SMPL
J2202-01 - 1	FK-GF-02-014-A	2164	0.4937	-7471B	PPB		04 Apr 2018 19:35:22	U	SMPL
J2202-03 - 1	FK-GF-02-014-B	1834	0.4164	-7471B	PPB		04 Apr 2018 19:37:38	U	SMPL
CCV55 - 1	CCV55	21224	4.9561	-7471B	PPB		04 Apr 2018 19:39:54	U	SMPL
CCB55 - 1	CCB55	-68	-0.0289	-7471B	PPB		04 Apr 2018 19:42:09	U	SMPL
J2203-01 - 1	FK-BPM-01-D02-D02A-E02-E02A	1933	0.4396	-7471B	PPB		04 Apr 2018 19:44:25	U	SMPL
J2203-01DUP - 1	FK-BPM-01-D02-D02A-E02-E02ADUP	1652	0.3738	-7471B	PPB		04 Apr 2018 19:46:41	U	SMPL
J2203-01MS - 1	FK-BPM-01-D02-D02A-E02-E02AMS	18301	4.2717	-7471B	PPB		04 Apr 2018 19:48:56	U	SMPL
J2203-01MSD - 1	FK-BPM-01-D02-D02A-E02-E02AMSD	17906	4.1792	-7471B	PPB		04 Apr 2018 19:51:13	U	SMPL
J2204-01 - 1	FK-BPM-01-F02-F02A-G02-G02A	1952	0.4440	-7471B	PPB		04 Apr 2018 19:53:30	U	SMPL
J2207-01 - 1	SP-1	29638	6.9260	-7471B	PPB		04 Apr 2018 19:55:46	U	SMPL
J2209-01 - 1	TRANSFORMER-OIL	-140	-0.0458	-7471B	PPB		04 Apr 2018 19:58:04	U	SMPL
J2211-02 - 1	DRUM-2-7-8	99	0.0102	-7471B	PPB		04 Apr 2018 20:00:23	U	SMPL
J2212-01 - 1	SP-1-A	2177	0.4967	-7471B	PPB		04 Apr 2018 20:02:41	U	SMPL
J2212-07 - 1	SP-1-B	1670	0.3780	-7471B	PPB		04 Apr 2018 20:04:57	U	SMPL
CCV56 - 1	CCV56	20526	4.7926	-7471B	PPB		04 Apr 2018 20:07:13	U	SMPL
CCB56 - 1	CCB56	-4	-0.0139	-7471B	PPB		04 Apr 2018 20:09:30	U	SMPL
J2203-01LX5 - 1		430	0.0877	-7471B	PPB		04 Apr 2018 20:11:46	U	SMPL
J2203-01A - 1		18688	4.3623	-7471B	PPB		04 Apr 2018 20:14:03	U	SMPL
CCV57 - 1	CCV57	20669	4.8261	-7471B	PPB		04 Apr 2018 20:16:19	U	SMPL
CCB57 - 1	CCB57	3	-0.0123	-7471B	PPB		04 Apr 2018 20:18:36	U	SMPL