



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB94389

Slope.: 97.5

BalanceID: WC SC-4

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.2	4.01	04/04/2018	17:26
2	CAL2	1	Water			20.2	7.01	04/04/2018	17:28
3	CAL3	1	Water			20.3	10.03	04/04/2018	17:30
4	ICV	1	Water			20.3	6.99	04/04/2018	17:31
5	CCV1	1	Water			20.5	2.01	04/04/2018	17:33
6	J2153-02	1	Solid	20.02	20	21.2	9.36	04/04/2018	17:34
7	J2153-02DUP	1	Solid	20.04	20	21.1	9.37	04/04/2018	17:35
8	J2153-04	1	Solid	20.03	20	21.5	9.12	04/04/2018	17:36
9	J2153-06	1	Solid	20.05	20	21.3	9.11	04/04/2018	17:37
10	J2214-03	1	Solid	20.02	20	22.5	8.96	04/04/2018	17:38
11	J2223-04	1	Solid	20.01	20	21.6	5.72	04/04/2018	17:40
12	J2228-03	1	Solid	20.03	20	21.5	6.97	04/04/2018	17:41
13	J2231-03	1	Solid	20.01	20	21.1	6.18	04/04/2018	17:43
14	J2232-03	1	Solid	20.03	20	21.5	5.51	04/04/2018	17:45
15	J2237-02	1	Solid	20.03	20	21.2	8.31	04/04/2018	17:46
16	CCV2	1	Water			20.4	12.01	04/04/2018	17:48

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity j2232

WorkList ID : 110427

Date : 4/4/2018 4:03:20 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/04/2018	Solid	J2153-02	Corrosivity	Cool 4 deg C	PSEG05	M32	HR-06-040418-A	04/04/2018	9045C
04/04/2018	Solid	J2153-04	Corrosivity	Cool 4 deg C	PSEG05	M32	HR-06-040418-B	04/04/2018	9045C
04/04/2018	Solid	J2153-06	Corrosivity	Cool 4 deg C	PSEG05	M32	HR-06-040418-C	04/04/2018	9045C
04/04/2018	Solid	J2214-03	Corrosivity	Cool 4 deg C	PSEG01	M21	HH-STONE	04/04/2018	9045C
04/04/2018	Solid	J2223-04	Corrosivity	Cool 4 deg C	PSEG01	M23	VW	04/04/2018	9045C
04/04/2018	Solid	J2228-03	Corrosivity	Cool 4 deg C	PSEG01	M23	11-WC	04/04/2018	9045C
04/04/2018	Solid	J2231-03	Corrosivity	Cool 4 deg C	PSEG01	M32	42ND-STREET	04/04/2018	9045C
04/04/2018	Solid	J2232-03	Corrosivity	Cool 4 deg C	PSEG01	M23	LA	04/04/2018	9045C
04/04/2018	Solid	J2237-02	Corrosivity	Cool 4 deg C	BRIN05	L22	SC-1	04/04/2018	9045C

Date/Time 4.4.18 4:05 PM

Received by: Jg

Relinquished by: CR

Date/Time

Received by:

Relinquished by:

4.4.18 6:30 PM

Jg

Jh