


7470A

LB94396
INSTRUMENT ID: CV1

Linear

 μ Abs.:

33343



A= 0.0000e+000

B= 2.9934e-004

C= 4.9302e-002

Rho= 0.9997842

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.057	0.057	25	0.000	25				
0.2	0.200	0.227	0.027	592	0.0 %	592				
2.5	2.500	2.478	-0.022	8113	0.0 %	8113				
5.0	5.000	4.841	-0.159	16007	0.0 %	16007				
7.5	7.500	7.568	0.068	25117	0.0 %	25117				
10.0	10.000	10.030	0.030	33343	0.0 %	33343				

LB94396 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 05 Apr 2018 11:06:05

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	25	-	0.0007470A	PPB		05 Apr 2018 11:10:06	S	Std
0.2 - 1	50.2	592	-	0.20007470A	PPB		05 Apr 2018 11:12:15	S	Std
2.5 - 1	52.5	8113	-	2.50007470A	PPB		05 Apr 2018 11:14:23	S	Std
5.0 - 1	55	16007	-	5.00007470A	PPB		05 Apr 2018 11:16:31	S	Std
7.5 - 1	57.5	25117	-	7.50007470A	PPB		05 Apr 2018 11:18:39	S	Std
10.0 - 1	510	33343	-	10.00007470A	PPB		05 Apr 2018 11:20:46	S	Std
ICV77 - 1	ICV77	13067	3.9608	-7470A	PPB		05 Apr 2018 11:24:03	U	SMPL
ICB77 - 1	ICB77	-151	0.0041	-7470A	PPB		05 Apr 2018 11:26:09	U	SMPL
CCV58 - 1	CCV58	17012	5.1417	-7470A	PPB		05 Apr 2018 11:28:16	U	SMPL
CCB58 - 1	CCB58	-143	0.0065	-7470A	PPB		05 Apr 2018 11:30:23	U	SMPL
CRA - 1	CRA	583	0.2238	-7470A	PPB		05 Apr 2018 11:32:30	U	SMPL
HighStd - 1	HighStd	33925	10.2044	-7470A	PPB		05 Apr 2018 11:34:37	U	SMPL
ChkStd - 1	ChkStd	23676	7.1365	-7470A	PPB		05 Apr 2018 11:36:44	U	SMPL
PB107989BL - 1	PBW	-253	-0.0264	-7470A	PPB		05 Apr 2018 11:38:53	U	SMPL
PB107989BS - 1	LCSW	12021	3.6477	-7470A	PPB		05 Apr 2018 11:41:02	U	SMPL
PB107989TB - 1	PB107956TB	-104	0.0182	-7470A	PPB		05 Apr 2018 11:43:10	U	SMPL
J2151-02 - 1	AU-06-040318-A	-4	0.0481	-7470A	PPB		05 Apr 2018 11:45:18	U	SMPL
J2151-02DUP - 1	AU-06-040318-ADUP	-1	0.0490	-7470A	PPB		05 Apr 2018 11:47:26	U	SMPL
J2151-02MS - 1	AU-06-040318-AMS	12546	3.8048	-7470A	PPB		05 Apr 2018 11:49:35	U	SMPL
J2151-02MSD - 1	AU-06-040318-AMSD	12461	3.7794	-7470A	PPB		05 Apr 2018 11:51:42	U	SMPL
CCV59 - 1	CCV59	16959	5.1258	-7470A	PPB		05 Apr 2018 11:53:48	U	SMPL
CCB59 - 1	CCB59	-152	0.0038	-7470A	PPB		05 Apr 2018 11:55:55	U	SMPL
J2151-04 - 1	AU-06-040318-B	7	0.0514	-7470A	PPB		05 Apr 2018 11:58:02	U	SMPL
J2151-06 - 1	AU-06-040318-C	0	0.0493	-7470A	PPB		05 Apr 2018 12:00:09	U	SMPL
J2166-02 - 1	TR-05-040318	25	0.0568	-7470A	PPB		05 Apr 2018 12:02:16	U	SMPL
J2192-02 - 1	CHRT25895- SOIL- CONCRETE	-12	0.0457	-7470A	PPB		05 Apr 2018 12:04:23	U	SMPL
J2197-03 - 1	DRUM-STONES	-1	0.0490	-7470A	PPB		05 Apr 2018 12:06:31	U	SMPL
J2197-05 - 1	SP	2	0.0499	-7470A	PPB		05 Apr 2018 12:08:38	U	SMPL
J2199-02 - 1	S6-A	1	0.0496	-7470A	PPB		05 Apr 2018 12:10:46	U	SMPL
J2207-06 - 1	SP-1	85	0.0747	-7470A	PPB		05 Apr 2018 12:12:55	U	SMPL
J2209-02 - 1	OILY-DEBRIS	28	0.0577	-7470A	PPB		05 Apr 2018 12:15:03	U	SMPL
J2211-01 - 1	DRUM-1-3-4-5-6	11	0.0526	-7470A	PPB		05 Apr 2018 12:17:10	U	SMPL
CCV60 - 1	CCV60	16916	5.1129	-7470A	PPB		05 Apr 2018 12:19:17	U	SMPL
CCB60 - 1	CCB60	-142	0.0068	-7470A	PPB		05 Apr 2018 12:21:24	U	SMPL
J2211-03 - 1	DRUM-9	-25	0.0418	-7470A	PPB		05 Apr 2018 12:23:32	U	SMPL
J2212-06 - 1	SP-1-A	-2	0.0487	-7470A	PPB		05 Apr 2018 12:25:39	U	SMPL
J2212-12 - 1	SP-1-B	17	0.0544	-7470A	PPB		05 Apr 2018 12:27:47	U	SMPL
J2213-01 - 1	DRUM-1	25	0.0568	-7470A	PPB		05 Apr 2018 12:29:54	U	SMPL
PB107992BL - 1	PBW	25	0.0568	-7470A	PPB		05 Apr 2018 12:32:01	U	SMPL
PB107992BS - 1	LCSW	12547	3.8051	-7470A	PPB		05 Apr 2018 12:34:09	U	SMPL
J2074-02 - 1	FB-032618	2	0.0499	-7470A	PPB		05 Apr 2018 12:36:16	U	SMPL
J2197-06 - 1	DRUM-WATER	437	0.1801	-7470A	PPB		05 Apr 2018 12:42:58	U	SMPL
J2197-06DUP - 1	DRUM-WATERDUP	369	0.1598	-7470A	PPB		05 Apr 2018 12:45:06	U	SMPL
J2197-06MS - 1	DRUM-WATERMS	14379	4.3535	-7470A	PPB		05 Apr 2018 12:47:13	U	SMPL
CCV61 - 1	CCV61	16512	4.9920	-7470A	PPB		05 Apr 2018 12:49:20	U	SMPL
CCB61 - 1	CCB61	-114	0.0152	-7470A	PPB		05 Apr 2018 12:51:28	U	SMPL
J2197-06MSD - 1	DRUM-WATERMSD	14190	4.2969	-7470A	PPB		05 Apr 2018 12:53:36	U	SMPL
PB107993BL - 1	PBW	-130	0.0104	-7470A	PPB		05 Apr 2018 12:55:43	U	SMPL
PB107993BS - 1	LCSW	11093	3.3699	-7470A	PPB		05 Apr 2018 12:57:51	U	SMPL
PB107993TB - 1	PB107963TB	-82	0.0248	-7470A	PPB		05 Apr 2018 12:59:58	U	SMPL
J2197-07 - 1	DRUM-WATER	-11	0.0460	-7470A	PPB		05 Apr 2018 13:02:05	U	SMPL
J2197-07DUP - 1	DRUM-WATERDUP	37	0.0604	-7470A	PPB		05 Apr 2018 13:04:12	U	SMPL
J2197-07MS - 1	DRUM-WATERMS	12197	3.7004	-7470A	PPB		05 Apr 2018 13:06:20	U	SMPL
J2197-07MSD - 1	DRUM-WATERMSD	12170	3.6923	-7470A	PPB		05 Apr 2018 13:08:29	U	SMPL
CCV62 - 1	CCV62	16264	4.9178	-7470A	PPB		05 Apr 2018 13:10:37	U	SMPL
CCB62 - 1	CCB62	-197	-0.0097	-7470A	PPB		05 Apr 2018 13:12:45	U	SMPL
J2151-02LX5 - 1		-27	0.0412	-7470A	PPB		05 Apr 2018 13:14:53	U	SMPL
J2151-02A - 1		13754	4.1664	-7470A	PPB		05 Apr 2018 13:17:00	U	SMPL
J2197-06LX5 - 1		24	0.0565	-7470A	PPB		05 Apr 2018 13:19:07	U	SMPL
J2197-06A - 1		14863	4.4984	-7470A	PPB		05 Apr 2018 13:21:14	U	SMPL
J2197-07LX5 - 1		122	0.0858	-7470A	PPB		05 Apr 2018 13:23:21	U	SMPL
J2197-07A - 1		13388	4.0569	-7470A	PPB		05 Apr 2018 13:25:30	U	SMPL
CCV63 - 1	CCV63	16786	5.0740	-7470A	PPB		05 Apr 2018 13:27:37	U	SMPL
CCB63 - 1	CCB63	-176	-0.0034	-7470A	PPB		05 Apr 2018 13:29:45	U	SMPL