

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

4/5/2018 12:19

Test: Phenolics-mg/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.039	0.0	0.113	
ICB1	0.024	0.0	0.006	
CCV1	1.004	0.0	0.110	
CCB1	0.018	0.0	0.006	
PB107948BL	0.009	0.0	0.005	
PB107948BS	1.059	0.0	0.115	
J2133-01	0.025	0.0	0.007	
J2133-02	0.108	0.0	0.015	
J2133-03	0.022	0.0	0.006	
PB107950BL	-0.003	0.0	0.004	
PB107950BS	1.021	0.0	0.111	
J2133-07	0.017	0.0	0.006	
J2133-08	0.097	0.0	0.014	
J2133-09	0.017	0.0	0.006	
CCV2	1.000	0.0	0.109	
CCB2	-0.006	0.0	0.003	
J2215-01	0.007	0.0	0.005	
J2215-01DUP	0.010	0.0	0.005	
J2215-01MS	0.993	0.0	0.108	
J2215-01MSD	0.987	0.0	0.108	
J2215-02	0.007	0.0	0.005	
CCV3	0.982	0.0	0.107	
CCB3	0.007	0.0	0.005	

N	23
Mean	0.367
SD	0.4814
CV%	131.14

AS

Aquakem v. 7.2AQ1

Results from time period:

Thu Apr 05 11:27:41 2018

Thu Apr 05 12:14:02 2018

Sample Id	Sam/Ctr/c	Test short name	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0578	mg/l	4/5/2018 11:27:41	
0.1PPM	A	Phenolics-	P	0.0994	mg/l	4/5/2018 11:27:42	
0.25PPM	A	Phenolics-	P	0.2508	mg/l	4/5/2018 11:27:43	
0.50PPM	A	Phenolics-	P	0.5018	mg/l	4/5/2018 11:27:44	
1.0PPM	A	Phenolics-	P	0.9818	mg/l	4/5/2018 11:27:45	
2.0PPM	A	Phenolics-	P	2.0084	mg/l	4/5/2018 11:27:46	
ICV1	S	Phenolics-	P	1.0389	mg/l	4/5/2018 12:00:01	
ICB1	S	Phenolics-	P	0.0241	mg/l	4/5/2018 12:00:02	
CCV1	S	Phenolics-	P	1.004	mg/l	4/5/2018 12:00:03	
CCB1	S	Phenolics-	P	0.0183	mg/l	4/5/2018 12:00:04	
PB107948BL	S	Phenolics-	P	0.009	mg/l	4/5/2018 12:00:05	
PB107948BS	S	Phenolics-	P	1.0595	mg/l	4/5/2018 12:00:06	
J2133-01	S	Phenolics-	P	0.0253	mg/l	4/5/2018 12:00:08	
J2133-02	S	Phenolics-	P	0.1078	mg/l	4/5/2018 12:00:09	
J2133-03	S	Phenolics-	P	0.0218	mg/l	4/5/2018 12:00:10	
PB107950BL	S	Phenolics-	P	-0.0035	mg/l	4/5/2018 12:00:11	
PB107950BS	S	Phenolics-	P	1.0206	mg/l	4/5/2018 12:00:12	
J2133-07	S	Phenolics-	P	0.0169	mg/l	4/5/2018 12:10:03	
J2133-08	S	Phenolics-	P	0.0967	mg/l	4/5/2018 12:10:04	
J2133-09	S	Phenolics-	P	0.0174	mg/l	4/5/2018 12:10:05	
CCV2	S	Phenolics-	P	0.9996	mg/l	4/5/2018 12:10:06	
CCB2	S	Phenolics-	P	-0.0058	mg/l	4/5/2018 12:10:07	
J2215-01	S	Phenolics-	P	0.0072	mg/l	4/5/2018 12:10:08	
J2215-01DUP	S	Phenolics-	P	0.0097	mg/l	4/5/2018 12:10:09	
J2215-01MS	S	Phenolics-	P	0.9929	mg/l	4/5/2018 12:10:10	
J2215-01MSD	S	Phenolics-	P	0.9873	mg/l	4/5/2018 12:10:11	
J2215-02	S	Phenolics-	P	0.0072	mg/l	4/5/2018 12:10:12	
CCV3	S	Phenolics-	P	0.9815	mg/l	4/5/2018 12:10:13	
CCB3	S	Phenolics-	P	0.0071	mg/l	4/5/2018 12:10:14	

CHEMTECH
 284 Sheffield Street,
 Mountainside, NJ 07092
 Reviewed by : AS

4/5/2018 11:30


Test Phenolics-

Accepted 4/5/2018 11:30

Factor 9.497
 Bias 0.004

Coeff. of det. 0.999833

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors

1	Phenol-2	0.010	0.0578	0.0500	
2	Phenol-2	0.014	0.0994	0.1000	
3	Phenol-2	0.030	0.2508	0.2500	
4	Phenol-2	0.057	0.5018	0.5000	
5	Phenol-2	0.107	0.9818	1.0000	
6	Phenol-2	0.215	2.0084	2.0000	