

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092

Reviewed by : AS

4/5/2018 13:37

Test: Phenolics- *mg/d*

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.017	0.0	0.111	
ICB1	0.001	0.0	0.004	
CCV1	0.983	0.0	0.107	
CCB1	0.001	0.0	0.004	
PB107949BL	0.000	0.0	0.004	
PB107949BS	1.030	0.0	0.112	
J2133-11	0.122	0.0	0.017	
J2133-12	0.012	0.0	0.005	
J2133-10	0.013	0.0	0.005	
CCV2	1.027	0.0	0.112	
CCB2	0.004	0.0	0.004	

N	11
Mean	0.383
SD	0.5021
CV%	131.17

AS

Aquakem v. 7.2AQ1

Results from time period:

Thu Apr 05 12:46:20 2018

Thu Apr 05 13:36:04 2018

Sample Id	Sam/Ctr/c/	Test short name	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0578	mg/l	4/5/2018 11:27:41	
0.1PPM	A	Phenolics-	P	0.0994	mg/l	4/5/2018 11:27:42	
0.25PPM	A	Phenolics-	P	0.2508	mg/l	4/5/2018 11:27:43	
0.50PPM	A	Phenolics-	P	0.5018	mg/l	4/5/2018 11:27:44	
1.0PPM	A	Phenolics-	P	0.9818	mg/l	4/5/2018 11:27:45	
2.0PPM	A	Phenolics-	P	2.0084	mg/l	4/5/2018 11:27:46	
ICV1	S	Phenolics-	P	1.0172	mg/l	4/5/2018 12:46:20	
ICB1	S	Phenolics-	P	0.0013	mg/l	4/5/2018 12:46:21	
CCV1	S	Phenolics-	P	0.9831	mg/l	4/5/2018 12:46:22	
CCB1	S	Phenolics-	P	0.0006	mg/l	4/5/2018 12:46:23	
PB107949BL	S	Phenolics-	P	0	mg/l	4/5/2018 12:46:24	
PB107949BS	S	Phenolics-	P	1.0304	mg/l	4/5/2018 12:46:25	
J2133-11	S	Phenolics-	P	0.1222	mg/l	4/5/2018 13:11:40	
J2133-12	S	Phenolics-	P	0.0119	mg/l	4/5/2018 13:11:41	
J2133-10	S	Phenolics-	P	0.013	mg/l	4/5/2018 13:36:02	
CCV2	S	Phenolics-	P	1.0267	mg/l	4/5/2018 13:36:03	
CCB2	S	Phenolics-	P	0.0039	mg/l	4/5/2018 13:36:04	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

4/5/2018 11:30


Test Phenolics-

Accepted 4/5/2018 11:30

Factor 9.497
Bias 0.004

Coeff. of det. 0.999833

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.010	0.0578	0.0500	
2	Phenol-2	0.014	0.0994	0.1000	
3	Phenol-2	0.030	0.2508	0.2500	
4	Phenol-2	0.057	0.5018	0.5000	
5	Phenol-2	0.107	0.9818	1.0000	
6	Phenol-2	0.215	2.0084	2.0000	