

7471B


LB94405

INSTRUMENT ID: CVL

Linear

 μ Abs.:

40945



A= 0.0000e+000

B= 2.4250e-004

C= -1.0107e-003

Rho= 0.9996837

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.007	0.007	32	0.000	32				
0.2	0.200	0.207	0.007	856	0.0 %	856				
2.5	2.500	2.520	0.020	10397	0.0 %	10397				
5.0	5.000	4.869	-0.131	20082	0.0 %	20082				
7.5	7.500	7.670	0.170	31632	0.0 %	31632				
10	10.000	9.928	-0.072	40945	0.0 %	40945				

LB94405 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 05 Apr 2018 15:31:20

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	32	-	0.00007471B	PPB		05 Apr 2018 15:33:53	S	Std
0.2 - 1	50.2	856	-	0.20007471B	PPB		05 Apr 2018 15:36:10	S	Std
2.5 - 1	52.5	10397	-	2.50007471B	PPB		05 Apr 2018 15:38:28	S	Std
5.0 - 1	55	20082	-	5.00007471B	PPB		05 Apr 2018 15:40:45	S	Std
7.5 - 1	57.5	31632	-	7.50007471B	PPB		05 Apr 2018 15:43:03	S	Std
10 - 1	510	40945	-	10.00007471B	PPB		05 Apr 2018 15:48:20	S	Std
ICV78 - 1	ICV78	16218	3.9318	-7471B	PPB		05 Apr 2018 15:51:05	U	SMPL
ICB78 - 1	ICB78	-118	-0.0296	-7471B	PPB		05 Apr 2018 15:53:21	U	SMPL
CCV64 - 1	CCV64	20168	4.8897	-7471B	PPB		05 Apr 2018 15:55:37	U	SMPL
CCB64 - 1	CCB64	-125	-0.0313	-7471B	PPB		05 Apr 2018 15:57:53	U	SMPL
CRA - 1	CRA	841	0.2029	-7471B	PPB		05 Apr 2018 16:00:10	U	SMPL
HighStd - 1	HighStd	41681	10.1065	-7471B	PPB		05 Apr 2018 16:02:26	U	SMPL
ChkStd - 1	ChkStd	28193	6.8357	-7471B	PPB		05 Apr 2018 16:04:42	U	SMPL
PB108013BL - 1	PBS	-228	-0.0563	-7471B	PPB		05 Apr 2018 16:07:00	U	SMPL
PB108013BS - 1	LCSS	14303	3.4674	-7471B	PPB		05 Apr 2018 16:09:18	U	SMPL
J2151-01 - 1	AU-06-040318-A	3879	0.9396	-7471B	PPB		05 Apr 2018 16:11:35	U	SMPL
J2151-03 - 1	AU-06-040318-B	4375	1.0599	-7471B	PPB		05 Apr 2018 16:13:53	U	SMPL
J2151-05 - 1	AU-06-040318-C	1565	0.3785	-7471B	PPB		05 Apr 2018 16:16:10	U	SMPL
J2153-01 - 1	HR-06-040418-A	484	0.1164	-7471B	PPB		05 Apr 2018 16:18:28	U	SMPL
J2153-01DUP - 1	HR-06-040418-ADUP	458	0.1101	-7471B	PPB		05 Apr 2018 16:20:44	U	SMPL
CCV65 - 1	CCV65	19986	4.8455	-7471B	PPB		05 Apr 2018 16:23:00	U	SMPL
CCB65 - 1	CCB65	-94	-0.0238	-7471B	PPB		05 Apr 2018 16:25:16	U	SMPL
J2153-01MS - 1	HR-06-040418-AMS	15520	3.7625	-7471B	PPB		05 Apr 2018 16:27:33	U	SMPL
J2153-01MSD - 1	HR-06-040418-AMSD	15492	3.7557	-7471B	PPB		05 Apr 2018 16:29:48	U	SMPL
J2153-03 - 1	HR-06-040418-B	1343	0.3247	-7471B	PPB		05 Apr 2018 16:32:04	U	SMPL
J2153-05 - 1	HR-06-040418-C	3587	0.8688	-7471B	PPB		05 Apr 2018 16:34:20	U	SMPL
J2166-01 - 1	TR-05-040318	3258	0.7890	-7471B	PPB		05 Apr 2018 16:36:37	U	SMPL
J2214-01 - 1	HH-STONE	233	0.0555	-7471B	PPB		05 Apr 2018 16:38:53	U	SMPL
J2223-02 - 1	CV-RB47188RT	1470	0.3555	-7471B	PPB		05 Apr 2018 16:41:11	U	SMPL
J2228-01 - 1	ACV-102-20-ACV-43-30	2568	0.6217	-7471B	PPB		05 Apr 2018 16:43:29	U	SMPL
J2231-01 - 1	42ND-STREET	12332	2.9895	-7471B	PPB		05 Apr 2018 16:45:46	U	SMPL
J2232-01 - 1	CV-RB26667RT	6962	1.6873	-7471B	PPB		05 Apr 2018 16:48:03	U	SMPL
CCV66 - 1	CCV66	20155	4.8865	-7471B	PPB		05 Apr 2018 16:50:19	U	SMPL
CCB66 - 1	CCB66	-112	-0.0282	-7471B	PPB		05 Apr 2018 16:52:35	U	SMPL
J2237-01 - 1	SC-1	936	0.2260	-7471B	PPB		05 Apr 2018 16:54:52	U	SMPL
J2240-01 - 1	PF-01	380	0.0911	-7471B	PPB		05 Apr 2018 16:57:08	U	SMPL
PB108020BL - 1	PBS	0	-0.0010	-7471B	PPB		05 Apr 2018 16:59:24	U	SMPL
PB108020BS - 1	LCSS	14926	3.6185	-7471B	PPB		05 Apr 2018 17:04:23	U	SMPL
J2222-01 - 1	EP-03	4930	1.1945	-7471B	PPB		05 Apr 2018 17:06:40	U	SMPL
J2222-02 - 1	EP-04	2832	0.6857	-7471B	PPB		05 Apr 2018 17:08:56	U	SMPL
J2222-03 - 1	EP-05	554	0.1333	-7471B	PPB		05 Apr 2018 17:11:14	U	SMPL
J2222-04 - 1	EP-11	682	0.1644	-7471B	PPB		05 Apr 2018 17:13:31	U	SMPL
J2222-05 - 1	EP-12	448	0.1076	-7471B	PPB		05 Apr 2018 17:15:49	U	SMPL
J2222-06 - 1	EP-13	890	0.2148	-7471B	PPB		05 Apr 2018 17:18:05	U	SMPL
CCV67 - 1	CCV67	20597	4.9937	-7471B	PPB		05 Apr 2018 17:20:22	U	SMPL
CCB67 - 1	CCB67	-95	-0.0240	-7471B	PPB		05 Apr 2018 17:22:38	U	SMPL
J2222-07 - 1	EP-19	1003	0.2422	-7471B	PPB		05 Apr 2018 17:24:55	U	SMPL
J2222-08 - 1	EP-20	20083	4.8690	-7471B	PPB		05 Apr 2018 17:27:11	U	SMPL
J2222-09 - 1	EP-21	371	0.0890	-7471B	PPB		05 Apr 2018 17:29:28	U	SMPL
J2222-09DUP - 1	EP-21DUP	429	0.1030	-7471B	PPB		05 Apr 2018 17:31:44	U	SMPL
J2222-09MS - 1	EP-21MS	17953	4.3525	-7471B	PPB		05 Apr 2018 17:34:01	U	SMPL
J2222-09MSD - 1	EP-21MSD	17768	4.3077	-7471B	PPB		05 Apr 2018 17:36:18	U	SMPL
J2153-01LX5 - 1		10	0.0014	-7471B	PPB		05 Apr 2018 17:38:35	U	SMPL
J2153-01A - 1		17190	4.1675	-7471B	PPB		05 Apr 2018 17:40:52	U	SMPL
J2222-09LX5 - 1		29	0.0060	-7471B	PPB		05 Apr 2018 17:43:10	U	SMPL
J2222-09A - 1		17215	4.1736	-7471B	PPB		05 Apr 2018 17:45:27	U	SMPL
CCV68 - 1	CCV68	20365	4.9374	-7471B	PPB		05 Apr 2018 17:47:43	U	SMPL
CCB68 - 1	CCB68	-115	-0.0289	-7471B	PPB		05 Apr 2018 17:50:00	U	SMPL