



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB94406

Slope : 97.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			19.1	4.01	04/05/2018	13:35
2	CAL2	1	Water			19.2	7.00	04/05/2018	13:36
3	CAL3	1	Water			19.2	10.01	04/05/2018	13:37
4	ICV	1	Water			19.3	7.00	04/05/2018	13:39
5	CCV1	1	Water			19.2	2.02	04/05/2018	13:40
6	J2254-01	1	Solid	20.02	20	19.1	7.36	04/05/2018	13:41
7	J2254-01DUP	1	Solid	20.03	20	19.2	7.37	04/05/2018	13:43
8	CCV2	1	Water			19.3	12.02	04/05/2018	13:45

1894406

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph j2254

WorkList ID : 110474

Date : 4/5/2018 1:29:35 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/04/2018	Solid	J2254-01	pH	Cool 4 deg C	TULL01	L22	STOCKPILE	04/04/2018	9045C

Date/Time 04/05/18 11:30 AM
Received by: J2
Relinquished by: LO

Date/Time 04/05/18 3:00 PM
Received by: CP
Relinquished by: J2