



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Sweta

Run Number: LB94409

Slope.: 97.5

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			18.2	3.99	04/05/2018	15:00
2	CAL2	1	Water			18.2	7.01	04/05/2018	15:01
3	CAL3	1	Water			18.3	10.02	04/05/2018	15:02
4	ICV	1	Water			18.4	7.00	04/05/2018	15:03
5	CCV1	1	Water			18.4	2.01	04/05/2018	15:05
6	J2247-04	1	Water			17.8	6.65	04/05/2018	15:06
7	J2247-04DUP	1	Water			17.9	6.66	04/05/2018	15:08
8	CCV2	1	Water			18.1	12.01	04/05/2018	15:09

2894409

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph j2247

WorkList ID : 110480

Date : 4/5/2018 2:38:11 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/05/2018	Water	J2247-04	pH	Cool 4 deg C	PSEG02	M33	CG-1	04/05/2018	9040C

Date/Time 04-05-18 2:38 PM
Received by: JP
Relinquished by: CP

Date/Time 04-05-18 3:00 PM
Received by: CP
Relinquished by: JP