

7471B


LB94429

INSTRUMENT ID: CX1

Linear

 μ Abs.:

42314



A= 0.0000e+000

B= 2.3703e-004

C= 3.4520e-002

Rho= 0.9997758

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.043	0.043	36	0.000	36				
0.2	0.200	0.229	0.029	822	0.0 %	822				
2.5	2.500	2.517	0.017	10475	0.0 %	10475				
5.0	5.000	4.829	-0.171	20228	0.0 %	20228				
7.5	7.500	7.517	0.017	31569	0.0 %	31569				
10	10.000	10.064	0.064	42314	0.0 %	42314				

LB94429 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 06 Apr 2018 09:41:50					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	36	-	0.00007471B		PPB	06 Apr 2018 09:56:48	S	Std
0.2 - 1	80.2	822	-	0.20007471B		PPB	06 Apr 2018 09:59:06	S	Std
2.5 - 1	52.5	10475	-	2.50007471B		PPB	06 Apr 2018 10:01:23	S	Std
5.0 - 1	55	20228	-	5.00007471B		PPB	06 Apr 2018 10:03:40	S	Std
7.5 - 1	57.5	31569	-	7.50007471B		PPB	06 Apr 2018 10:05:58	S	Std
10 - 1		42314	-	10.00007471B		PPB	06 Apr 2018 10:08:15	S	Std
ICV80 - 1	ICV80	16454	3.9345	-7471B		PPB	06 Apr 2018 10:11:18	U	SMPL
ICB80 - 1	ICB80	-145	0.0002	-7471B		PPB	06 Apr 2018 10:13:34	U	SMPL
CCV76 - 1	CCV76	20519	4.8980	-7471B		PPB	06 Apr 2018 10:15:50	U	SMPL
CCB76 - 1	CCB76	-173	-0.0065	-7471B		PPB	06 Apr 2018 10:18:06	U	SMPL
CRA - 1	CRA	797	0.2234	-7471B		PPB	06 Apr 2018 10:20:22	U	SMPL
HighStd - 1	HighStd	42623	10.1372	-7471B		PPB	06 Apr 2018 10:22:38	U	SMPL
ChkStd - 1	ChkStd	29026	6.9144	-7471B		PPB	06 Apr 2018 10:24:54	U	SMPL
PB108045BL - 1	PBS	-260	-0.0271	-7471B		PPB	06 Apr 2018 10:27:13	U	SMPL
PB108045BS - 1	LCSS	14489	3.4688	-7471B		PPB	06 Apr 2018 10:29:32	U	SMPL
J2164-01 - 1	SA-01-040518	1808	0.4631	-7471B		PPB	06 Apr 2018 10:31:49	U	SMPL
J2165-01 - 1	SU-04-040518	2176	0.5503	-7471B		PPB	06 Apr 2018 10:34:06	U	SMPL
J2248-01 - 1	11944	365	0.1210	-7471B		PPB	06 Apr 2018 10:36:24	U	SMPL
J2248-03 - 1	12000	1639	0.4230	-7471B		PPB	06 Apr 2018 10:38:41	U	SMPL
J2248-05 - 1	RT3059	1980	0.5038	-7471B		PPB	06 Apr 2018 10:40:57	U	SMPL
CCV77 - 1	CCV77	20430	4.8769	-7471B		PPB	06 Apr 2018 10:43:14	U	SMPL
CCB77 - 1	CCB77	-97	0.0115	-7471B		PPB	06 Apr 2018 10:45:30	U	SMPL
J2254-01 - 1	STOCKPILE	20473	4.8871	-7471B		PPB	06 Apr 2018 10:47:46	U	SMPL
J2256-01 - 1	FK-BPM-02-004-ABCD	1103	0.2960	-7471B		PPB	06 Apr 2018 10:50:02	U	SMPL
J2256-01DUP - 1	FK-BPM-02-004-ABCDUP	1269	0.3353	-7471B		PPB	06 Apr 2018 10:54:50	U	SMPL
J2256-01MS - 1	FK-BPM-02-004-ABCDMS	16679	3.9879	-7471B		PPB	06 Apr 2018 10:57:06	U	SMPL
J2256-01MSD - 1	FK-BPM-02-004-ABCDMSD	16310	3.9004	-7471B		PPB	06 Apr 2018 10:59:23	U	SMPL
J2257-01 - 1	LP-OILY-STONE	5372	1.3078	-7471B		PPB	06 Apr 2018 11:01:40	U	SMPL
J2258-01 - 1	11984	752	0.2128	-7471B		PPB	06 Apr 2018 11:03:57	U	SMPL
J2265-01 - 1	PF-2	447	0.1405	-7471B		PPB	06 Apr 2018 11:06:15	U	SMPL
J2265-02 - 1	PF-3	428	0.1360	-7471B		PPB	06 Apr 2018 11:08:33	U	SMPL
J2265-03 - 1	PF-4	356	0.1189	-7471B		PPB	06 Apr 2018 11:10:49	U	SMPL
CCV78 - 1	CCV78	20501	4.8938	-7471B		PPB	06 Apr 2018 11:13:05	U	SMPL
CCB78 - 1	CCB78	-59	0.0205	-7471B		PPB	06 Apr 2018 11:15:22	U	SMPL
J2265-04 - 1	PF-5	596	0.1758	-7471B		PPB	06 Apr 2018 11:17:38	U	SMPL
J2265-05 - 1	PF-6	382	0.1251	-7471B		PPB	06 Apr 2018 11:19:54	U	SMPL
J2265-06 - 1	PF-7	387	0.1262	-7471B		PPB	06 Apr 2018 11:22:11	U	SMPL
J2256-01LX5 - 1		244	0.0924	-7471B		PPB	06 Apr 2018 11:24:27	U	SMPL
J2256-01A - 1		18537	4.4283	-7471B		PPB	06 Apr 2018 11:26:44	U	SMPL
CCV79 - 1	CCV79	20464	4.8850	-7471B		PPB	06 Apr 2018 11:29:01	U	SMPL
CCB79 - 1	CCB79	-112	0.0080	-7471B		PPB	06 Apr 2018 11:31:18	U	SMPL