


7470A

LB94434
INSTRUMENT ID: CV1

Linear

 μ Abs.:

34133



A= 0.0000e+000

B= 2.9239e-004

C= 3.7761e-002

Rho= 0.9998305

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.040	0.040	8	0.000	8				
0.2	0.200	0.237	0.037	683	0.0 %	683				
2.5	2.500	2.467	-0.033	8308	0.0 %	8308				
5.0	5.000	4.865	-0.135	16511	0.0 %	16511				
7.5	7.500	7.572	0.072	25767	0.0 %	25767				
10.0	10.000	10.018	0.018	34133	0.0 %	34133				

LB94434 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 06 Apr 2018 12:28:55

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	8	-	0.00007470A	PPB	06 Apr 2018 12:30:39	S	Std
0.2 - 1	50.2	683	-	0.20007470A	PPB	06 Apr 2018 12:32:47	S	Std
2.5 - 1	52.5	8308	-	2.50007470A	PPB	06 Apr 2018 12:34:56	S	Std
5.0 - 1	55	16511	-	5.00007470A	PPB	06 Apr 2018 12:37:04	S	Std
7.5 - 1	57.5	25767	-	7.50007470A	PPB	06 Apr 2018 12:39:12	S	Std
10.0 - 1	510	34133	-	10.00007470A	PPB	06 Apr 2018 12:41:20	S	Std
ICV81 - 1	ICV81	13136	3.8787	-7470A	PPB	06 Apr 2018 12:44:00	U	SMPL
ICB81 - 1	ICB81	-161	-0.0093	-7470A	PPB	06 Apr 2018 12:46:06	U	SMPL
CCV80 - 1	CCV80	16422	4.8395	-7470A	PPB	06 Apr 2018 12:48:14	U	SMPL
CCB80 - 1	CCB80	-154	-0.0073	-7470A	PPB	06 Apr 2018 12:50:20	U	SMPL
CRA - 1	CRA	513	0.1878	-7470A	PPB	06 Apr 2018 12:52:27	U	SMPL
HighStd - 1	HighStd	33740	9.9032	-7470A	PPB	06 Apr 2018 12:54:34	U	SMPL
ChkStd - 1	ChkStd	23581	6.9327	-7470A	PPB	06 Apr 2018 12:56:41	U	SMPL
PB108042BL - 1	PBW	-248	-0.0348	-7470A	PPB	06 Apr 2018 12:58:50	U	SMPL
PB108042BS - 1	LCSW	13213	3.9012	-7470A	PPB	06 Apr 2018 13:01:00	U	SMPL
PB108042TB - 1	PB108042TB PB107997TB	-129	0.0000	-7470A	PPB	06 Apr 2018 13:03:07	U	SMPL
J2153-02 - 1	HR-06-040418-A	-31	0.0287	-7470A	PPB	06 Apr 2018 13:05:16	U	SMPL
J2153-02DUP - 1	HR-06-040418-ADUP	-29	0.0293	-7470A	PPB	06 Apr 2018 13:07:24	U	SMPL
J2153-02MS - 1	HR-06-040418-AMS	12662	3.7401	-7470A	PPB	06 Apr 2018 13:09:33	U	SMPL
J2153-02MSD - 1	HR-06-040418-AMSD	12838	3.7915	-7470A	PPB	06 Apr 2018 13:11:39	U	SMPL
CCV81 - 1	CCV81	16229	4.7830	-7470A	PPB	06 Apr 2018 13:13:46	U	SMPL
CCB81 - 1	CCB81	-152	-0.0067	-7470A	PPB	06 Apr 2018 13:15:53	U	SMPL
J2153-04 - 1	HR-06-040418-B	28	0.0459	-7470A	PPB	06 Apr 2018 13:18:01	U	SMPL
J2153-06 - 1	HR-06-040418-C	-15	0.0334	-7470A	PPB	06 Apr 2018 13:20:07	U	SMPL
J2214-03 - 1	HH-STONE	18	0.0430	-7470A	PPB	06 Apr 2018 13:22:14	U	SMPL
J2223-04 - 1	CV-RB47188RT	16	0.0424	-7470A	PPB	06 Apr 2018 13:24:21	U	SMPL
J2228-03 - 1	ACV-102-20-ACV-43-30	-27	0.0299	-7470A	PPB	06 Apr 2018 13:26:28	U	SMPL
J2231-03 - 1	42ND-STREET	0	0.0378	-7470A	PPB	06 Apr 2018 13:28:36	U	SMPL
J2232-03 - 1	CV-RB26667RT	-6	0.0360	-7470A	PPB	06 Apr 2018 13:30:44	U	SMPL
PB108043BL - 1	PBW	-11	0.0345	-7470A	PPB	06 Apr 2018 13:32:53	U	SMPL
PB108043BS - 1	LCSW	12563	3.7111	-7470A	PPB	06 Apr 2018 13:35:01	U	SMPL
J2236-01 - 1	MW-103	1464	0.4658	-7470A	PPB	06 Apr 2018 13:37:08	U	SMPL
CCV82 - 1	CCV82	16083	4.7403	-7470A	PPB	06 Apr 2018 13:39:15	U	SMPL
CCB82 - 1	CCB82	-98	0.0091	-7470A	PPB	06 Apr 2018 13:41:21	U	SMPL
J2236-02 - 1	MW-104	50	0.0524	-7470A	PPB	06 Apr 2018 13:43:30	U	SMPL
J2236-03 - 1	MW-105	281	0.1199	-7470A	PPB	06 Apr 2018 13:45:36	U	SMPL
J2236-04 - 1	MW-101	183	0.0913	-7470A	PPB	06 Apr 2018 13:47:43	U	SMPL
J2236-04DUP - 1	MW-101DUP	157	0.0837	-7470A	PPB	06 Apr 2018 13:49:50	U	SMPL
J2236-05 - 1	J2236-04MS	15604	4.6003	-7470A	PPB	06 Apr 2018 13:51:58	U	SMPL
J2236-06 - 1	J2236-04MSD	15234	4.4921	-7470A	PPB	06 Apr 2018 13:54:05	U	SMPL
J2236-07 - 1	DUP-01	-14	0.0337	-7470A	PPB	06 Apr 2018 13:56:14	U	SMPL
J2236-08 - 1	FB-040418	-31	0.0287	-7470A	PPB	06 Apr 2018 13:58:22	U	SMPL
CCV83 - 1	CCV83	16203	4.7754	-7470A	PPB	06 Apr 2018 14:00:30	U	SMPL
CCB83 - 1	CCB83	-102	0.0079	-7470A	PPB	06 Apr 2018 14:02:37	U	SMPL
J2153-02LX5 - 1		-3	0.0369	-7470A	PPB	06 Apr 2018 14:04:45	U	SMPL
J2153-02A - 1		13447	3.9696	-7470A	PPB	06 Apr 2018 14:06:52	U	SMPL
J2236-04LX5 - 1		-49	0.0234	-7470A	PPB	06 Apr 2018 14:08:59	U	SMPL
J2236-04A - 1		13482	3.9798	-7470A	PPB	06 Apr 2018 14:13:41	U	SMPL
J2247-04 - 1	CG-1	-62	0.0196	-7470A	PPB	06 Apr 2018 14:15:47	U	SMPL
CCV84 - 1	CCV84	16360	4.8213	-7470A	PPB	06 Apr 2018 14:17:55	U	SMPL
CCB84 - 1	CCB84	-134	-0.0014	-7470A	PPB	06 Apr 2018 14:20:03	U	SMPL