


ISM02.4

LB94447
INSTRUMENT ID: CV1

Linear

 μ Abs.:

40859



A= 0.0000e+000

B= 2.4573e-004

C= -1.0005e-002

Rho= 0.9999027

Accept=Accepted

slope
 $y = 1.0712 \times 10^{-4}x$

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0 D
0.00	0.000	-0.002	-0.002	33	0.000	33					2
0.2	0.200	0.205	0.005	876	0.0 %	876					-1
2.5	2.500	2.473	-0.027	10105	0.0 %	10105					2
5.0	5.000	5.081	0.081	20717	0.0 %	20717					-1
7.5	7.500	7.412	-0.088	30205	0.0 %	30205					0
10.0	10.000	10.030	0.030	40859	0.0 %	40859					

123
4/6/18

LB94447
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	33	-		0 ISM02.4	PPB	4/6/2018 15:27	Std
	0.2 S0.2	876	-		0.2 ISM02.4	PPB	4/6/2018 15:30	Std
	2.5 S2.5	10105	-		2.5 ISM02.4	PPB	4/6/2018 15:35	Std
	5 S5	20717	-		5 ISM02.4	PPB	4/6/2018 15:38	Std
	7.5 S7.5	30205	-		7.5 ISM02.4	PPB	4/6/2018 15:40	Std
	10 S10	40859	-		10 ISM02.4	PPB	4/6/2018 15:45	Std
ICV	ICV	15852	3.8853	-	ISM02.4	PPB	4/6/2018 15:48	SMPL
ICB	ICB	-109	-0.0368	-	ISM02.4	PPB	4/6/2018 15:50	SMPL
CCV085	CCV085	20960	5.1405	-	ISM02.4	PPB	4/6/2018 15:52	SMPL
CCB085	CCB085	-139	-0.0442	-	ISM02.4	PPB	4/6/2018 15:55	SMPL
PB108062BL	PBW003	-3	-0.0107	-	ISM02.4	PPB	4/6/2018 15:57	SMPL
J2215-01	RW-12D-20180328	72	0.0077	-	ISM02.4	PPB	4/6/2018 15:59	SMPL
J2215-01D	RW-12D-20180328D	51	0.0025	-	ISM02.4	PPB	4/6/2018 16:01	SMPL
J2215-01S	RW-12D-20180328S	3332	0.8088	-	ISM02.4	PPB	4/6/2018 16:04	SMPL
J2215-02	RW-12I-20180328	33	-0.0019	-	ISM02.4	PPB	4/6/2018 16:07	SMPL
J2215-03	RW-12D-20180328	98	0.0141	-	ISM02.4	PPB	4/6/2018 16:09	SMPL
J2215-04	RW-12I-20180328	33	-0.0019	-	ISM02.4	PPB	4/6/2018 16:11	SMPL
J2252-01	FIELD-BLANK-20180404	43	0.0006	-	ISM02.4	PPB	4/6/2018 16:14	SMPL
CCV086	CCV086	20599	5.0518	-	ISM02.4	PPB	4/6/2018 16:16	SMPL
CCB086	CCB086	-117	-0.0388	-	ISM02.4	PPB	4/6/2018 16:18	SMPL