



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB94457

Slope.: 98.1

BalanceID: WC SC-4

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			18.3	3.99	04/06/2018	17:10
2	CAL2	1	Water			18.3	7.00	04/06/2018	17:11
3	CAL3	1	Water			18.3	10.02	04/06/2018	17:12
4	ICV	1	Water			18.2	6.99	04/06/2018	17:15
5	CCV1	1	Water			18.2	2.01	04/06/2018	17:16
6	J2155-02	1	Solid	20.02	20	19.4	6.85	04/06/2018	17:19
7	J2155-02DUP	1	Solid	20.03	20	19.4	6.86	04/06/2018	17:20
8	J2155-04	1	Solid	20.04	20	19.7	5.66	04/06/2018	17:21
9	J2155-06	1	Solid	20.01	20	20.1	6.30	04/06/2018	17:23
10	J2157-02	1	Solid	20.03	20	19.5	8.92	04/06/2018	17:25
11	J2157-04	1	Solid	20.02	20	19.6	8.52	04/06/2018	17:26
12	J2157-06	1	Solid	20.05	20	20.2	8.54	04/06/2018	17:28
13	J2157-08	1	Solid	20.02	20	19.7	8.45	04/06/2018	17:30
14	J2157-10	1	Solid	20.03	20	19.5	9.11	04/06/2018	17:31
15	J2158-02	1	Solid	20.02	20	20.1	8.64	04/06/2018	17:32
16	CCV2	1	Water			18.2	12.01	04/06/2018	17:33
17	J2270-02	1	Solid	20.04	20	19.6	7.80	04/06/2018	17:35
18	J2270-04	1	Solid	20.02	20	19.2	8.20	04/06/2018	17:36
19	J2270-06	1	Solid	20.04	20	20.2	8.00	04/06/2018	17:37
20	J2278-03	1	Solid	20.05	20	20.4	9.51	04/06/2018	17:39

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	J2280-02	1	Solid	20.06	20	19.8	8.45	04/06/2018	17:40
22	J2280-04	1	Solid	20.03	20	20.3	8.11	04/06/2018	17:42
23	J2280-04DUP	1	Solid	20.04	20	20.4	8.12	04/06/2018	17:43
24	J2283-02	1	Solid	20.03	20	20.1	11.18	04/06/2018	17:45
25	CCV3	1	Water			18.2	2.01	04/06/2018	17:46

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity j2278

WorkList ID : 110539

Date : 4/6/2018 2:49:05 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/06/2018	Solid	J2155-02	Corrosivity	Cool 4 deg C	PSEG05	M33	NB-01-040618-A	04/06/2018	9045C
04/06/2018	Solid	J2155-04	Corrosivity	Cool 4 deg C	PSEG05	M33	NB-01-040618-B	04/06/2018	9045C
04/06/2018	Solid	J2155-06	Corrosivity	Cool 4 deg C	PSEG05	M33	NB-01-040618-C	04/06/2018	9045C
04/06/2018	Solid	J2157-02	Corrosivity	Cool 4 deg C	PSEG05	M53	CL-01-040618-A	04/06/2018	9045C
04/06/2018	Solid	J2157-04	Corrosivity	Cool 4 deg C	PSEG05	M53	CL-01-040618-B	04/06/2018	9045C
04/06/2018	Solid	J2157-06	Corrosivity	Cool 4 deg C	PSEG05	M53	CL-01-040618-C	04/06/2018	9045C
04/06/2018	Solid	J2157-08	Corrosivity	Cool 4 deg C	PSEG05	M53	CL-01-040618-D	04/06/2018	9045C
04/06/2018	Solid	J2157-10	Corrosivity	Cool 4 deg C	PSEG05	M53	CL-01-040618-E	04/06/2018	9045C
04/06/2018	Solid	J2158-02	Corrosivity	Cool 4 deg C	PSEG05	M43	OK-03-040618	04/06/2018	9045C
04/05/2018	Solid	J2270-02	Corrosivity	Cool 4 deg C	ENVI27	L22	SOUTHLAWN-PRECHAR-AR	04/05/2018	9045C
04/05/2018	Solid	J2270-04	Corrosivity	Cool 4 deg C	ENVI27	L22	SOUTHLAWN-PRECHAR-AR	04/05/2018	9045C
04/05/2018	Solid	J2270-06	Corrosivity	Cool 4 deg C	ENVI27	L22	SOUTHLAWN-PRECHAR-AR	04/05/2018	9045C
04/06/2018	Solid	J2278-03	Corrosivity	Cool 4 deg C	PSEG02	M51	DRUM-2-3-6-8-9-11-13	04/06/2018	9045C
04/06/2018	Solid	J2280-02	Corrosivity	Cool 4 deg C	PSEG01	M53	S1	04/06/2018	9045C
04/06/2018	Solid	J2280-04	Corrosivity	Cool 4 deg C	PSEG01	M53	S2	04/06/2018	9045C
04/06/2018	Solid	J2283-02	Corrosivity	Cool 4 deg C	SEAC01	L11	SOIL-CUTTINGS	04/06/2018	9045C

Date/Time 04.06.18 2:50 PM

Received by: JP

Relinquished by: CP

Date/Time 04.06.18 630 PM

Received by: CP

Relinquished by: JP