



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: Non-Polar Material
Run Number: LB94465
Analysis Date: 04/09/2018
BalanceID: WC SC-3
OvenID: EX Oven-2

ANALYST: JIGNESH
REVIEWED BY: apatel
Extraction Date: 04/09/2018
Extraction IN Time: 08:10
Extraction OUT Time: 09:11

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB94465BL	LB94465BL	WATER	2	1000	100	2.8403	2.8403	3.02	2.8403	2.8403	0.0000	0
2	LB94465BS	LB94465BS	WATER	2	1000	100	2.9144	2.9144	3.01	2.9334	2.9334	0.0190	19
3	LB94465BSD	LB94465BSD	WATER	2	1000	100	2.8811	2.8811	3.03	2.9003	2.9003	0.0192	19.2
4	J2268-01	MH-92A	WATER	2	970	100	2.9815	2.9815	3.04	2.9817	2.9817	0.0002	0.21
5	J2268-02	MH-33	WATER	2	1000	100	2.9886	2.9886	3.05	2.9899	2.9899	0.0013	1.3



QC Batch# LB94465 Test: Non-Polar Ma Analysis Date: 04/09/2018

Chemicals Used:

Chemical Name	Chemical Lot #
HEXANE	W2341
pH Paper 0-14	W2248
Sodium Sulfate	EP1832
1:1 HCL	WP59249
Silica Gel	W2199
Sand	NA

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	10.00 ML	WP58005
LCSWD	10.00 ML	WP58006
MS/MSD	NA	NA

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-3

Before Analysis

0.0020 gram Balance:	<u>0.0018</u>	In OVEN TEMP :	<u>71</u>	Dessicator Time In :	<u>10:40</u>
1.0000 gram Balance:	<u>1.0003</u>	In Time:	<u>09:37</u>		
Bal Check Time:	<u>08:30</u>	Out OVEN TEMP:	<u>71</u>	Dessicator Time Out:	<u>11:15</u>
		Out Time:	<u>10:39</u>		

After Analysis

0.0020 gram Balance:	<u>0.0021</u>	In OVEN TEMP :	<u>70</u>	Dessicator Time In :	<u>12:17</u>
1.0000 gram Balance:	<u>1.0005</u>	In Time:	<u>11:42</u>		
Bal Check Time:	<u>13:26</u>	Out OVEN TEMP:	<u>70</u>	Dessicator Time Out:	<u>12:50</u>
		Out Time:	<u>12:16</u>		

1894465

WORKLIST(Hardcopy Internal Chain)

WorkList Name : NON POPLAR J2282 WorkList ID : 110544 Date : 4/9/2018 7:53:04 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/20/2018	Water	J2268-01 - O	Non-Polar Material	Conc H2SO4 4c	DIST01	L21	MH-92A	04/06/2018	1664A
04/20/2018	Water	J2268-02 - O	Non-Polar Material	Conc H2SO4 4c	DIST01	L21	MH-33	04/06/2018	1664A

Date/Time 04-09-18 8:00 PM
Received by: JP
Relinquished by: JP

Date/Time 04-09-18 9:30 AM
Received by: CP
Relinquished by: JP