

LB94469


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

31779



A= 0.0000e+000

B= 3.1198e-004

C= 2.3565e-002

Rho= 0.9998115

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.020	0.020	-12	0.000	-12				
0.2	0.200	0.222	0.022	635	0.0 %	635				
2.5	2.500	2.445	-0.055	7762	0.0 %	7762				
5.0	5.000	4.937	-0.063	15750	0.0 %	15750				
7.5	7.500	7.638	0.138	24408	0.0 %	24408				
10.0	10.000	9.938	-0.062	31779	0.0 %	31779				

LB94469 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 09 Apr 2018 09:02:45

Sample ID	Extended ID	μ Abs.	Conc.	Std Cond	Method	Units	Date	Type	Type
0.0 - 1	50	-12	-	0.0000	7470A	PPB	09 Apr 2018 09:13:48	S	Std
0.2 - 1	50.2	635	-	0.2000	7470A	PPB	09 Apr 2018 09:15:57	S	Std
2.5 - 1	52.5	7762	-	2.5000	7470A	PPB	09 Apr 2018 09:18:05	S	Std
5.0 - 1	55	15750	-	5.0000	7470A	PPB	09 Apr 2018 09:20:13	S	Std
7.5 - 1	57.5	24408	-	7.5000	7470A	PPB	09 Apr 2018 09:22:22	S	Std
10.0 - 1	51.0	31779	-	10.0000	7470A	PPB	09 Apr 2018 09:29:27	S	Std
ICV82 - 1	ICV82	12704	3.9869	-7470A	PPB	09 Apr 2018 09:32:03	U	SMPL	
ICB82 - 1	ICB82	-42	0.0105	-7470A	PPB	09 Apr 2018 09:36:32	U	SMPL	
CCV87 - 1	CCV87	16118	5.0520	-7470A	PPB	09 Apr 2018 09:38:38	U	SMPL	
CCB87 - 1	CCB87	-155	-0.0248	-7470A	PPB	09 Apr 2018 09:40:44	U	SMPL	
CRA - 1	CRA	622	0.2176	-7470A	PPB	09 Apr 2018 09:42:52	U	SMPL	
HighStd - 1	HighStd	32077	10.0308	-7470A	PPB	09 Apr 2018 09:44:58	U	SMPL	
ChkStd - 1	ChkStd	23069	7.2206	-7470A	PPB	09 Apr 2018 09:47:05	U	SMPL	
PB108077BL - 1	PBW	-183	-0.0335	-7470A	PPB	09 Apr 2018 09:49:14	U	SMPL	
PB108077BS - 1	LCSW	13567	4.2562	-7470A	PPB	09 Apr 2018 09:51:23	U	SMPL	
PB108077TB - 1	PB108037TB	-6	0.0217	-7470A	PPB	09 Apr 2018 09:53:30	U	SMPL	
J2164-02 - 1	SA-01-040518	170	0.0766	-7470A	PPB	09 Apr 2018 09:55:39	U	SMPL	
J2165-02 - 1	SU-04-040518	-4	0.0223	-7470A	PPB	09 Apr 2018 09:57:47	U	SMPL	
J2248-02 - 1	11944	344	0.1309	-7470A	PPB	09 Apr 2018 09:59:56	U	SMPL	
J2248-02DUP - 1	11944DUP	335	0.1281	-7470A	PPB	09 Apr 2018 10:04:35	U	SMPL	
CCV88 - 1	CCV88	16067	5.0361	-7470A	PPB	09 Apr 2018 10:06:41	U	SMPL	
CCB88 - 1	CCB88	179	0.0794	-7470A	PPB	09 Apr 2018 10:08:48	U	SMPL	
J2248-02MS - 1	11944MS	14123	4.4296	-7470A	PPB	09 Apr 2018 10:10:56	U	SMPL	
J2248-02MSD - 1	11944MSD	13754	4.3145	-7470A	PPB	09 Apr 2018 10:13:03	U	SMPL	
J2248-04 - 1	12000	100	0.0548	-7470A	PPB	09 Apr 2018 10:15:10	U	SMPL	
J2248-06 - 1	RT3059	66	0.0442	-7470A	PPB	09 Apr 2018 10:17:18	U	SMPL	
J2254-02 - 1	STOCKPILE	43	0.0370	-7470A	PPB	09 Apr 2018 10:19:26	U	SMPL	
J2257-03 - 1	LP-OILY-STONE	86	0.0504	-7470A	PPB	09 Apr 2018 10:21:34	U	SMPL	
J2258-02 - 1	11984	53	0.0401	-7470A	PPB	09 Apr 2018 10:23:42	U	SMPL	
PB108078BL - 1	PBW	75	0.0470	-7470A	PPB	09 Apr 2018 10:25:50	U	SMPL	
PB108078BS - 1	LCSW	13977	4.3841	-7470A	PPB	09 Apr 2018 10:27:58	U	SMPL	
PB108078TB - 1	PB108041TB	29	0.0326	-7470A	PPB	09 Apr 2018 10:30:05	U	SMPL	
CCV89 - 1	CCV89	16666	5.2230	-7470A	PPB	09 Apr 2018 10:34:51	U	SMPL	
CCB89 - 1	CCB89	78	0.0479	-7470A	PPB	09 Apr 2018 10:36:58	U	SMPL	
J2247-05 - 1	CG-1	146	0.0691	-7470A	PPB	09 Apr 2018 10:39:06	U	SMPL	
J2247-05DUP - 1	CG-1DUP	110	0.0579	-7470A	PPB	09 Apr 2018 10:41:13	U	SMPL	
J2247-05MS - 1	CG-1MS	14438	4.5279	-7470A	PPB	09 Apr 2018 10:43:20	U	SMPL	
J2247-05MSD - 1	CG-1MSD	13071	4.1014	-7470A	PPB	09 Apr 2018 10:45:27	U	SMPL	
J2248-02LX5 - 1		-9	0.0208	-7470A	PPB	09 Apr 2018 10:47:35	U	SMPL	
J2248-02A - 1		12950	4.0637	-7470A	PPB	09 Apr 2018 10:49:43	U	SMPL	
J2247-05LX5 - 1		-31	0.0139	-7470A	PPB	09 Apr 2018 10:51:50	U	SMPL	
J2247-05A - 1		12802	4.0175	-7470A	PPB	09 Apr 2018 10:53:59	U	SMPL	
CCV90 - 1	CCV90	15389	4.8246	-7470A	PPB	09 Apr 2018 11:00:59	U	SMPL	
CCB90 - 1	CCB90	1	0.0239	-7470A	PPB	09 Apr 2018 11:03:05	U	SMPL	