

245.1


LB94476

INSTRUMENT ID: CVA

Linear

 μ Abs.:

34451



A= 0.0000e+000

B= 2.9311e-004

C= -2.2813e-002

Rho= 0.9998629

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.009	0.009	110	0.000	110				
0.05	0.050	0.053	0.003	258	0.0 %	258				
0.2	0.200	0.199	-0.001	758	0.0 %	758				
2.5	2.500	2.502	0.002	8613	0.0 %	8613				
5.0	5.000	5.047	0.047	17298	0.0 %	17298				
7.5	7.500	7.364	-0.136	25202	0.0 %	25202				
10.0	10.000	10.075	0.075	34451	0.0 %	34451				

LB94476 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 09 Apr 2018 13:17:45

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	110	-	0.0000	245.1	PPB	09 Apr 2018 13:20:19	S	Std
0.05 - 1	50.05	258	-	0.0500	245.1	PPB	09 Apr 2018 13:22:27	S	Std
0.2 - 1	50.2	758	-	0.2000	245.1	PPB	09 Apr 2018 13:24:36	S	Std
2.5 - 1	52.5	8613	-	2.5000	245.1	PPB	09 Apr 2018 13:26:44	S	Std
5.0 - 1	55	17298	-	5.0000	245.1	PPB	09 Apr 2018 13:28:52	S	Std
7.5 - 1	57.5	25202	-	7.5000	245.1	PPB	09 Apr 2018 13:30:59	S	Std
10.0 - 1	510	34451	-	10.0000	245.1	PPB	09 Apr 2018 13:35:54	S	Std
ICV83 - 1	ICV83	13072	3.8087	-245.1		PPB	09 Apr 2018 13:43:42	U	SMPL
ICB83 - 1	ICB83	5	-0.0213	-245.1		PPB	09 Apr 2018 13:45:48	U	SMPL
CCV91 - 1	CCV91	17560	5.1242	-245.1		PPB	09 Apr 2018 13:47:55	U	SMPL
CCB91 - 1	CCB91	-25	-0.0301	-245.1		PPB	09 Apr 2018 13:50:02	U	SMPL
CRA - 1	CRA	244	0.0487	-245.1		PPB	09 Apr 2018 13:52:09	U	SMPL
HighStd - 1	HighStd	34187	9.9977	-245.1		PPB	09 Apr 2018 13:54:16	U	SMPL
ChkStd - 1	ChkStd	23115	6.7524	-245.1		PPB	09 Apr 2018 13:56:23	U	SMPL
PB108079BL - 1	PBW	-102	-0.0527	-245.1		PPB	09 Apr 2018 13:58:31	U	SMPL
PB108079BS - 1	LCSW	12926	3.7659	-245.1		PPB	09 Apr 2018 14:00:41	U	SMPL
J2255-01 - 1	040418-JETB-E-D-B	173	0.0279	-245.1		PPB	09 Apr 2018 14:02:49	U	SMPL
J2255-02 - 1	040418-JETB-E-D-M	177	0.0291	-245.1		PPB	09 Apr 2018 14:04:57	U	SMPL
J2255-03 - 1	040418-JETB-E-D-S	196	0.0346	-245.1		PPB	09 Apr 2018 14:07:05	U	SMPL
J2255-04 - 1	040418-JETB-E-U-B	190	0.0329	-245.1		PPB	09 Apr 2018 14:09:13	U	SMPL
J2255-05 - 1	040418-JETB-E-U-M	187	0.0320	-245.1		PPB	09 Apr 2018 14:11:20	U	SMPL
CCV92 - 1	CCV92	17037	4.9709	-245.1		PPB	09 Apr 2018 14:13:27	U	SMPL
CCB92 - 1	CCB92	-24	-0.0298	-245.1		PPB	09 Apr 2018 14:15:34	U	SMPL
J2255-06 - 1	040418-JETB-E-U-S	168	0.0264	-245.1		PPB	09 Apr 2018 14:17:42	U	SMPL
J2255-07 - 1	040418-DBRI-E-D-B	174	0.0282	-245.1		PPB	09 Apr 2018 14:19:48	U	SMPL
J2255-08 - 1	040418-DBRI-E-D-M	198	0.0352	-245.1		PPB	09 Apr 2018 14:21:55	U	SMPL
J2255-09 - 1	040418-DBRI-E-D-S	189	0.0326	-245.1		PPB	09 Apr 2018 14:24:03	U	SMPL
J2255-10 - 1	040418-DBRI-E-U-B	178	0.0294	-245.1		PPB	09 Apr 2018 14:26:10	U	SMPL
J2255-11 - 1	040418-DBRI-E-U-M	138	0.0176	-245.1		PPB	09 Apr 2018 14:28:18	U	SMPL
J2255-12 - 1	040418-DBRI-E-U-S	221	0.0420	-245.1		PPB	09 Apr 2018 14:30:26	U	SMPL
J2264-02 - 1	BOTTLE-2-PH-TANK	159	0.0238	-245.1		PPB	09 Apr 2018 14:32:34	U	SMPL
J2268-01 - 1	MH-92A	622	0.1595	-245.1		PPB	09 Apr 2018 14:34:43	U	SMPL
J2268-01DUP - 1	MH-92ADUP	616	0.1577	-245.1		PPB	09 Apr 2018 14:41:29	U	SMPL
CCV93 - 1	CCV93	17471	5.0981	-245.1		PPB	09 Apr 2018 14:43:36	U	SMPL
CCB93 - 1	CCB93	107	0.0086	-245.1		PPB	09 Apr 2018 14:45:43	U	SMPL
J2268-01MS - 1	MH-92AMS	16379	4.7780	-245.1		PPB	09 Apr 2018 14:47:49	U	SMPL
J2268-01MSD - 1	MH-92AMSD	16343	4.7675	-245.1		PPB	09 Apr 2018 14:49:56	U	SMPL
J2268-02 - 1	MH-33	1768	0.4954	-245.1		PPB	09 Apr 2018 14:52:03	U	SMPL
J2273-01 - 1	040518-DBRI-F-D-B	65	-0.0038	-245.1		PPB	09 Apr 2018 14:54:11	U	SMPL
J2273-02 - 1	040518-DBRI-F-D-S	117	0.0115	-245.1		PPB	09 Apr 2018 14:56:19	U	SMPL
J2273-03 - 1	040518-DBRI-F-U-B	111	0.0097	-245.1		PPB	09 Apr 2018 14:58:26	U	SMPL
J2273-04 - 1	040518-DBRI-F-U-S	93	0.0044	-245.1		PPB	09 Apr 2018 15:00:35	U	SMPL
J2268-01LX5 - 1	CCV94	164	0.0253	-245.1		PPB	09 Apr 2018 15:02:43	U	SMPL
J2268-01A - 1		15347	4.4755	-245.1		PPB	09 Apr 2018 15:04:51	U	SMPL
CCV94 - 1	CCV94	17477	5.0999	-245.1		PPB	09 Apr 2018 15:06:59	U	SMPL
CCB94 - 1	CCB94	-22	-0.0293	-245.1		PPB	09 Apr 2018 15:09:06	U	SMPL