

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

4/9/2018 13:05

Test: Phenolics-mg/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.017	0.0	0.108	
ICB1	0.006	0.0	0.004	
CCV1	1.024	0.0	0.108	
CCB1	0.002	0.0	0.004	
PB108074BL	0.001	0.0	0.004	
PB108074BS	1.007	0.0	0.107	
J2268-01	0.014	0.0	0.005	
J2268-01DUP	0.016	0.0	0.005	
J2268-01MS	0.664	0.0	0.072	
J2268-01MSD	0.662	0.0	0.071	
J2268-02	-0.002	0.0	0.004	
CCV2	1.017	0.0	0.108	
CCB2	0.006	0.0	0.004	

N	13
Mean	0.418
SD	0.4779
CV%	114.37

Aquakem v. 7.2AQ1

AS

Results from time period:

Mon Apr 09 12:35:36 2018

Mon Apr 09 12:38:24 2018

Sample Id	Sam/Ctr/c/	Test short nan	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0526	mg/l	4/9/2018 11:36:26	
0.1PPM	A	Phenolics-	P	0.1059	mg/l	4/9/2018 11:36:27	
0.25PPM	A	Phenolics-	P	0.2529	mg/l	4/9/2018 11:36:28	
0.50PPM	A	Phenolics-	P	0.4916	mg/l	4/9/2018 11:36:29	
1.0PPM	A	Phenolics-	P	0.9913	mg/l	4/9/2018 11:36:30	
2.0PPM	A	Phenolics-	P	2.0057	mg/l	4/9/2018 11:36:31	
ICV1	S	Phenolics-	P	1.0172	mg/l	4/9/2018 12:35:36	
ICB1	S	Phenolics-	P	0.0058	mg/l	4/9/2018 12:35:37	
CCV1	S	Phenolics-	P	1.0244	mg/l	4/9/2018 12:35:38	
CCB1	S	Phenolics-	P	0.0023	mg/l	4/9/2018 12:35:39	
PB108074BL	S	Phenolics-	P	0.0008	mg/l	4/9/2018 12:35:40	
PB108074BS	S	Phenolics-	P	1.0067	mg/l	4/9/2018 12:35:41	
J2268-01	S	Phenolics-	P	0.0137	mg/l	4/9/2018 12:35:42	
J2268-01DUP	S	Phenolics-	P	0.0156	mg/l	4/9/2018 12:35:43	
J2268-01MS	S	Phenolics-	P	0.6637	mg/l	4/9/2018 12:35:44	
J2268-01MSD	S	Phenolics-	P	0.6616	mg/l	4/9/2018 12:35:45	
J2268-02	S	Phenolics-	P	-0.0021	mg/l	4/9/2018 12:35:46	
CCV2	S	Phenolics-	P	1.0174	mg/l	4/9/2018 12:35:47	
CCB2	S	Phenolics-	P	0.0055	mg/l	4/9/2018 12:38:24	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

4/9/2018 11:43


Test Phenolics-

Accepted 4/9/2018 11:42

Factor 9.795
Bias 0.004

Coeff. of det. 0.999918

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.009	0.0526	0.0500	
2	Phenol-2	0.015	0.1059	0.1000	
3	Phenol-2	0.030	0.2529	0.2500	
4	Phenol-2	0.054	0.4916	0.5000	
5	Phenol-2	0.105	0.9913	1.0000	
6	Phenol-2	0.209	2.0057	2.0000	