


7471B

LB94487

INSTRUMENT ID: CX1

 μ Abs.:

42821



Linear

A= 0.0000e+000

B= 2.3531e-004

C= -3.1543e-002

Rho= 0.9998875

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.007	-0.007	104	0.000	104				
0.2	0.200	0.199	-0.001	978	0.0 %	978				
2.5	2.500	2.584	0.084	11117	0.0 %	11117				
5.0	5.000	4.909	-0.091	20994	0.0 %	20994				
7.5	7.500	7.471	-0.029	31883	0.0 %	31883				
10	10.000	10.045	0.045	42821	0.0 %	42821				

LB94487 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 09 Apr 2018 16:34:33

Sample ID	Extended ID	µ Abs.	Conc.	Std Cond	Method	Units	Date	Type	Type
0.0 - 1	80	104	-	0.0007471B	PPB		09 Apr 2018 16:37:12	S	Std
0.2 - 1	50.2	978	-	0.20007471B	PPB		09 Apr 2018 16:39:30	S	Std
2.5 - 1	52.5	11117	-	2.50007471B	PPB		09 Apr 2018 16:41:48	S	Std
5.0 - 1	55	20994	-	5.00007471B	PPB		09 Apr 2018 16:44:05	S	Std
7.5 - 1	57.5	31883	-	7.50007471B	PPB		09 Apr 2018 16:46:22	S	Std
10 - 1	58	42821	-	10.00007471B	PPB		09 Apr 2018 16:51:36	S	Std
ICV84 - 1	ICV84	16794	3.9203	-7471B	PPB		09 Apr 2018 16:54:16	U	SMPL
ICB84 - 1	ICB84	-80	-0.0504	-7471B	PPB		09 Apr 2018 16:56:32	U	SMPL
CCV95 - 1	CCV95	22750	5.3218	-7471B	PPB		09 Apr 2018 16:58:48	U	SMPL
CCB95 - 1	CCB95	-89	-0.0525	-7471B	PPB		09 Apr 2018 17:01:03	U	SMPL
CRA - 1	CRA	942	0.1901	-7471B	PPB		09 Apr 2018 17:03:20	U	SMPL
HighStd - 1	HighStd	43601	10.2282	-7471B	PPB		09 Apr 2018 17:05:36	U	SMPL
ChkStd - 1	ChkStd	28667	6.7141	-7471B	PPB		09 Apr 2018 17:07:52	U	SMPL
PB108080BL - 1	PBS	-125	-0.0610	-7471B	PPB		09 Apr 2018 17:10:10	U	SMPL
PB108080BS - 1	LCSS	16178	3.7753	-7471B	PPB		09 Apr 2018 17:12:28	U	SMPL
J2155-01 - 1	NB-01-040618-A	2136	0.4711	-7471B	PPB		09 Apr 2018 17:14:45	U	SMPL
J2155-03 - 1	NB-01-040618-B	2597	0.5796	-7471B	PPB		09 Apr 2018 17:17:03	U	SMPL
J2155-05 - 1	NB-01-040618-C	3024	0.6800	-7471B	PPB		09 Apr 2018 17:19:20	U	SMPL
J2157-01 - 1	CL-01-040618-A	2433	0.5410	-7471B	PPB		09 Apr 2018 17:21:38	U	SMPL
J2157-03 - 1	CL-01-040618-B	2341	0.5193	-7471B	PPB		09 Apr 2018 17:23:54	U	SMPL
CCV96 - 1	CCV96	22600	5.2865	-7471B	PPB		09 Apr 2018 17:26:10	U	SMPL
CCB96 - 1	CCB96	-74	-0.0490	-7471B	PPB		09 Apr 2018 17:28:26	U	SMPL
J2157-05 - 1	CL-01-040618-C	3408	0.7704	-7471B	PPB		09 Apr 2018 17:30:42	U	SMPL
J2157-07 - 1	CL-01-040618-D	3975	0.9038	-7471B	PPB		09 Apr 2018 17:32:58	U	SMPL
J2157-09 - 1	CL-01-040618-E	1688	0.3657	-7471B	PPB		09 Apr 2018 17:35:14	U	SMPL
J2158-01 - 1	OK-03-040618	1951	0.4275	-7471B	PPB		09 Apr 2018 17:37:30	U	SMPL
J2270-01 - 1	SOUTHLAWN-PRECHAR-AREA-1	20111	4.7008	-7471B	PPB		09 Apr 2018 17:39:47	U	SMPL
J2270-03 - 1	SOUTHLAWN-PRECHAR-AREA-2	45827	10.7520	-7471B	PPB		09 Apr 2018 17:42:04	U	SMPL
J2270-05 - 1	SOUTHLAWN-PRECHAR-AREA-3	15160	3.5358	-7471B	PPB		09 Apr 2018 17:44:22	U	SMPL
J2275-01 - 1	N-NARANJITO-S002-0001-01	2631	0.5876	-7471B	PPB		09 Apr 2018 17:46:42	U	SMPL
J2275-01DUP - 1	N-NARANJITO-S002-0001-01DUP	2618	0.5845	-7471B	PPB		09 Apr 2018 17:49:00	U	SMPL
J2275-02 - 1	J2275-01MS	21152	4.9457	-7471B	PPB		09 Apr 2018 17:51:16	U	SMPL
CCV97 - 1	CCV97	22798	5.3331	-7471B	PPB		09 Apr 2018 17:53:32	U	SMPL
CCB97 - 1	CCB97	-61	-0.0459	-7471B	PPB		09 Apr 2018 17:55:48	U	SMPL
J2275-03 - 1	J2275-01MSD	19567	4.5728	-7471B	PPB		09 Apr 2018 17:58:05	U	SMPL
J2275-04 - 1	N-NARANJITO-S002-0001-02	2235	0.4944	-7471B	PPB		09 Apr 2018 18:00:21	U	SMPL
J2278-01 - 1	DRUM-2-3-6-8-9-11-13	314	0.0423	-7471B	PPB		09 Apr 2018 18:02:37	U	SMPL
J2278-04 - 1	DRUM-1-5-4-12-14-15-17	98	-0.0085	-7471B	PPB		09 Apr 2018 18:04:54	U	SMPL
J2279-02 - 1	DRUM-1	95	-0.0092	-7471B	PPB		09 Apr 2018 18:07:10	U	SMPL
J2280-01 - 1	S1	2172	0.4796	-7471B	PPB		09 Apr 2018 18:09:27	U	SMPL
J2280-03 - 1	S2	1118	0.2315	-7471B	PPB		09 Apr 2018 18:11:44	U	SMPL
J2275-01LX5 - 1		548	0.0974	-7471B	PPB		09 Apr 2018 18:14:02	U	SMPL
J2275-01A - 1		19722	4.6092	-7471B	PPB		09 Apr 2018 18:16:20	U	SMPL
J2270-03DLX2 - 1		23132	5.4117	-7471B	PPB		09 Apr 2018 18:18:36	U	SMPL
CCV98 - 1	CCV98	22869	5.3498	-7471B	PPB		09 Apr 2018 18:20:52	U	SMPL
CCB98 - 1	CCB98	-92	-0.0532	-7471B	PPB		09 Apr 2018 18:23:09	U	SMPL