



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB94488

Ambient Barometric Pressure (mmHg): 765

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	28.0	82.4	80.8	04/09/2018	09:05
2	J2247-05		1	>100.0	>212.0	>212.0	04/09/2018	09:35
3	J2247-05DUP		1	>100.0	>212.0	>212.0	04/09/2018	09:55
4	J2268-01		1	>100.0	>212.0	>212.0	04/09/2018	10:25
5	J2268-02		1	>100.0	>212.0	>212.0	04/09/2018	10:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.33

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP4918

WorkList ID : 110552

Date : 4/9/2018 8:02:53 AM

LB 94488

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/10/2018	Water	J2247-05	Flash Point	Cool 4 deg C	PSEG02	M33	CG-1	04/05/2018	1010A
04/20/2018	Water	J2268-01	Flash Point	Cool 4 deg C	DIST01	L21	MH-92A	04/06/2018	1010A
04/20/2018	Water	J2268-02	Flash Point	Cool 4 deg C	DIST01	L21	MH-33	04/06/2018	1010A

Date/Time 4/9/18 8:30 AM
 Received by: KS
 Relinquished by: CP

Date/Time 4/9/18 10:55 AM
 Received by: CP
 Relinquished by: KS