



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB94491

Slope : 98.2

BalanceID: WC SC-3

pH Meter ID : WC pH meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.2	4.01	04/09/2018	17:20
2	CAL2	1	Water			21.2	7.00	04/09/2018	17:21
3	CAL3	1	Water			21.3	10.02	04/09/2018	17:22
4	ICV	1	Water			21.2	7.01	04/09/2018	17:24
5	CCV1	1	Water			21.3	2.01	04/09/2018	17:25
6	J2156-02	1	Solid	20.03	20	21.6	7.66	04/09/2018	17:26
7	J2156-02DUP	1	Solid	20.04	20	21.5	7.67	04/09/2018	17:29
8	J2297-04	1	Solid	20.02	20	21.7	6.65	04/09/2018	17:30
9	CCV2	1	Water			21.3	12.01	04/09/2018	17:31

LB94491

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosion j12156 WorkList ID : 110582 Date : 4/9/2018 5:11:52 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/09/2018	Solid	J2156-02	Corrosivity	Cool 4 deg C	PSEG05	M62	HD-01-040918	04/09/2018	9045C
04/09/2018	Solid	J2297-04	Corrosivity	Cool 4 deg C	PSEG01	M62	RO-1-3	04/09/2018	9045C

Date/Time 04-09-18 5:11 PM
Received by: SP
Relinquished by: W

Date/Time 04-09-18 6:35 PM
Received by: SP
Relinquished by: J8