



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB94497

Slope : 98.2

BalanceID: WC SC-3

pH Meter ID : wc ph meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.2	4.01	04/09/2018	17:20
2	CAL2	1	Water			21.2	7.00	04/09/2018	17:21
3	CAL3	1	Water			21.3	10.02	04/09/2018	17:22
4	ICV	1	Water			21.2	7.01	04/09/2018	17:24
5	CCV1	1	Water			21.2	2.01	04/09/2018	17:34
6	J2302-03	1	Solid	20.03	20	22.5	6.07	04/09/2018	17:35
7	J2302-03DUP	1	Solid	20.04	20	22.7	6.08	04/09/2018	17:36
8	J2303-03	1	Solid	20.01	20	22.4	7.84	04/09/2018	17:37
9	J2304-12	1	Solid	20.05	20	22.1	7.03	04/09/2018	17:38
10	CCV2	1	Water			21.1	12.02	04/09/2018	17:40

WORKLIST(Hardcopy Internal Chain)

1394497

WorkList Name : corrosivity j2256 WorkList ID : 110585

Date : 4/9/2018 5:17:07 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/09/2018	Solid	J2302-03	Corrosivity	Cool 4 deg C	PSEG01	M62	RT-4235	04/09/2018	9045C
04/09/2018	Solid	J2303-03	Corrosivity	Cool 4 deg C	PSEG01	M63	M11	04/09/2018	9045C
04/09/2018	Solid	J2304-12	Corrosivity	Cool 4 deg C	PSEG01	M62	SP-1	04/09/2018	9045C

Date/Time 04.09.18 5.20 PM
 Received by: JL
 Relinquished by: [Signature]

Date/Time 04.09.18 6.40 PM
 Received by: [Signature]
 Relinquished by: JP