


7470A

LB94498
INSTRUMENT ID: CV1

Linear

 μ Abs.:

35255



A= 0.0000e+000

B= 2.8679e-004

C= -1.1672e-002

Rho= 0.9998304

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.020	0.020	110	0.000	110				
0.2	0.200	0.217	0.017	796	0.0 %	796				
2.5	2.500	2.540	0.040	8898	0.0 %	8898				
5.0	5.000	4.922	-0.078	17202	0.0 %	17202				
7.5	7.500	7.403	-0.097	25853	0.0 %	25853				
10.0	10.000	10.099	0.099	35255	0.0 %	35255				

LB94498 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 10 Apr 2018 10:31:35

Sample ID	Extended ID	µ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	110	-	0.00007470A	PPB	10 Apr 2018 10:48:33	S	Std
0.2 - 1	50.2	796	-	0.20007470A	PPB	10 Apr 2018 10:50:42	S	Std
2.5 - 1	52.5	8898	-	2.50007470A	PPB	10 Apr 2018 10:52:51	S	Std
5.0 - 1	55	17202	-	5.00007470A	PPB	10 Apr 2018 10:54:59	S	Std
7.5 - 1	57.5	25853	-	7.50007470A	PPB	10 Apr 2018 10:57:06	S	Std
10.0 - 1	510	35255	-	10.00007470A	PPB	10 Apr 2018 11:01:43	S	Std
ICV85 - 1	ICV85	13562	3.8777	-7470A	PPB	10 Apr 2018 11:04:15	U	SMPL
ICB85 - 1	ICB85	-47	-0.0252	-7470A	PPB	10 Apr 2018 11:06:21	U	SMPL
CCV01 - 1	CCV01	17859	5.1101	-7470A	PPB	10 Apr 2018 11:08:28	U	SMPL
CCB01 - 1	CCB01	-36	-0.0220	-7470A	PPB	10 Apr 2018 11:10:35	U	SMPL
CRA - 1	CRA	757	0.2054	-7470A	PPB	10 Apr 2018 11:12:42	U	SMPL
HighStd - 1	HighStd	35308	10.1142	-7470A	PPB	10 Apr 2018 11:14:49	U	SMPL
ChkStd - 1	ChkStd	24118	6.9051	-7470A	PPB	10 Apr 2018 11:16:55	U	SMPL
PB108110BL - 1	PBW	-104	-0.0415	-7470A	PPB	10 Apr 2018 11:19:04	U	SMPL
PB108110BS - 1	LCSW	13850	3.9603	-7470A	PPB	10 Apr 2018 11:21:13	U	SMPL
PB108110TB - 1	PB108075TB	-41	-0.0234	-7470A	PPB	10 Apr 2018 11:23:21	U	SMPL
J2155-02 - 1	NB-01-040618-A	50	0.0027	-7470A	PPB	10 Apr 2018 11:25:29	U	SMPL
J2155-04 - 1	NB-01-040618-B	54	0.0038	-7470A	PPB	10 Apr 2018 11:27:37	U	SMPL
J2155-06 - 1	NB-01-040618-C	69	0.0081	-7470A	PPB	10 Apr 2018 11:29:45	U	SMPL
J2157-02 - 1	CL-01-040618-A	279	0.0683	-7470A	PPB	10 Apr 2018 11:31:52	U	SMPL
CCV02 - 1	CCV02	17637	5.0464	-7470A	PPB	10 Apr 2018 11:33:58	U	SMPL
CCB02 - 1	CCB02	-56	-0.0277	-7470A	PPB	10 Apr 2018 11:36:05	U	SMPL
J2157-04 - 1	CL-01-040618-B	277	0.0678	-7470A	PPB	10 Apr 2018 11:38:12	U	SMPL
J2157-06 - 1	CL-01-040618-C	231	0.0546	-7470A	PPB	10 Apr 2018 11:40:19	U	SMPL
J2157-08 - 1	CL-01-040618-D	138	0.0279	-7470A	PPB	10 Apr 2018 11:42:25	U	SMPL
J2157-10 - 1	CL-01-040618-E	62	0.0061	-7470A	PPB	10 Apr 2018 11:44:32	U	SMPL
J2158-02 - 1	OK-03-040618	36	-0.0013	-7470A	PPB	10 Apr 2018 11:46:40	U	SMPL
J2270-02 - 1	SOUTHLAWN-PRECHAR-AREA-1	94	0.0153	-7470A	PPB	10 Apr 2018 11:48:47	U	SMPL
J2270-02DUP - 1	SOUTHLAWN-PRECHAR-AREA-1DUP	98	0.0164	-7470A	PPB	10 Apr 2018 11:50:56	U	SMPL
J2270-02MS - 1	SOUTHLAWN-PRECHAR-AREA-1MS	12623	3.6085	-7470A	PPB	10 Apr 2018 11:53:04	U	SMPL
J2270-02MSD - 1	SOUTHLAWN-PRECHAR-AREA-1MSD	12845	3.6721	-7470A	PPB	10 Apr 2018 11:55:12	U	SMPL
J2270-04 - 1	SOUTHLAWN-PRECHAR-AREA-2	96	0.0159	-7470A	PPB	10 Apr 2018 11:57:19	U	SMPL
CCV03 - 1	CCV03	17908	5.1241	-7470A	PPB	10 Apr 2018 11:59:26	U	SMPL
CCB03 - 1	CCB03	-77	-0.0338	-7470A	PPB	10 Apr 2018 12:01:33	U	SMPL
J2270-06 - 1	SOUTHLAWN-PRECHAR-AREA-3	305	0.0758	-7470A	PPB	10 Apr 2018 12:03:41	U	SMPL
J2278-03 - 1	DRUM-2-3-6-8-9-11-13	123	0.0236	-7470A	PPB	10 Apr 2018 12:05:48	U	SMPL
J2278-05 - 1	DRUM-1-5-4-12-14-15-17	122	0.0233	-7470A	PPB	10 Apr 2018 12:07:55	U	SMPL
J2279-03 - 1	DRUM-1	109	0.0196	-7470A	PPB	10 Apr 2018 12:10:02	U	SMPL
J2280-02 - 1	S1	101	0.0173	-7470A	PPB	10 Apr 2018 12:12:10	U	SMPL
J2280-04 - 1	S2	80	0.0113	-7470A	PPB	10 Apr 2018 12:14:17	U	SMPL
J2283-02 - 1	SOIL-CUTTINGS	82	0.0118	-7470A	PPB	10 Apr 2018 12:16:25	U	SMPL
J2270-02LX5 - 1		98	0.0164	-7470A	PPB	10 Apr 2018 12:18:34	U	SMPL
J2270-02A - 1		13576	3.8818	-7470A	PPB	10 Apr 2018 12:20:42	U	SMPL
CCV04 - 1	CCV04	17826	5.1006	-7470A	PPB	10 Apr 2018 12:22:49	U	SMPL
CCB04 - 1	CCB04	-18	-0.0168	-7470A	PPB	10 Apr 2018 12:24:56	U	SMPL