



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB94499

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.2	4.01	04/10/2018	11:10
2	CAL2	1	Water			20.3	7.00	04/10/2018	11:12
3	CAL3	1	Water			20.3	10.02	04/10/2018	11:13
4	ICV	1	Water			20.2	6.99	04/10/2018	11:15
5	CCV1	1	Water			20.1	2.01	04/10/2018	11:16
6	J2311-01	1	Water			19.8	4.16	04/10/2018	11:19
7	J2311-01DUP	1	Water			19.7	4.17	04/10/2018	11:20
8	CCV2	1	Water			20.4	12.02	04/10/2018	11:21

LB94499

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH J2311 WorkList ID : 110610 Date : 4/10/2018 10:54:23 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/02/2018	Water	J2311-01 -A	pH	Cool 4 deg C	STER11	L53	POOLWATER	04/02/2018	SM 4500-PH B

Date/Time 04-10-18 11:00 A.M.
Received by: JL
Relinquished by: CR

Date/Time 04-10-18 12:00 PM
Received by: LP
Relinquished by: JP