



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB94505

Slope : 97.2

pH Meter ID : wc ph meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.2	4.01	04/10/2018	13:05
2	CAL2	1	Water			20.2	7.01	04/10/2018	13:06
3	CAL3	1	Water			20.3	10.02	04/10/2018	13:07
4	ICV	1	Water			20.1	6.99	04/10/2018	13:09
5	CCV1	1	Water			20.3	2.01	04/10/2018	13:10
6	J2314-02	1	Water			16.2	8.18	04/10/2018	13:11
7	J2314-02DUP	1	Water			16.3	8.19	04/10/2018	13:12
8	CCV2	1	Water			20.3	12.01	04/10/2018	13:13

2394505

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity J2314 WorkList ID : 110615 Date : 4/10/2018 12:57:03 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/09/2018	Water	J2314-02-13	Corrosivity	Cool 4 deg C	TETR06	L31	CA-AQIDW01-20180409	04/09/2018	9040C

Date/Time 04.10.18 11:00 PM
Received by: JB
Relinquished by: CD

Date/Time 04.10.18 13:20 PM
Received by: SP
Relinquished by: SP