

=====

Test results

Aquakem 7.2AQ1

2B94510
Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

4/10/2018 14:03

Reactive
Test: Total CN ug/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	94.046	0.0	0.092	
ICB1	-2.126	0.0	0.003	
CCV1	237.480	0.0	0.225	
CCB1	-1.744	0.0	0.004	
PB108127BL	-1.749	0.0	0.004	
J2156-02	-1.558	0.0	0.004	
J2156-02DUP	-2.185	0.0	0.003	
J2297-04	-1.825	0.0	0.004	
J2302-03	-1.463	0.0	0.004	
J2303-03	-2.022	0.0	0.003	
J2304-12	-1.561	0.0	0.004	
CCV2	254.250	0.0	0.241	
CCB2	-2.195	0.0	0.003	

N	13
Mean	43.642
SD	93.6123
CV%	214.50

AS

Aquakem v. 7.2AQ1

Results from time period:

Tue Apr 10 11:43:40 2018

Tue Apr 10 14:03:00 2018

Sample Id	Sam/Ctr/c/	Test short nan	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	-1.5295	µg/l	4/10/2018 13:11:03	
5.OPPBCN	A	Reactive CN	P	3.5175	µg/l	4/10/2018 13:11:04	
10PPBCN	A	Reactive CN	P	8.0363	µg/l	4/10/2018 13:11:05	
50PPBCN	A	Reactive CN	P	49.4096	µg/l	4/10/2018 13:11:06	
100PPBCN	A	Reactive CN	P	103.2958	µg/l	4/10/2018 13:11:07	
250PPBCN	A	Reactive CN	P	255.6325	µg/l	4/10/2018 13:11:08	
500PPBCN	A	Reactive CN	P	496.6377	µg/l	4/10/2018 13:11:09	
ICV1	S	Reactive CN	P	94.0462	µg/l	4/10/2018 13:59:36	
ICB1	S	Reactive CN	P	-2.1258	µg/l	4/10/2018 13:59:37	
CCV1	S	Reactive CN	P	237.4805	µg/l	4/10/2018 13:59:38	
CCB1	S	Reactive CN	P	-1.7443	µg/l	4/10/2018 13:59:39	
PB108127BL	S	Reactive CN	P	-1.7486	µg/l	4/10/2018 13:59:40	
J2156-02	S	Reactive CN	P	-1.5577	µg/l	4/10/2018 13:59:41	
J2156-02DUP	S	Reactive CN	P	-2.185	µg/l	4/10/2018 13:59:42	
J2297-04	S	Reactive CN	P	-1.8249	µg/l	4/10/2018 13:59:43	
J2302-03	S	Reactive CN	P	-1.4631	µg/l	4/10/2018 13:59:44	
J2303-03	S	Reactive CN	P	-2.0221	µg/l	4/10/2018 13:59:45	
J2304-12	S	Reactive CN	P	-1.5614	µg/l	4/10/2018 13:59:46	
CCV2	S	Reactive CN	P	254.2502	µg/l	4/10/2018 14:02:59	
CCB2	S	Reactive CN	P	-2.195	µg/l	4/10/2018 14:03:00	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

4/10/2018 13:23


Test Total CN

Accepted 4/10/2018 13:23

Factor 1080
Bias 0.005

Coeff. of det. 0.999695

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	-1.5295	0.0000	
2	5.0PPBCN	0.009	3.5175	5.0000	
3	10PPBCN	0.013	8.0363	10.0000	
4	50PPBCN	0.051	49.4096	50.0000	
5	100PPBCN	0.101	103.2958	100.0000	
6	250PPBCN	0.242	255.6325	250.0000	
7	500PPBCN	0.465	496.6377	500.0000	