


7471B

LB94517
INSTRUMENT ID: CV4

Linear

 μ Abs.:

42802



A= 0.0000e+000

B= 2.3298e-004

C= -2.5335e-002

Rho= 0.9999532

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.001	-0.001	105	0.000	105				
0.2	0.200	0.184	-0.016	900	0.0 %	900				
2.5	2.500	2.496	-0.004	10820	0.0 %	10820				
5.0	5.000	5.007	0.007	21600	0.0 %	21600				
7.5	7.500	7.567	0.067	32588	0.0 %	32588				
10	10.000	9.947	-0.053	42802	0.0 %	42802				

LB94517 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 10 Apr 2018 15:38:29

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	105	-	0.0000	7471B	PPB	10 Apr 2018 15:41:21	S	Std
0.2 - 1	50.2	900	-	0.2000	7471B	PPB	10 Apr 2018 15:43:38	S	Std
2.5 - 1	52.5	10820	-	2.5000	7471B	PPB	10 Apr 2018 15:45:56	S	Std
5.0 - 1	55	21600	-	5.0000	7471B	PPB	10 Apr 2018 15:48:13	S	Std
7.5 - 1	57.5	32588	-	7.5000	7471B	PPB	10 Apr 2018 15:50:30	S	Std
10 - 1	510	42802	-	10.0000	7471B	PPB	10 Apr 2018 15:55:30	S	Std
ICV87 - 1	ICV87	17142	3.9685	-7471B	PPB	PPB	10 Apr 2018 15:58:06	U	SMPL
ICB87 - 1	ICB87	-75	-0.0428	-7471B	PPB	PPB	10 Apr 2018 16:00:22	U	SMPL
CCV08 - 1	CCV08	21663	5.0218	-7471B	PPB	PPB	10 Apr 2018 16:02:39	U	SMPL
CCB08 - 1	CCB08	-80	-0.0440	-7471B	PPB	PPB	10 Apr 2018 16:04:55	U	SMPL
CRA - 1	CRA	877	0.1790	-7471B	PPB	PPB	10 Apr 2018 16:07:11	U	SMPL
HighStd - 1	HighStd	43030	9.9999	-7471B	PPB	PPB	10 Apr 2018 16:09:27	U	SMPL
ChkStd - 1	ChkStd	31302	7.2675	-7471B	PPB	PPB	10 Apr 2018 16:11:43	U	SMPL
PB108115BL - 1	PBS	-55	-0.0381	-7471B	PPB	PPB	10 Apr 2018 16:14:01	U	SMPL
PB108115BS - 1	LCSS	15612	3.6120	-7471B	PPB	PPB	10 Apr 2018 16:16:20	U	SMPL
J2156-01 - 1	HD-01-040918	2250	0.4989	-7471B	PPB	PPB	10 Apr 2018 16:18:37	U	SMPL
J2282-01 - 1	EPE-L-6(20-22)	480	0.0865	-7471B	PPB	PPB	10 Apr 2018 16:20:54	U	SMPL
J2282-02 - 1	EPE-L-6(25-27)	460	0.0818	-7471B	PPB	PPB	10 Apr 2018 16:23:11	U	SMPL
J2285-01 - 1	SPRING-18	931	0.1916	-7471B	PPB	PPB	10 Apr 2018 16:25:29	U	SMPL
J2297-01 - 1	RO-1-3	3477	0.7847	-7471B	PPB	PPB	10 Apr 2018 16:27:45	U	SMPL
CCV09 - 1	CCV09	21751	5.0423	-7471B	PPB	PPB	10 Apr 2018 16:30:00	U	SMPL
CCB09 - 1	CCB09	-30	-0.0323	-7471B	PPB	PPB	10 Apr 2018 16:32:16	U	SMPL
J2297-01DUP - 1	RO-1-3DUP	3241	0.7298	-7471B	PPB	PPB	10 Apr 2018 16:34:31	U	SMPL
J2297-01MS - 1	RO-1-3MS	18861	4.3690	-7471B	PPB	PPB	10 Apr 2018 16:36:48	U	SMPL
J2297-01MSD - 1	RO-1-3MSD	19285	4.4677	-7471B	PPB	PPB	10 Apr 2018 16:39:04	U	SMPL
J2301-01 - 1	EPE-L-6(0-5)	7217	1.6561	-7471B	PPB	PPB	10 Apr 2018 16:41:20	U	SMPL
J2301-02 - 1	EPE-L-6(5-10)	858	0.1746	-7471B	PPB	PPB	10 Apr 2018 16:43:37	U	SMPL
J2302-01 - 1	RT4235-RT3407	2653	0.5928	-7471B	PPB	PPB	10 Apr 2018 16:45:54	U	SMPL
J2303-01 - 1	M11	5234	1.1941	-7471B	PPB	PPB	10 Apr 2018 16:48:11	U	SMPL
J2304-01 - 1	SP-1	1602	0.3479	-7471B	PPB	PPB	10 Apr 2018 16:50:28	U	SMPL
J2297-01LX5 - 1		681	0.1333	-7471B	PPB	PPB	10 Apr 2018 16:52:46	U	SMPL
J2297-01A - 1		20100	4.6576	-7471B	PPB	PPB	10 Apr 2018 16:55:02	U	SMPL
CCV10 - 1	CCV10	21767	5.0460	-7471B	PPB	PPB	10 Apr 2018 16:57:18	U	SMPL
CCB10 - 1	CCB10	-44	-0.0356	-7471B	PPB	PPB	10 Apr 2018 16:59:35	U	SMPL