



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB94521

Slope : 97.8

pH Meter ID : wc ph meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.5	3.99	04/10/2018	17:00
2	CAL2	1	Water			20.5	7.01	04/10/2018	17:01
3	CAL3	1	Water			20.5	10.02	04/10/2018	17:02
4	ICV	1	Water			20.4	7.00	04/10/2018	17:04
5	CCV1	1	Water			20.3	2.01	04/10/2018	17:05
6	J2318-06	1	Water			18.9	7.61	04/10/2018	17:06
7	J2318-06DUP	1	Water			18.8	7.62	04/10/2018	17:08
8	CCV2	1	Water			20.6	12.02	04/10/2018	17:10

LB 94521

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph J2318

WorkList ID : 110624

Date : 4/10/2018 4:42:43 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/10/2018	Water	J2318-06 V	pH	Cool 4 deg C	CARP01	L13	CLEARWELL	04/10/2018	SM 4500-PH B

Date/Time 04-10-18 4:50 PM
 Received by: SO
 Relinquished by: CO

Date/Time 04-10-18 5:15 PM
 Received by: CO
 Relinquished by: SO