

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 04/11/2018

TEMP1 IN: 103 04/11/2018 07:10 TEMP OUT: 104 04/11/2018 08:20

TEMP2 IN: 104 04/11/2018 09:00 TEMP OUT: 104 04/11/2018 10:20

TEMP3 IN: 105 04/11/2018 12:20 TEMP OUT: 105 04/11/2018 14:10

TEMP4 IN: 105 04/11/2018 14:55 TEMP OUT: 105 04/11/2018 16:10

Run Number: LB94523

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB94523BL	LB94523BL	1.3859	1.3859	100	1.3859	1.3859	1.3859	0.0000	0
2	LB94523BS	LB94523BS	1.3909	1.3909	100	1.4455	1.4455	1.4455	0.0546	546
3	J2321-01	041018-JETB-E-D-B	1.3944	1.3944	200	1.4116	1.4116	1.4116	0.0172	86
4	J2321-01DUP	041018-JETB-E-D-BDUP	1.3949	1.3949	200	1.4120	1.4120	1.4120	0.0171	85.5
5	J2321-02	041018-JETB-E-D-M	1.3864	1.3864	300	1.3954	1.3954	1.3954	0.0090	30
6	J2321-03	041018-JETB-E-D-S	1.3872	1.3872	300	1.3914	1.3914	1.3914	0.0042	14
7	J2321-04	041018-JETB-E-U-B	1.3884	1.3884	300	1.4057	1.4057	1.4057	0.0173	57.7
8	J2321-05	041018-JETB-E-U-M	1.3962	1.3962	400	1.4028	1.4028	1.4028	0.0066	16.5
9	J2321-06	041018-JETB-E-U-S	1.3945	1.3945	400	1.3998	1.3998	1.3998	0.0053	13.3
10	J2321-07	041018-DBRI-E-D-B	1.3939	1.3939	400	1.3995	1.3995	1.3995	0.0056	14
11	J2321-08	041018-DBRI-E-D-S	1.3942	1.3942	400	1.4007	1.4007	1.4007	0.0065	16.3
12	J2321-09	041018-DBRI-E-U-B	1.4019	1.4019	400	1.4077	1.4077	1.4077	0.0058	14.5
13	J2321-10	041018-DBRI-E-U-S	1.3869	1.3869	400	1.3926	1.3926	1.3926	0.0057	14.3



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatelANALYST: AparnaDate: 04/11/2018TEMP1 IN: 103 04/11/2018 07:10 TEMP OUT: 104 04/11/2018 08:20TEMP2 IN: 104 04/11/2018 09:00 TEMP OUT: 104 04/11/2018 10:20TEMP3 IN: 105 04/11/2018 12:20 TEMP OUT: 105 04/11/2018 14:10TEMP4 IN: 105 04/11/2018 14:55 TEMP OUT: 105 04/11/2018 16:10Run Number: LB94523BalanceID: WC SC-3OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
A = Sample Volume (ml)						Weight (g) = C - B				
B = Empty Dish Weight (g)						Result mg/L = $\frac{D}{A} \times 1000 \times 1000$				
C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C (g)										
D = Weight (g)										

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS041118-2

WorkList ID : 110643

Date : 4/11/2018 8:16:25 AM

2394523

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/17/2018	Water	J2321-01 C	TSS	Cool 4 deg C	HDR103	L22	041018-JETB-E-D-B	04/10/2018	SM2540 D
04/17/2018	Water	J2321-02 C	TSS	Cool 4 deg C	HDR103	L22	041018-JETB-E-D-M	04/10/2018	SM2540 D
04/17/2018	Water	J2321-03 C	TSS	Cool 4 deg C	HDR103	L22	041018-JETB-E-D-S	04/10/2018	SM2540 D
04/17/2018	Water	J2321-04 C	TSS	Cool 4 deg C	HDR103	L22	041018-JETB-E-U-B	04/10/2018	SM2540 D
04/17/2018	Water	J2321-05 C	TSS	Cool 4 deg C	HDR103	L22	041018-JETB-E-U-M	04/10/2018	SM2540 D
04/17/2018	Water	J2321-06 C	TSS	Cool 4 deg C	HDR103	L22	041018-JETB-E-U-S	04/10/2018	SM2540 D
04/17/2018	Water	J2321-07 C	TSS	Cool 4 deg C	HDR103	L22	041018-DBRI-E-D-B	04/10/2018	SM2540 D
04/17/2018	Water	J2321-08 C	TSS	Cool 4 deg C	HDR103	L22	041018-DBRI-E-D-S	04/10/2018	SM2540 D
04/17/2018	Water	J2321-09 C	TSS	Cool 4 deg C	HDR103	L22	041018-DBRI-E-U-B	04/10/2018	SM2540 D
04/17/2018	Water	J2321-10 C	TSS	Cool 4 deg C	HDR103	L22	041018-DBRI-E-U-S	04/10/2018	SM2540 D

Date/Time 04/11/18, 08:30
 Received by: AS
 Relinquished by: DP

Date/Time 04/11/18, 12:40
 Received by: DP
 Relinquished by: AS