


7470A

LB94524
INSTRUMENT ID: CV1

Linear

 μ Abs.:

34987



A= 0.0000e+000

B= 2.8890e-004

C= 1.5276e-002

Rho= 0.9997564

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.039	0.039	83	0.000	83				
0.2	0.200	0.218	0.018	703	0.0 %	703				
2.5	2.500	2.526	0.026	8690	0.0 %	8690				
5.0	5.000	4.916	-0.084	16963	0.0 %	16963				
7.5	7.500	7.377	-0.123	25483	0.0 %	25483				
10.0	10.000	10.123	0.123	34987	0.0 %	34987				

LB94524 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 11 Apr 2018 10:22:06

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc.	Method	Units	Date	Type	Type
0.0 - 1	50	83	-	0.0000	7470A	PPB	11 Apr 2018 10:25:39	S	Std
0.2 - 1	50.2	703	-	0.2000	7470A	PPB	11 Apr 2018 10:27:48	S	Std
2.5 - 1	52.5	8690	-	2.5000	7470A	PPB	11 Apr 2018 10:29:56	S	Std
5.0 - 1	55	16963	-	5.0000	7470A	PPB	11 Apr 2018 10:32:04	S	Std
7.5 - 1	57.5	25483	-	7.5000	7470A	PPB	11 Apr 2018 10:34:12	S	Std
10.0 - 1	510	34987	-	10.0000	7470A	PPB	11 Apr 2018 10:38:47	S	Std
ICV88 - 1	ICV88	13589	3.9412	-	7470A	PPB	11 Apr 2018 10:41:19	U	SMPL
ICB88 - 1	ICB88	-75	-0.0064	-	7470A	PPB	11 Apr 2018 10:43:25	U	SMPL
CCV11 - 1	CCV11	16982	4.9214	-	7470A	PPB	11 Apr 2018 10:45:33	U	SMPL
CCB11 - 1	CCB11	-84	-0.0090	-	7470A	PPB	11 Apr 2018 10:47:40	U	SMPL
CRA - 1	CRA	686	0.2135	-	7470A	PPB	11 Apr 2018 10:49:47	U	SMPL
HighStd - 1	HighStd	34868	10.0888	-	7470A	PPB	11 Apr 2018 10:51:53	U	SMPL
ChkStd - 1	ChkStd	24435	7.0746	-	7470A	PPB	11 Apr 2018 10:54:00	U	SMPL
PB108141BL - 1	PBW	-219	-0.0480	-	7470A	PPB	11 Apr 2018 10:56:09	U	SMPL
PB108141BS - 1	LCSW	13180	3.8230	-	7470A	PPB	11 Apr 2018 10:58:18	U	SMPL
PB108141TB - 1	PB108113TB	-106	-0.0153	-	7470A	PPB	11 Apr 2018 11:00:26	U	SMPL
J2156-02 - 1	HD-01-040918	-6	0.0135	-	7470A	PPB	11 Apr 2018 11:02:33	U	SMPL
J2297-04 - 1	RO-3-1	15	0.0196	-	7470A	PPB	11 Apr 2018 11:04:42	U	SMPL
J2302-03 - 1	RT4235-RT3407	60	0.0326	-	7470A	PPB	11 Apr 2018 11:06:50	U	SMPL
J2303-03 - 1	M11	95	0.0427	-	7470A	PPB	11 Apr 2018 11:08:56	U	SMPL
CCV12 - 1	CCV12	17145	4.9685	-	7470A	PPB	11 Apr 2018 11:11:03	U	SMPL
CCB12 - 1	CCB12	-77	-0.0070	-	7470A	PPB	11 Apr 2018 11:13:10	U	SMPL
J2304-12 - 1	SP-1	31	0.0242	-	7470A	PPB	11 Apr 2018 11:15:17	U	SMPL
J2306-01 - 1	RT3075	45	0.0283	-	7470A	PPB	11 Apr 2018 11:17:24	U	SMPL
J2306-03 - 1	CHRT26884	17	0.0202	-	7470A	PPB	11 Apr 2018 11:19:31	U	SMPL
J2306-04 - 1	RB36275	10	0.0182	-	7470A	PPB	11 Apr 2018 11:21:38	U	SMPL
J2306-04DUP - 1	RB36275DUP	11	0.0185	-	7470A	PPB	11 Apr 2018 11:23:46	U	SMPL
J2306-04MS - 1	RB36275MS	13032	3.7803	-	7470A	PPB	11 Apr 2018 11:25:53	U	SMPL
J2306-04MSD - 1	RB36275MSD	13123	3.8066	-	7470A	PPB	11 Apr 2018 11:28:02	U	SMPL
J2306-04LX5 - 1		-49	0.0011	-	7470A	PPB	11 Apr 2018 11:30:10	U	SMPL
J2306-04A - 1		13748	3.9871	-	7470A	PPB	11 Apr 2018 11:32:18	U	SMPL
CCV13 - 1	CCV13	17262	5.0023	-	7470A	PPB	11 Apr 2018 11:34:25	U	SMPL
CCB13 - 1	CCB13	-64	-0.0032	-	7470A	PPB	11 Apr 2018 11:36:32	U	SMPL