



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB94526

Ambient Barometric Pressure (mmHg): 765

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	78.9	04/11/2018	08:50
2	J2314-02		1	>100.0	>212.0	>212.0	04/11/2018	09:15
3	J2314-02DUP		1	>100.0	>212.0	>212.0	04/11/2018	09:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.33

LB94526

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP-041118

WorkList ID : 110640

Date : 4/11/2018 7:57:20 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/17/2018	Water	J2314-02	Flash Point	Cool 4 deg C	TETR06	L31	CA-AQIDW01-20180409	04/09/2018	1010A

Date/Time 04/11/18 8:15 AM
Received by: KS
Relinquished by: CP

Date/Time 04/11/18 09:50 AM
Received by: CP
Relinquished by: KS