



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB94527

Slope : 99.2

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.5	4.01	04/11/2018	11:10
2	CAL2	1	Water			20.2	7.01	04/11/2018	11:12
3	CAL3	1	Water			20.3	10.03	04/11/2018	11:14
4	ICV	1	Water			20.5	7.00	04/11/2018	11:16
5	CCV1	1	Water			20.3	2.00	04/11/2018	11:18
6	J2326-01	1	Water			12.6	4.46	04/11/2018	11:20
7	J2326-01DUP	1	Water			12.6	4.47	04/11/2018	11:22
8	J2326-02	1	Water			12.7	4.57	04/11/2018	11:24
9	J2326-03	1	Water			14.0	5.63	04/11/2018	11:26
10	CCV2	1	Water			20.5	12.00	04/11/2018	11:28

WORKLIST(Hardcopy Internal Chain)

LB94527

Worklist Name : pH J2326

Worklist ID : 110651

Date : 4/11/2018 11:01:33 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/10/2018	Water	J2326-01	pH	Cool 4 deg C	STER01	N11	CELL-A-POOL-WATER(DE-1C	04/10/2018	SM 4500-PH B
04/10/2018	Water	J2326-02	pH	Cool 4 deg C	STER01	N11	CELL-B-C-POOL-WATER(DE	04/10/2018	SM 4500-PH B
04/10/2018	Water	J2326-03	pH	Cool 4 deg C	STER01	N11	CELL-D-POOL-WATER(DE-1C	04/10/2018	SM 4500-PH B

Date/Time 04/11/18 11:05
 Received by: SP
 Relinquished by: SP

Date/Time 04/11/18 11:35
 Received by: SP
 Relinquished by: SP