



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB94546

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : wc ph meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.5	4.01	04/11/2018	17:22
2	CAL2	1	Water			20.5	7.00	04/11/2018	17:23
3	CAL3	1	Water			20.5	10.02	04/11/2018	17:24
4	ICV	1	Water			20.6	6.99	04/11/2018	17:25
5	CCV1	1	Water			20.4	2.01	04/11/2018	17:26
6	J2162-02	1	Solid	20.02	20	20.1	7.21	04/11/2018	17:28
7	J2162-02DUP	1	Solid	20.03	20	20.2	7.22	04/11/2018	17:30
8	J2162-04	1	Solid	20.02	20	20.7	9.21	04/11/2018	17:31
9	J2162-06	1	Solid	20.01	20	20.4	8.78	04/11/2018	17:32
10	J2163-02	1	Solid	20.03	20	20.2	8.63	04/11/2018	17:35
11	J2163-04	1	Solid	20.02	20	20.4	8.18	04/11/2018	17:36
12	J2163-06	1	Solid	20.03	20	20.3	7.45	04/11/2018	17:37
13	J2163-08	1	Solid	20.01	20	20.6	7.99	04/11/2018	17:38
14	J2165-04	1	Solid	20.04	20	20.4	7.85	04/11/2018	17:40
15	J2335-02	1	Solid	20.02	20	20.5	7.21	04/11/2018	17:42
16	CCV2	1	Water			20.1	12.01	04/11/2018	17:44
17	J2335-04	1	Solid	20.05	20	20.4	8.18	04/11/2018	17:45
18	J2335-06	1	Solid	20.02	20	20.0	7.45	04/11/2018	17:46
19	CCV3	1	Water			20.2	2.01	04/11/2018	17:48

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity j2335

WorkList ID : 110663

Date : 4/11/2018 2:46:53 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/11/2018	Solid	J2162-02	Corrosivity	Cool 4 deg C	PSEG05	N63	EO-01-041118-A	04/11/2018	9045C
04/11/2018	Solid	J2162-04	Corrosivity	Cool 4 deg C	PSEG05	N63	EO-01-041118-B	04/11/2018	9045C
04/11/2018	Solid	J2162-06	Corrosivity	Cool 4 deg C	PSEG05	N63	EO-01-041118-C	04/11/2018	9045C
04/11/2018	Solid	J2163-02	Corrosivity	Cool 4 deg C	PSEG05	N62	PL-01-041118-A	04/11/2018	9045C
04/11/2018	Solid	J2163-04	Corrosivity	Cool 4 deg C	PSEG05	N62	PL-01-041118-B	04/11/2018	9045C
04/11/2018	Solid	J2163-06	Corrosivity	Cool 4 deg C	PSEG05	N62	PL-01-041118-C	04/11/2018	9045C
04/11/2018	Solid	J2163-08	Corrosivity	Cool 4 deg C	PSEG05	N62	PL-01-041118-D	04/11/2018	9045C
04/11/2018	Solid	J2165-04	Corrosivity	Cool 4 deg C	PSEG05		SU-03-041118	04/11/2018	9045C
04/11/2018	Solid	J2335-02	Corrosivity	Cool 4 deg C	PSEG01	N62	S1	04/11/2018	9045C
04/11/2018	Solid	J2335-04	Corrosivity	Cool 4 deg C	PSEG01	N62	S2	04/11/2018	9045C
04/11/2018	Solid	J2335-06	Corrosivity	Cool 4 deg C	PSEG01	N62	S3	04/11/2018	9045C

Date/Time 4-11-18 2:00 PM
 Received by: SP
 Relinquished by: CB

Date/Time 4-11-18 6:30 PM
 Received by: CB
 Relinquished by: JP