

245.1


LB 94553

INSTRUMENT ID: CV2

Linear

 μ Abs.:

3406E



A= 0.0000e+000

B= 2.9650e-004

C= -3.7099e-003

Rho= 0.9998468

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.008	0.008	40	0.000	40				
0.05	0.050	0.060	0.010	214	0.0 %	214				
0.2	0.200	0.205	0.005	704	0.0 %	704				
2.5	2.500	2.521	0.021	8514	0.0 %	8514				
5.0	5.000	4.999	-0.001	16873	0.0 %	16873				
7.5	7.500	7.360	-0.140	24836	0.0 %	24836				
10.0	10.000	10.097	0.097	34066	0.0 %	34066				

LB94553 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 12 Apr 2018 10:08:33

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc.	Method	Units	Date	Type	Type
0.0 - 1	50	40	-	0.0000	245.1	PPB	12 Apr 2018 10:12:06	S	Std
0.05 - 1	50.02	214	-	0.0500	245.1	PPB	12 Apr 2018 10:14:15	S	Std
0.2 - 1	50.2	704	-	0.2000	245.1	PPB	12 Apr 2018 10:16:24	S	Std
2.5 - 1	52.5	8514	-	2.5000	245.1	PPB	12 Apr 2018 10:18:33	S	Std
5.0 - 1	55	16873	-	5.0000	245.1	PPB	12 Apr 2018 10:20:41	S	Std
7.5 - 1	57.5	24836	-	7.5000	245.1	PPB	12 Apr 2018 10:22:48	S	Std
10.0 - 1	510	34066	-	10.0000	245.1	PPB	12 Apr 2018 10:24:56	S	Std
ICV89 - 1	ICV89	13196	3.9089	-245.1		PPB	12 Apr 2018 10:27:34	U	SMPL
ICB89 - 1	ICB89	-100	-0.0334	-245.1		PPB	12 Apr 2018 10:29:41	U	SMPL
CCV14 - 1	CCV14	17009	5.0395	-245.1		PPB	12 Apr 2018 10:31:48	U	SMPL
CCB14 - 1	CCB14	-100	-0.0334	-245.1		PPB	12 Apr 2018 10:33:55	U	SMPL
CRA - 1	CRA	207	0.0577	-245.1		PPB	12 Apr 2018 10:36:03	U	SMPL
HighStd - 1	HighStd	34148	10.1213	-245.1		PPB	12 Apr 2018 10:38:09	U	SMPL
ChkStd - 1	ChkStd	23384	6.9297	-245.1		PPB	12 Apr 2018 10:40:16	U	SMPL
PB108162BL - 1	PBW	-201	-0.0633	-245.1		PPB	12 Apr 2018 10:42:26	U	SMPL
PB108162BS - 1	LCSW	13470	3.9902	-245.1		PPB	12 Apr 2018 10:44:35	U	SMPL
J2320-01 - 1	040918-DBRI-E-D-B	-107	-0.0354	-245.1		PPB	12 Apr 2018 10:46:43	U	SMPL
J2320-02 - 1	040918-DBRI-E-D-S	-11	-0.0070	-245.1		PPB	12 Apr 2018 10:48:51	U	SMPL
J2320-03 - 1	040918-DBRI-E-U-B	2	-0.0031	-245.1		PPB	12 Apr 2018 10:50:59	U	SMPL
J2320-04 - 1	040918-DBRI-E-U-S	12	-0.0002	-245.1		PPB	12 Apr 2018 10:53:07	U	SMPL
J2320-05 - 1	040918-JETB-E-D-B	59	0.0138	-245.1		PPB	12 Apr 2018 10:55:14	U	SMPL
CCV15 - 1	CCV15	16704	4.9491	-245.1		PPB	12 Apr 2018 10:57:21	U	SMPL
CCB15 - 1	CCB15	-96	-0.0322	-245.1		PPB	12 Apr 2018 10:59:28	U	SMPL
J2320-06 - 1	040918-JETB-E-D-M	3	-0.0028	-245.1		PPB	12 Apr 2018 11:01:36	U	SMPL
J2320-07 - 1	040918-JETB-E-D-S	6	-0.0019	-245.1		PPB	12 Apr 2018 11:03:43	U	SMPL
J2320-08 - 1	040918-JETB-E-U-B	39	0.0079	-245.1		PPB	12 Apr 2018 11:05:50	U	SMPL
J2320-09 - 1	040918-JETB-E-U-M	14	0.0004	-245.1		PPB	12 Apr 2018 11:07:57	U	SMPL
J2320-10 - 1	040918-JETB-E-U-S	4	-0.0025	-245.1		PPB	12 Apr 2018 11:10:05	U	SMPL
J2321-01 - 1	041018-JETB-E-D-B	35	0.0067	-245.1		PPB	12 Apr 2018 11:12:13	U	SMPL
J2321-02 - 1	041018-JETB-E-D-M	33	0.0061	-245.1		PPB	12 Apr 2018 11:14:21	U	SMPL
J2321-03 - 1	041018-JETB-E-D-S	12	-0.0002	-245.1		PPB	12 Apr 2018 11:16:29	U	SMPL
J2321-04 - 1	041018-JETB-E-U-B	32	0.0058	-245.1		PPB	12 Apr 2018 11:18:37	U	SMPL
J2321-05 - 1	041018-JETB-E-U-M	2	-0.0031	-245.1		PPB	12 Apr 2018 11:20:44	U	SMPL
CCV16 - 1	CCV16	17046	5.0505	-245.1		PPB	12 Apr 2018 11:22:51	U	SMPL
CCB16 - 1	CCB16	-106	-0.0351	-245.1		PPB	12 Apr 2018 11:24:58	U	SMPL
J2321-06 - 1	041018-JETB-E-U-S	-18	-0.0090	-245.1		PPB	12 Apr 2018 11:27:06	U	SMPL
J2321-07 - 1	041018-DBRI-E-D-B	5	-0.0022	-245.1		PPB	12 Apr 2018 11:29:14	U	SMPL
J2321-08 - 1	041018-DBRI-E-D-S	4	-0.0025	-245.1		PPB	12 Apr 2018 11:31:21	U	SMPL
J2321-09 - 1	041018-DBRI-E-U-B	11	-0.0004	-245.1		PPB	12 Apr 2018 11:33:28	U	SMPL
J2321-10 - 1	041018-DBRI-E-U-S	-1	-0.0040	-245.1		PPB	12 Apr 2018 11:35:35	U	SMPL
J2321-10DUP - 1	041018-DBRI-E-U-SDUP	-18	-0.0090	-245.1		PPB	12 Apr 2018 11:37:43	U	SMPL
J2321-10MS - 1	041018-DBRI-E-U-SMS	13080	3.8745	-245.1		PPB	12 Apr 2018 11:39:51	U	SMPL
J2321-10MSD - 1	041018-DBRI-E-U-SMSD	13154	3.8965	-245.1		PPB	12 Apr 2018 11:41:59	U	SMPL
J2321-10LX5 - 1		-82	-0.0280	-245.1		PPB	12 Apr 2018 11:44:09	U	SMPL
J2321-10A - 1		13310	3.9427	-245.1		PPB	12 Apr 2018 11:46:15	U	SMPL
CCV17 - 1	CCV17	16885	5.0027	-245.1		PPB	12 Apr 2018 11:48:22	U	SMPL
CCB17 - 1	CCB17	-119	-0.0390	-245.1		PPB	12 Apr 2018 11:50:30	U	SMPL