


7470A

LB94557
INSTRUMENT ID: CV1

Linear

 μ Abs.:

35698



A= 0.0000e+000

B= 2.7793e-004

C= -2.8332e-002

Rho= 0.9998286

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.023	-0.023	18	0.000	18				
0.2	0.200	0.177	-0.023	738	0.0 %	738				
2.5	2.500	2.480	-0.020	9025	0.0 %	9025				
5.0	5.000	5.082	0.082	18388	0.0 %	18388				
7.5	7.500	7.591	0.091	27416	0.0 %	27416				
10.0	10.000	9.893	-0.107	35698	0.0 %	35698				

LB94557 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 12 Apr 2018 12:23:04

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	18	-	0.0000	7470A	PPB	12 Apr 2018 12:25:00	S	Std
0.2 - 1	50.2	738	-	0.2000	7470A	PPB	12 Apr 2018 12:27:09	S	Std
2.5 - 1	32.5	9025	-	2.5000	7470A	PPB	12 Apr 2018 12:29:17	S	Std
5.0 - 1	55	18388	-	5.0000	7470A	PPB	12 Apr 2018 12:31:26	S	Std
7.5 - 1	57.5	27416	-	7.5000	7470A	PPB	12 Apr 2018 12:33:33	S	Std
10.0 - 1	510	35698	-	10.0000	7470A	PPB	12 Apr 2018 12:35:41	S	Std
ICV90 - 1	ICV90	13915	3.8390	-	7470A	PPB	12 Apr 2018 12:38:40	U	SMPL
ICB90 - 1	ICB90	-155	-0.0714	-	7470A	PPB	12 Apr 2018 12:40:47	U	SMPL
CCV18 - 1	CCV18	18318	5.0627	-	7470A	PPB	12 Apr 2018 12:42:54	U	SMPL
CCB18 - 1	CCB18	-151	-0.0703	-	7470A	PPB	12 Apr 2018 12:45:00	U	SMPL
CRA - 1	CRA	705	0.1676	-	7470A	PPB	12 Apr 2018 12:47:08	U	SMPL
HighStd - 1	HighStd	35497	9.8372	-	7470A	PPB	12 Apr 2018 12:49:14	U	SMPL
ChkStd - 1	ChkStd	24990	6.9171	-	7470A	PPB	12 Apr 2018 12:51:21	U	SMPL
PB108171BL - 1	PBW	-46	-0.0411	-	7470A	PPB	12 Apr 2018 12:55:59	U	SMPL
PB108171BS - 1	LCSW	12870	3.5486	-	7470A	PPB	12 Apr 2018 12:58:06	U	SMPL
PB108171TB - 1	PB108142TB	-112	-0.0595	-	7470A	PPB	12 Apr 2018 13:00:14	U	SMPL
J2161-02 - 1	OR-02-041018-A	-14	-0.0322	-	7470A	PPB	12 Apr 2018 13:02:22	U	SMPL
J2161-02DUP - 1	OR-02-041018-ADUP	-6	-0.0300	-	7470A	PPB	12 Apr 2018 13:04:30	U	SMPL
J2161-02MS - 1	OR-02-041018-AMS	12841	3.5405	-	7470A	PPB	12 Apr 2018 13:06:39	U	SMPL
J2161-02MSD - 1	OR-02-041018-AMSD	13128	3.6203	-	7470A	PPB	12 Apr 2018 13:08:46	U	SMPL
CCV19 - 1	CCV19	18402	5.0861	-	7470A	PPB	12 Apr 2018 13:10:52	U	SMPL
CCB19 - 1	CCB19	-163	-0.0736	-	7470A	PPB	12 Apr 2018 13:13:00	U	SMPL
J2161-04 - 1	OR-02-041018-B	-27	-0.0358	-	7470A	PPB	12 Apr 2018 13:15:08	U	SMPL
J2307-04 - 1	WC-PH3-ARE-01	-13	-0.0319	-	7470A	PPB	12 Apr 2018 13:17:15	U	SMPL
J2307-05 - 1	WC-PH3-AR-02	1	-0.0281	-	7470A	PPB	12 Apr 2018 13:19:22	U	SMPL
J2307-06 - 1	WC-PH3-SZ-03	8	-0.0261	-	7470A	PPB	12 Apr 2018 13:21:29	U	SMPL
J2319-02 - 1	DRUM-1	0	-0.0283	-	7470A	PPB	12 Apr 2018 13:23:37	U	SMPL
J2161-02LX5 - 1		5	-0.0269	-	7470A	PPB	12 Apr 2018 13:25:45	U	SMPL
J2161-02A - 1		13381	3.6906	-	7470A	PPB	12 Apr 2018 13:27:53	U	SMPL
CCV20 - 1	CCV20	17744	4.9032	-	7470A	PPB	12 Apr 2018 13:30:01	U	SMPL
CCB20 - 1	CCB20	-149	-0.0697	-	7470A	PPB	12 Apr 2018 13:32:10	U	SMPL