



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB94560

Reagent/Standard	Lot/Loq #
TURBIDITY CCV 10 NTU	WP63643
Turbidity std, 40NTU	WP63638
turbidity std 5 NTU	WP63641
turbidity std 20 NTU	WP63639
Turbidity std 1 NTU	WP63642
turbidity std 0 NTU	WP63637
turbidity std 10 NTU	WP63640
TURBIDITY LOD STD, 0.5NTU	WP63644
TURBIDITY LOQ STD, 2.0NTU	WP63645

Intercept: -0.0194

Slope: 0.9038

Regression: 0.9989

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.190	04/12/2018	10:13
2	CAL2	1	1	1.090	04/12/2018	10:16
3	CAL3	5	1	4.610	04/12/2018	10:20
4	CAL4	10	1	9.180	04/12/2018	10:23
5	CAL5	20	1	16.800	04/12/2018	10:27
6	CAL6	40	1	36.700	04/12/2018	10:31



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB94560

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.280	10.289	04/12/2018	10:34
2	ICB		1	0.189	0.231	04/12/2018	10:37
3	CCV1	10	1	9.200	10.201	04/12/2018	10:57
4	CCB1		1	0.186	0.227	04/12/2018	11:00
5	LB94560BL		1	0.185	0.226	04/12/2018	11:04
6	J2133-10		1	0.809	0.917	04/12/2018	11:09
7	J2133-11		1	2.310	2.577	04/12/2018	11:12
8	J2133-12		1	0.810	0.918	04/12/2018	11:16
9	CCV2	10	1	8.780	9.736	04/12/2018	11:19
10	CCB2		1	0.190	0.232	04/12/2018	11:23