



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB94561

Slope : 97.2

pH Meter ID : wc ph meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.3	4.01	04/12/2018	12:35
2	CAL2	1	Water			20.4	7.00	04/12/2018	12:36
3	CAL3	1	Water			20.4	10.02	04/12/2018	12:37
4	ICV	1	Water			20.3	7.01	04/12/2018	12:39
5	CCV1	1	Water			20.2	2.01	04/12/2018	12:40
6	J2336-02	1	Water			18.8	5.77	04/12/2018	12:41
7	J2336-02DUP	1	Water			18.9	5.78	04/12/2018	12:42
8	CCV2	1	Water			20.3	12.02	04/12/2018	12:43

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity j12339

WorkList ID : 110706

Date : 4/12/2018 12:25:10 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/11/2018	Water	J2336-02 -A	Corrosivity	Cool 4 deg C	PSEG01	N63	SL	04/11/2018	9040C

Date/Time 04-12-18 12:30 PM
 Received by: JP
 Relinquished by: CP

Date/Time 04-12-18 1:30 PM
 Received by: JP
 Relinquished by: CP