



## Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB94562

Slope : 97.2

pH Meter ID : wc ph meter-1

| Calibration Standards          | Chemtech Log# |
|--------------------------------|---------------|
| PH 4 BUFFER SOLUTION           | W2271         |
| buffer solution pH 7 yellow    | W2159         |
| PH 10.01 BUFFER,COLOR CD 475ML | W2220         |
| buffer solution pH 7 yellow    | W2219         |
| pH 2.00 Buffer                 | W2289         |
| Buffer Solution, PH12 (500ml)  | W2267         |

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

| Seq | LabID       | DF | Matrix | Weight (gm) | Volume (ml) | Temperature (°C) | Result (pH) | Anal Date  | Anal Time |
|-----|-------------|----|--------|-------------|-------------|------------------|-------------|------------|-----------|
| 1   | CAL1        | 1  | Water  |             |             | 20.3             | 4.01        | 04/12/2018 | 12:35     |
| 2   | CAL2        | 1  | Water  |             |             | 20.4             | 7.00        | 04/12/2018 | 12:36     |
| 3   | CAL3        | 1  | Water  |             |             | 20.4             | 10.02       | 04/12/2018 | 12:37     |
| 4   | ICV         | 1  | Water  |             |             | 20.3             | 7.01        | 04/12/2018 | 12:39     |
| 5   | CCV1        | 1  | Water  |             |             | 20.2             | 2.01        | 04/12/2018 | 12:40     |
| 6   | J2339-01    | 1  | Water  |             |             | 18.8             | 5.77        | 04/12/2018 | 12:41     |
| 7   | J2339-01DUP | 1  | Water  |             |             | 18.9             | 5.78        | 04/12/2018 | 12:42     |
| 8   | CCV2        | 1  | Water  |             |             | 20.3             | 12.02       | 04/12/2018 | 12:43     |

1944562

# WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH J2339

WorkList ID : 110705

Date : 4/12/2018 12:24:45 PM

| Due Date   | Matrix | Sample      | Test | Preservative | Customer | Storage Location | Customer Sample         | Collect Date | Method |
|------------|--------|-------------|------|--------------|----------|------------------|-------------------------|--------------|--------|
| 04/11/2018 | Water  | J2339-01-1A | pH   | Cool 4 deg C | PSEG01   | N03              | SL                      | 04/11/2018   | 9040C  |
| 04/10/2018 | Water  | J2339-01-1B | pH   | Cool 4 deg C | STER08   | N11              | STERIGENICS-TUSTIN-WATI | 04/10/2018   | 9040C  |

Date/Time 04-12-18 12:30 PM  
 Received by: Jo  
 Relinquished by: CO

Date/Time 04-12-18 1:30 PM  
 Received by: CO  
 Relinquished by: SI