

7471B


LB 94567

INSTRUMENT ID: CV2

Linear

 μ Abs.:

41247



A= 0.0000e+000

B= 2.4125e-004

C= 4.0148e-003

Rho= 0.9999268

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.008	0.008	15	0.000	15				
0.2	0.200	0.219	0.019	891	0.0 %	891				
2.5	2.500	2.436	-0.064	10082	0.0 %	10082				
5.0	5.000	5.009	0.009	20745	0.0 %	20745				
7.5	7.500	7.573	0.073	31375	0.0 %	31375				
10	10.000	9.955	-0.045	41247	0.0 %	41247				

LB94567 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 12 Apr 2018 14:44:24

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	15	-	0.0000	7471B	PPB	12 Apr 2018 14:46:16	S	Std
0.2 - 1	50.2	891	-	0.2000	7471B	PPB	12 Apr 2018 14:48:33	S	Std
2.5 - 1	52.5	10082	-	2.5000	7471B	PPB	12 Apr 2018 14:50:51	S	Std
5.0 - 1	55	20745	-	5.0000	7471B	PPB	12 Apr 2018 14:53:08	S	Std
7.5 - 1	57.5	31375	-	7.5000	7471B	PPB	12 Apr 2018 14:55:25	S	Std
10 - 1	510	41247	-	10.0000	7471B	PPB	12 Apr 2018 15:00:24	S	Std
ICV91 - 1	ICV91	15884	3.8361	-	7471B	PPB	12 Apr 2018 15:03:08	U	SMPL
ICB91 - 1	ICB91	-157	-0.0339	-	7471B	PPB	12 Apr 2018 15:05:24	U	SMPL
CCV21 - 1	CCV21	20645	4.9847	-	7471B	PPB	12 Apr 2018 15:07:40	U	SMPL
CCB21 - 1	CCB21	-181	-0.0397	-	7471B	PPB	12 Apr 2018 15:09:56	U	SMPL
CRA - 1	CRA	862	0.2120	-	7471B	PPB	12 Apr 2018 15:12:13	U	SMPL
HighStd - 1	HighStd	42404	10.2341	-	7471B	PPB	12 Apr 2018 15:14:29	U	SMPL
ChkStd - 1	ChkStd	31047	7.4942	-	7471B	PPB	12 Apr 2018 15:16:45	U	SMPL
PB108172BL - 1	PBS	-225	-0.0503	-	7471B	PPB	12 Apr 2018 15:19:04	U	SMPL
PB108172BS - 1	LCSS	15084	3.6431	-	7471B	PPB	12 Apr 2018 15:21:22	U	SMPL
J2161-01 - 1	OR-02-041018-A	3118	0.7562	-	7471B	PPB	12 Apr 2018 15:23:39	U	SMPL
J2161-03 - 1	OR-02-041018-B	5648	1.3666	-	7471B	PPB	12 Apr 2018 15:25:57	U	SMPL
J2162-01 - 1	EO-01-041118-A	15293	3.6935	-	7471B	PPB	12 Apr 2018 15:28:15	U	SMPL
J2162-03 - 1	EO-01-041118-B	5135	1.2428	-	7471B	PPB	12 Apr 2018 15:30:33	U	SMPL
J2162-05 - 1	EO-01-041118-C	2285	0.5553	-	7471B	PPB	12 Apr 2018 15:32:49	U	SMPL
CCV22 - 1	CCV22	20429	4.9326	-	7471B	PPB	12 Apr 2018 15:35:05	U	SMPL
CCB22 - 1	CCB22	-124	-0.0259	-	7471B	PPB	12 Apr 2018 15:37:21	U	SMPL
J2163-01 - 1	PL-01-041118-A	2126	0.5169	-	7471B	PPB	12 Apr 2018 15:39:37	U	SMPL
J2163-03 - 1	PL-01-041118-B	1399	0.3415	-	7471B	PPB	12 Apr 2018 15:41:53	U	SMPL
J2163-05 - 1	PL-01-041118-C	1087	0.2663	-	7471B	PPB	12 Apr 2018 15:44:09	U	SMPL
J2163-07 - 1	PL-01-041118-D	1120	0.2742	-	7471B	PPB	12 Apr 2018 15:46:25	U	SMPL
J2165-03 - 1	SU-03-041118	3134	0.7601	-	7471B	PPB	12 Apr 2018 15:48:42	U	SMPL
J2310-01 - 1	P004-S003-0001-01	3854	0.9338	-	7471B	PPB	12 Apr 2018 15:55:59	U	SMPL
J2310-01DUP - 1	P004-S003-0001-01DUP	3834	0.9290	-	7471B	PPB	12 Apr 2018 15:58:16	U	SMPL
J2310-02 - 1	J2310-01MS	18241	4.4047	-	7471B	PPB	12 Apr 2018 16:00:34	U	SMPL
J2310-03 - 1	J2310-01MSD	18339	4.4283	-	7471B	PPB	12 Apr 2018 16:02:51	U	SMPL
J2310-04 - 1	P004-S003-0001-02	1824	0.4441	-	7471B	PPB	12 Apr 2018 16:05:07	U	SMPL
CCV23 - 1	CCV23	20642	4.9840	-	7471B	PPB	12 Apr 2018 16:07:24	U	SMPL
CCB23 - 1	CCB23	-94	-0.0187	-	7471B	PPB	12 Apr 2018 16:09:40	U	SMPL
J2310-05 - 1	P004-S004-0001-01	1666	0.4059	-	7471B	PPB	12 Apr 2018 16:11:57	U	SMPL
J2319-01 - 1	DRUM-1	79	0.0231	-	7471B	PPB	12 Apr 2018 16:14:14	U	SMPL
J2325-01 - 1	FK-GF-02-015-A	2385	0.5794	-	7471B	PPB	12 Apr 2018 16:16:31	U	SMPL
J2325-03 - 1	FK-GF-02-015-B	2391	0.5808	-	7471B	PPB	12 Apr 2018 16:18:47	U	SMPL
J2335-01 - 1	S1	12737	3.0768	-	7471B	PPB	12 Apr 2018 16:21:04	U	SMPL
J2335-03 - 1	S2	13384	3.2329	-	7471B	PPB	12 Apr 2018 16:23:21	U	SMPL
J2335-05 - 1	S3	4091	0.9910	-	7471B	PPB	12 Apr 2018 16:25:38	U	SMPL
J2310-01LX5 - 1		695	0.1717	-	7471B	PPB	12 Apr 2018 16:27:56	U	SMPL
J2310-01A - 1		20714	5.0013	-	7471B	PPB	12 Apr 2018 16:30:14	U	SMPL
CCV24 - 1	CCV24	20679	4.9929	-	7471B	PPB	12 Apr 2018 16:32:30	U	SMPL
CCB24 - 1	CCB24	-169	-0.0368	-	7471B	PPB	12 Apr 2018 16:34:46	U	SMPL