



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB94579

Slope : 97.6

BalanceID: WC SC-4

pH Meter ID : wc ph meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.3	4.01	04/12/2018	16:35
2	CAL2	1	Water			20.3	7.01	04/12/2018	16:36
3	CAL3	1	Water			20.3	10.01	04/12/2018	16:37
4	ICV	1	Water			20.2	7.00	04/12/2018	16:38
5	CCV1	1	Water			20.3	2.01	04/12/2018	16:40
6	J2152-02	1	Solid	20.02	20	20.9	5.96	04/12/2018	16:42
7	J2152-02DUP	1	Solid	20.03	20	20.5	5.98	04/12/2018	16:44
8	J2166-04	1	Solid	20.02	20	21.5	7.36	04/12/2018	16:45
9	J2348-02	1	Solid	20.03	20	21.1	8.70	04/12/2018	16:46
10	J2348-04	1	Solid	20.03	20	20.9	8.03	04/12/2018	16:47
11	J2348-06	1	Solid	20.01	20	20.9	8.20	04/12/2018	16:48
12	J2348-08	1	Solid	20.03	20	21.0	9.46	04/12/2018	16:50
13	J2348-10	1	Solid	20.02	20	20.8	8.63	04/12/2018	16:51
14	J2357-02	1	Solid	20.03	20	20.9	10.45	04/12/2018	17:00
15	J2364-04	1	Solid	20.02	20	20.8	8.18	04/12/2018	17:01
16	CCV2	1	Water			20.2	12.01	04/12/2018	17:02
17	J2364-08	1	Solid	20.05	20	21.0	8.30	04/12/2018	17:03
18	J2364-12	1	Solid	20.04	20	21.1	7.48	04/12/2018	17:05
19	CCV3	1	Water			20.3	2.02	04/12/2018	17:06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity J12364

WorkList ID : 110717

Date : 4/12/2018 2:54:07 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/12/2018	Solid	J2152-02	Corrosivity	Cool 4 deg C	PSEG05	N63	BU-01-041218	04/12/2018	9045C
04/12/2018	Solid	J2166-04	Corrosivity	Cool 4 deg C	PSEG05		TR-04-041218	04/12/2018	9045C
04/12/2018	Solid	J2348-02	Corrosivity	Cool 4 deg C	PSEG01	N63	135-348	04/12/2018	9045C
04/12/2018	Solid	J2348-04	Corrosivity	Cool 4 deg C	PSEG01	N63	132	04/12/2018	9045C
04/12/2018	Solid	J2348-06	Corrosivity	Cool 4 deg C	PSEG01	N63	155	04/12/2018	9045C
04/12/2018	Solid	J2348-08	Corrosivity	Cool 4 deg C	PSEG01	N63	275	04/12/2018	9045C
04/12/2018	Solid	J2348-10	Corrosivity	Cool 4 deg C	PSEG01	N63	281	04/12/2018	9045C
04/12/2018	Solid	J2357-02	Corrosivity	Cool 4 deg C	PSEG01	N63	RBR2520258	04/12/2018	9045C
04/12/2018	Solid	J2364-04	Corrosivity	Cool 4 deg C	PSEG01	N63	SP-A	04/12/2018	9045C
04/12/2018	Solid	J2364-08	Corrosivity	Cool 4 deg C	PSEG01	N63	SP-B	04/12/2018	9045C
04/12/2018	Solid	J2364-12	Corrosivity	Cool 4 deg C	PSEG01	N63	SP-C	04/12/2018	9045C

Date/Time 04-12-18 6:10 PM

Received by: JP

Relinquished by: JP

Date/Time 04-12-18 6:10 PM

Received by: JP

Relinquished by: JP